

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** HDRI08
Lab Code: ACE **SDG NO.:** Q3649
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/17/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115154.D **Time Analyzed:** 20:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.485	5.387	5.587	507.590	500.000	1.5
Aroclor-1016-2	5.507	5.408	5.608	503.560	500.000	0.7
Aroclor-1016-3	5.568	5.470	5.670	525.400	500.000	5.1
Aroclor-1016-4	5.664	5.566	5.766	510.160	500.000	2.0
Aroclor-1016-5	5.956	5.858	6.058	489.090	500.000	-2.2
Aroclor-1260-1	7.069	6.971	7.171	484.690	500.000	-3.1
Aroclor-1260-2	7.323	7.225	7.425	522.230	500.000	4.4
Aroclor-1260-3	7.680	7.582	7.782	519.970	500.000	4.0
Aroclor-1260-4	7.902	7.803	8.003	505.390	500.000	1.1
Aroclor-1260-5	8.209	8.110	8.310	501.090	500.000	0.2
Decachlorobiphenyl	9.927	9.830	10.030	55.570	50.000	11.1
Tetrachloro-m-xylene	4.346	4.248	4.448	52.550	50.000	5.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** HDRI08
Lab Code: ACE **SDG NO.:** Q3649
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/17/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115154.D **Time Analyzed:** 20:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.725	4.628	4.828	502.490	500.000	0.5
Aroclor-1016-2	4.744	4.647	4.847	500.710	500.000	0.1
Aroclor-1016-3	4.919	4.821	5.021	514.870	500.000	3.0
Aroclor-1016-4	4.960	4.864	5.064	504.420	500.000	0.9
Aroclor-1016-5	5.172	5.076	5.276	522.500	500.000	4.5
Aroclor-1260-1	6.202	6.106	6.306	469.120	500.000	-6.2
Aroclor-1260-2	6.389	6.294	6.494	453.530	500.000	-9.3
Aroclor-1260-3	6.542	6.446	6.646	468.490	500.000	-6.3
Aroclor-1260-4	7.012	6.917	7.117	472.980	500.000	-5.4
Aroclor-1260-5	7.254	7.158	7.358	472.000	500.000	-5.6
Decachlorobiphenyl	8.636	8.541	8.741	52.610	50.000	5.2
Tetrachloro-m-xylene	3.648	3.548	3.748	51.260	50.000	2.5