

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** HDRI08
Lab Code: ACE **SDG NO.:** Q3649
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/17/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076540.D **Time Analyzed:** 15:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.791	5.695	5.895	489.750	500.000	-2.1
Aroclor-1016-2	5.813	5.717	5.917	496.120	500.000	-0.8
Aroclor-1016-3	5.876	5.780	5.980	491.300	500.000	-1.7
Aroclor-1016-4	5.973	5.877	6.077	498.060	500.000	-0.4
Aroclor-1016-5	6.265	6.169	6.369	475.460	500.000	-4.9
Aroclor-1260-1	7.383	7.287	7.487	524.740	500.000	4.9
Aroclor-1260-2	7.636	7.540	7.740	526.150	500.000	5.2
Aroclor-1260-3	7.994	7.899	8.099	541.370	500.000	8.3
Aroclor-1260-4	8.222	8.127	8.327	537.080	500.000	7.4
Aroclor-1260-5	8.547	8.452	8.652	489.240	500.000	-2.2
Decachlorobiphenyl	10.410	10.319	10.519	46.120	50.000	-7.8
Tetrachloro-m-xylene	4.639	4.542	4.742	49.110	50.000	-1.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** HDRI08
Lab Code: ACE **SDG NO.:** Q3649
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/17/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076540.D **Time Analyzed:** 15:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.871	4.778	4.978	507.140	500.000	1.4
Aroclor-1016-2	4.928	4.836	5.036	521.340	500.000	4.3
Aroclor-1016-3	5.048	4.954	5.154	484.200	500.000	-3.2
Aroclor-1016-4	5.088	4.996	5.196	525.010	500.000	5.0
Aroclor-1016-5	5.303	5.210	5.410	490.170	500.000	-2.0
Aroclor-1260-1	6.330	6.238	6.438	533.610	500.000	6.7
Aroclor-1260-2	6.518	6.425	6.625	515.420	500.000	3.1
Aroclor-1260-3	6.671	6.578	6.778	502.990	500.000	0.6
Aroclor-1260-4	7.139	7.047	7.247	512.210	500.000	2.4
Aroclor-1260-5	7.380	7.287	7.487	473.160	500.000	-5.4
Decachlorobiphenyl	8.772	8.682	8.882	42.800	50.000	-14.4
Tetrachloro-m-xylene	3.777	3.681	3.881	51.440	50.000	2.9