

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** HDRI08
Lab Code: ACE **SDG NO.:** Q3649
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/17/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076555.D **Time Analyzed:** 20:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.791	5.695	5.895	483.500	500.000	-3.3
Aroclor-1016-2	5.812	5.717	5.917	501.210	500.000	0.2
Aroclor-1016-3	5.875	5.780	5.980	485.120	500.000	-3.0
Aroclor-1016-4	5.973	5.877	6.077	494.330	500.000	-1.1
Aroclor-1016-5	6.265	6.169	6.369	491.800	500.000	-1.6
Aroclor-1260-1	7.383	7.287	7.487	563.340	500.000	12.7
Aroclor-1260-2	7.636	7.540	7.740	510.020	500.000	2.0
Aroclor-1260-3	7.994	7.899	8.099	560.910	500.000	12.2
Aroclor-1260-4	8.221	8.127	8.327	511.620	500.000	2.3
Aroclor-1260-5	8.547	8.452	8.652	501.380	500.000	0.3
Decachlorobiphenyl	10.410	10.319	10.519	46.520	50.000	-7.0
Tetrachloro-m-xylene	4.639	4.542	4.742	50.780	50.000	1.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** HDRI08
Lab Code: ACE **SDG NO.:** Q3649
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/17/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076555.D **Time Analyzed:** 20:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.870	4.778	4.978	527.910	500.000	5.6
Aroclor-1016-2	4.929	4.836	5.036	569.940	500.000	14.0
Aroclor-1016-3	5.047	4.954	5.154	546.760	500.000	9.4
Aroclor-1016-4	5.090	4.996	5.196	536.870	500.000	7.4
Aroclor-1016-5	5.303	5.210	5.410	512.750	500.000	2.6
Aroclor-1260-1	6.329	6.238	6.438	417.310	500.000	-16.5
Aroclor-1260-2	6.517	6.425	6.625	472.870	500.000	-5.4
Aroclor-1260-3	6.669	6.578	6.778	507.970	500.000	1.6
Aroclor-1260-4	7.139	7.047	7.247	471.800	500.000	-5.6
Aroclor-1260-5	7.379	7.287	7.487	462.940	500.000	-7.4
Decachlorobiphenyl	8.770	8.682	8.882	43.060	50.000	-13.9
Tetrachloro-m-xylene	3.776	3.681	3.881	53.220	50.000	6.4