

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3649

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No. (PEM): PEM - PD091219.D Date Analyzed: 11/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:04

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.067	8.970	9.170	20.390	20.000	2.0
Tetrachloro-m-xylene	3.544	3.490	3.590	21.040	20.000	5.2
alpha-BHC	3.993	3.940	4.040	8.920	10.000	-10.8
beta-BHC	4.511	4.460	4.560	9.940	10.000	-0.6
gamma-BHC (Lindane)	4.324	4.270	4.370	9.120	10.000	-8.8
Endrin	6.567	6.500	6.640	44.720	50.000	-10.6
4,4'-DDT	7.016	6.950	7.090	91.520	100.000	-8.5
Methoxychlor	7.489	7.420	7.560	211.580	250.000	-15.4

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No. (PEM): PEM - PD091219.D Date Analyzed: 11/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:04

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.060	7.960	8.160	20.690	20.000	3.5
Tetrachloro-m-xylene	2.873	2.820	2.920	22.610	20.000	13.1
alpha-BHC	3.385	3.330	3.440	11.010	10.000	10.1
beta-BHC	4.017	3.970	4.070	11.240	10.000	12.4
gamma-BHC (Lindane)	3.721	3.670	3.770	11.030	10.000	10.3
Endrin	5.777	5.710	5.850	47.210	50.000	-5.6
4,4'-DDT	6.171	6.100	6.240	98.490	100.000	-1.5
Methoxychlor	6.741	6.670	6.810	191.760	250.000	-23.3