

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following "Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

LAB CHRONICLE

OrderID:	Q3649	OrderDate:	11/14/2025 3:02:13 PM
Client:	HDR, Inc.	Project:	PVWC Linear Construction
Contact:	Andrew Wadden	Location:	D41,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3649-01	B5-0.0-0.5-20251114	SOIL			11/14/25			11/14/25
			PCB	8082A		11/17/25	11/17/25	
			Pesticide-TCL	8081B		11/17/25	11/17/25	
Q3649-03	DUPE-0.0-0.5-20251114	SOIL			11/14/25			11/14/25
			PCB	8082A		11/17/25	11/17/25	
			Pesticide-TCL	8081B		11/17/25	11/17/25	
Q3649-06	B4-0.0-0.5-20251114	SOIL			11/14/25			11/14/25
			PCB	8082A		11/17/25	11/17/25	
			Pesticide-TCL	8081B		11/17/25	11/17/25	
Q3649-08	B3-0.0-0.5-20251114	SOIL			11/14/25			11/14/25
			PCB	8082A		11/17/25	11/17/25	
			Pesticide-TCL	8081B		11/17/25	11/17/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3649

Order ID: Q3649

Client: HDR, Inc.

Project ID: PVWC Linear Construction

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :	DUPE-0.0-0.5-20251114							
Q3649-03	DUPE-0.0-0.5-202511	SOIL	4,4-DDT	2.60	P	0.16	2.00	ug/kg
Total Concentration:				2.600				



QC SUMMARY

Surrogate Summary

SDG No.: Q3649

Client: HDR, Inc.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD091193.D	PIBLK-PD091193.D	Decachlorobiphen	1	20	20.8	104		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	19.5	98		30 (41)	150 (134)
		Decachlorobiphen	2	20	20.1	100		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	20.0	100		30 (41)	150 (134)
I.BLK-PD091218.D	PIBLK-PD091218.D	Decachlorobiphen	1	20	20.0	100		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	19.6	98		30 (41)	150 (134)
		Decachlorobiphen	2	20	19.3	96		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	22.6	113		30 (41)	150 (134)
PB170575BL	PB170575BL	Decachlorobiphen	1	20	19.1	96		30 (10)	150 (143)
		Tetrachloro-m-xyl	1	20	18.2	91		30 (10)	150 (131)
		Decachlorobiphen	2	20	19.9	100		30 (10)	150 (143)
		Tetrachloro-m-xyl	2	20	19.7	98		30 (10)	150 (131)
PB170575BS	PB170575BS	Decachlorobiphen	1	20	23.6	118		30 (10)	150 (143)
		Tetrachloro-m-xyl	1	20	21.9	109		30 (10)	150 (131)
		Decachlorobiphen	2	20	25.0	125		30 (10)	150 (143)
		Tetrachloro-m-xyl	2	20	24.0	120		30 (10)	150 (131)
I.BLK-PD091234.D	PIBLK-PD091234.D	Decachlorobiphen	1	20	20.8	104		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	20.0	100		30 (41)	150 (134)
		Decachlorobiphen	2	20	18.9	94		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	21.4	107		30 (41)	150 (134)
Q3649-01	B5-0.0-0.5-20251114	Decachlorobiphen	1	20	16.7	84		30 (10)	150 (143)
		Tetrachloro-m-xyl	1	20	16.8	84		30 (10)	150 (131)
		Decachlorobiphen	2	20	15.4	77		30 (10)	150 (143)
		Tetrachloro-m-xyl	2	20	17.9	89		30 (10)	150 (131)
Q3649-03	DUPE-0.0-0.5-20251114	Decachlorobiphen	1	20	14.2	71		30 (10)	150 (143)
		Tetrachloro-m-xyl	1	20	17.7	88		30 (10)	150 (131)
		Decachlorobiphen	2	20	12.6	63		30 (10)	150 (143)
		Tetrachloro-m-xyl	2	20	16.6	83		30 (10)	150 (131)
Q3649-03MS	DUPE-0.0-0.5-20251114MS	Decachlorobiphen	1	20	14.4	72		30 (10)	150 (143)
		Tetrachloro-m-xyl	1	20	16.8	84		30 (10)	150 (131)
		Decachlorobiphen	2	20	12.1	61		30 (10)	150 (143)
		Tetrachloro-m-xyl	2	20	17.4	87		30 (10)	150 (131)
Q3649-03MSD	DUPE-0.0-0.5-20251114MSD	Decachlorobiphen	1	20	13.1	65		30 (10)	150 (143)
		Tetrachloro-m-xyl	1	20	16.3	81		30 (10)	150 (131)
		Decachlorobiphen	2	20	11.4	57		30 (10)	150 (143)
		Tetrachloro-m-xyl	2	20	17.1	86		30 (10)	150 (131)
Q3649-06	B4-0.0-0.5-20251114	Decachlorobiphen	1	20	13.5	67		30 (10)	150 (143)
		Tetrachloro-m-xyl	1	20	17.1	85		30 (10)	150 (131)
		Decachlorobiphen	2	20	11.5	58		30 (10)	150 (143)
		Tetrachloro-m-xyl	2	20	18.1	91		30 (10)	150 (131)
Q3649-08	B3-0.0-0.5-20251114	Decachlorobiphen	1	20	16.3	81		30 (10)	150 (143)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: Q3649

Client: HDR, Inc.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q3649-08	B3-0.0-0.5-20251114	Tetrachloro-m-xyl	1	20	19.1	96		30 (10)	150 (131)
		Decachlorobiphen	2	20	14.7	74		30 (10)	150 (143)
		Tetrachloro-m-xyl	2	20	19.2	96		30 (10)	150 (131)
I.BLK-PD091243.D	PIBLK-PD091243.D	Decachlorobiphen	1	20	20.3	102		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	19.8	99		30 (41)	150 (134)
		Decachlorobiphen	2	20	17.5	88		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	21.2	106		30 (41)	150 (134)

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3649</u>	Analytical Method:	<u>8081B</u>
Client:	<u>HDR, Inc.</u>	DataFile :	<u>PD091239.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits		RPD
			Result	Result							High		
Lab Sample ID:	Q3649-03MS (Column 1)	Client Sample ID:		DUPE-0.0-0.5-20251114MS									
	alpha-BHC	19.24	0	17.5	ug/kg	91				30 (60)	150 (144)		
	beta-BHC	19.24	0	16.4	ug/kg	85				30 (54)	150 (143)		
	delta-BHC	19.24	0	16.7	ug/kg	87				30 (47)	150 (129)		
	gamma-BHC (Lindane)	19.24	0	17.5	ug/kg	91				30 (39)	150 (136)		
	Heptachlor	19.24	0	18.2	ug/kg	95				30 (33)	150 (148)		
	Aldrin	19.24	0	16.7	ug/kg	87				30 (38)	150 (139)		
	Heptachlor epoxide	19.24	0	16.1	ug/kg	84				30 (23)	150 (155)		
	Endosulfan I	19.24	0	16.9	ug/kg	88				30 (30)	150 (142)		
	Dieldrin	19.24	0	24.0	ug/kg	125				30 (32)	150 (140)		
	4,4'-DDE	19.24	0	15.8	ug/kg	82				30 (20)	150 (156)		
	Endrin	19.24	0	18.7	ug/kg	97				30 (57)	150 (139)		
	Endosulfan II	19.24	0	16.9	ug/kg	88				30 (41)	150 (125)		
	4,4'-DDD	19.24	0	15.8	ug/kg	82				30 (47)	150 (163)		
	Endosulfan sulfate	19.24	0	15.9	ug/kg	83				30 (43)	150 (125)		
	4,4'-DDT	19.24	1.50	21.4	ug/kg	103				30 (25)	150 (149)		
	Methoxychlor	19.24	0	18.9	ug/kg	98				30 (32)	150 (132)		
	Endrin ketone	19.24	0	15.5	ug/kg	81				30 (27)	150 (142)		
	Endrin aldehyde	19.24	0	16.7	ug/kg	87				30 (19)	150 (148)		
	alpha-Chlordane	19.24	0	16.2	ug/kg	84				30 (26)	150 (147)		
	gamma-Chlordane	19.24	0	16.2	ug/kg	84				30 (30)	150 (142)		
Lab Sample ID:	Q3649-03MS (Column 2)	Client Sample ID:		DUPE-0.0-0.5-20251114MS									
	alpha-BHC	19.24	0	17.5	ug/kg	91				30 (60)	150 (144)		
	beta-BHC	19.24	0	18.2	ug/kg	95				30 (54)	150 (143)		
	delta-BHC	19.24	0	16.6	ug/kg	86				30 (47)	150 (129)		
	gamma-BHC (Lindane)	19.24	0	17.0	ug/kg	88				30 (39)	150 (136)		
	Heptachlor	19.24	0	17.4	ug/kg	90				30 (33)	150 (148)		
	Aldrin	19.24	0	16.7	ug/kg	87				30 (38)	150 (139)		
	Heptachlor epoxide	19.24	0	16.4	ug/kg	85				30 (23)	150 (155)		
	Endosulfan I	19.24	0	14.5	ug/kg	75				30 (30)	150 (142)		
	Dieldrin	19.24	0	23.1	ug/kg	120				30 (32)	150 (140)		
	4,4'-DDE	19.24	0	16.0	ug/kg	83				30 (20)	150 (156)		
	Endrin	19.24	0	17.3	ug/kg	90				30 (57)	150 (139)		
	Endosulfan II	19.24	0	15.9	ug/kg	83				30 (41)	150 (125)		
	4,4'-DDD	19.24	0	12.6	ug/kg	65				30 (47)	150 (163)		
	Endosulfan sulfate	19.24	0	14.8	ug/kg	77				30 (43)	150 (125)		

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3649</u>	Analytical Method:	<u>8081B</u>
Client:	<u>HDR, Inc.</u>	DataFile :	<u>PD091239.D</u>

Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits		RPD
		Result	Result							High		
4,4'-DDT	19.24	2.60	19.8	ug/kg	89				30 (25)	150 (149)		
Methoxychlor	19.24	0	15.4	ug/kg	80				30 (32)	150 (132)		
Endrin ketone	19.24	0	11.7	ug/kg	61				30 (27)	150 (142)		
Endrin aldehyde	19.24	0	15.1	ug/kg	78				30 (19)	150 (148)		
alpha-Chlordane	19.24	0	17.0	ug/kg	88				30 (26)	150 (147)		
gamma-Chlordane	19.24	0	16.6	ug/kg	86				30 (30)	150 (142)		

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3649</u>	Analytical Method:	<u>8081B</u>
Client:	<u>HDR, Inc.</u>	DataFile :	<u>PD091240.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec	RPD		Low	Limits	
			Result				Qual	RPD	Qual		High	RPD
Lab Sample ID:	Q3649-03MSD (Column 1)		Client Sample ID:	DUPE-0.0-0.5-20251114MS								
	alpha-BHC	19.22	0	16.9	ug/kg	88		3		30 (60)	150 (144)	30 (15)
	beta-BHC	19.22	0	16.5	ug/kg	86		1		30 (54)	150 (143)	30 (15)
	delta-BHC	19.22	0	16.4	ug/kg	85		2		30 (47)	150 (129)	30 (20)
	gamma-BHC (Lindane)	19.22	0	16.9	ug/kg	88		3		30 (39)	150 (136)	30 (30)
	Heptachlor	19.22	0	16.9	ug/kg	88		8		30 (33)	150 (148)	30 (20)
	Aldrin	19.22	0	16.2	ug/kg	84		4		30 (38)	150 (139)	30 (30)
	Heptachlor epoxide	19.22	0	15.7	ug/kg	82		2		30 (23)	150 (155)	30 (16)
	Endosulfan I	19.22	0	16.2	ug/kg	84		5		30 (30)	150 (142)	30 (15)
	Dieldrin	19.22	0	23.1	ug/kg	120		4		30 (32)	150 (140)	30 (20)
	4,4'-DDE	19.22	0	15.2	ug/kg	79		4		30 (20)	150 (156)	30 (16)
	Endrin	19.22	0	18.0	ug/kg	94		3		30 (57)	150 (139)	30 (16)
	Endosulfan II	19.22	0	16.3	ug/kg	85		3		30 (41)	150 (125)	30 (15)
	4,4'-DDD	19.22	0	15.3	ug/kg	80		2		30 (47)	150 (163)	30 (20)
	Endosulfan sulfate	19.22	0	15.2	ug/kg	79		5		30 (43)	150 (125)	30 (15)
	4,4'-DDT	19.22	1.50	20.0	ug/kg	96		7		30 (25)	150 (149)	30 (15)
	Methoxychlor	19.22	0	17.2	ug/kg	89		10		30 (32)	150 (132)	30 (20)
	Endrin ketone	19.22	0	14.8	ug/kg	77		5		30 (27)	150 (142)	30 (15)
	Endrin aldehyde	19.22	0	15.8	ug/kg	82		6		30 (19)	150 (148)	30 (15)
	alpha-Chlordane	19.22	0	15.4	ug/kg	80		5		30 (26)	150 (147)	30 (20)
	gamma-Chlordane	19.22	0	15.7	ug/kg	82		2		30 (30)	150 (142)	30 (15)
Lab Sample ID:	Q3649-03MSD (Column 2)		Client Sample ID:	DUPE-0.0-0.5-20251114MS								
	alpha-BHC	19.22	0	17.2	ug/kg	89		2		30 (60)	150 (144)	30 (15)
	beta-BHC	19.22	0	17.8	ug/kg	93		2		30 (54)	150 (143)	30 (15)
	delta-BHC	19.22	0	16.2	ug/kg	84		2		30 (47)	150 (129)	30 (20)
	gamma-BHC (Lindane)	19.22	0	16.6	ug/kg	86		2		30 (39)	150 (136)	30 (30)
	Heptachlor	19.22	0	16.3	ug/kg	85		6		30 (33)	150 (148)	30 (20)
	Aldrin	19.22	0	16.3	ug/kg	85		2		30 (38)	150 (139)	30 (30)
	Heptachlor epoxide	19.22	0	16.0	ug/kg	83		2		30 (23)	150 (155)	30 (16)
	Endosulfan I	19.22	0	14.1	ug/kg	73		3		30 (30)	150 (142)	30 (15)
	Dieldrin	19.22	0	22.5	ug/kg	117		3		30 (32)	150 (140)	30 (20)
	4,4'-DDE	19.22	0	15.4	ug/kg	80		4		30 (20)	150 (156)	30 (16)
	Endrin	19.22	0	16.6	ug/kg	86		5		30 (57)	150 (139)	30 (16)
	Endosulfan II	19.22	0	15.3	ug/kg	80		4		30 (41)	150 (125)	30 (15)
	4,4'-DDD	19.22	0	12.2	ug/kg	63		3		30 (47)	150 (163)	30 (20)
	Endosulfan sulfate	19.22	0	14.0	ug/kg	73		5		30 (43)	150 (125)	30 (15)

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3649</u>	Analytical Method:	<u>8081B</u>
Client:	<u>HDR, Inc.</u>	DataFile :	<u>PD091240.D</u>

Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits	
		Result	Result							High	RPD
4,4'-DDT	19.22	2.60	18.1	ug/kg	81		9		30 (25)	150 (149)	30 (15)
Methoxychlor	19.22	0	14.1	ug/kg	73		9		30 (32)	150 (132)	30 (20)
Endrin ketone	19.22	0	10.9	ug/kg	57		7		30 (27)	150 (142)	30 (15)
Endrin aldehyde	19.22	0	14.3	ug/kg	74		5		30 (19)	150 (148)	30 (15)
alpha-Chlordane	19.22	0	16.5	ug/kg	86		2		30 (26)	150 (147)	30 (20)
gamma-Chlordane	19.22	0	17.3	ug/kg	90		5		30 (30)	150 (142)	30 (15)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3649

Analytical Method: 8081B

Client: HDR, Inc.

Datafile : PD091224.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB170575BS (Column 1)	alpha-BHC	16.66	16.3	ug/kg	98				40 (84)	140 (123)	
	beta-BHC	16.66	15.6	ug/kg	94				40 (82)	140 (123)	
	delta-BHC	16.66	16.5	ug/kg	99				40 (83)	140 (126)	
	gamma-BHC (Lindane)	16.66	16.3	ug/kg	98				40 (83)	140 (125)	
	Heptachlor	16.66	16.3	ug/kg	98				40 (83)	140 (122)	
	Aldrin	16.66	16.2	ug/kg	97				40 (82)	140 (124)	
	Heptachlor epoxide	16.66	16.0	ug/kg	96				40 (83)	140 (120)	
	Endosulfan I	16.66	16.0	ug/kg	96				40 (81)	140 (124)	
	Dieldrin	16.66	16.3	ug/kg	98				40 (85)	140 (121)	
	4,4'-DDE	16.66	16.3	ug/kg	98				40 (81)	140 (123)	
	Endrin	16.66	15.0	ug/kg	90				40 (76)	140 (130)	
	Endosulfan II	16.66	16.1	ug/kg	97				40 (80)	140 (125)	
	4,4'-DDD	16.66	16.0	ug/kg	96				40 (80)	140 (131)	
	Endosulfan sulfate	16.66	15.9	ug/kg	95				40 (81)	140 (122)	
	4,4'-DDT	16.66	16.5	ug/kg	99				40 (70)	140 (129)	
	Methoxychlor	16.66	15.6	ug/kg	94				40 (60)	140 (119)	
	Endrin ketone	16.66	16.5	ug/kg	99				40 (77)	140 (132)	
	Endrin aldehyde	16.66	16.5	ug/kg	99				40 (79)	140 (124)	
	alpha-Chlordane	16.66	15.3	ug/kg	92				40 (84)	140 (120)	
	gamma-Chlordane	16.66	16.3	ug/kg	98				40 (83)	140 (122)	
PB170575BS (Column 2)	alpha-BHC	16.66	17.1	ug/kg	103				40 (84)	140 (123)	
	beta-BHC	16.66	16.9	ug/kg	101				40 (82)	140 (123)	
	delta-BHC	16.66	17.1	ug/kg	103				40 (83)	140 (126)	
	gamma-BHC (Lindane)	16.66	17.1	ug/kg	103				40 (83)	140 (125)	
	Heptachlor	16.66	16.9	ug/kg	101				40 (83)	140 (122)	
	Aldrin	16.66	17.2	ug/kg	103				40 (82)	140 (124)	
	Heptachlor epoxide	16.66	17.0	ug/kg	102				40 (83)	140 (120)	
	Endosulfan I	16.66	17.1	ug/kg	103				40 (81)	140 (124)	
	Dieldrin	16.66	17.1	ug/kg	103				40 (85)	140 (121)	
	4,4'-DDE	16.66	17.1	ug/kg	103				40 (81)	140 (123)	
	Endrin	16.66	15.6	ug/kg	94				40 (76)	140 (130)	
	Endosulfan II	16.66	17.0	ug/kg	102				40 (80)	140 (125)	
	4,4'-DDD	16.66	16.7	ug/kg	100				40 (80)	140 (131)	
	Endosulfan sulfate	16.66	17.0	ug/kg	102				40 (81)	140 (122)	
	4,4'-DDT	16.66	17.8	ug/kg	107				40 (70)	140 (129)	
	Methoxychlor	16.66	16.6	ug/kg	100				40 (60)	140 (119)	
	Endrin ketone	16.66	17.7	ug/kg	106				40 (77)	140 (132)	
	Endrin aldehyde	16.66	17.6	ug/kg	106				40 (79)	140 (124)	
	alpha-Chlordane	16.66	17.3	ug/kg	104				40 (84)	140 (120)	
	gamma-Chlordane	16.66	17.3	ug/kg	104				40 (83)	140 (122)	

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID

PB170575BL

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3649

Lab Sample ID: PB170575BL

Lab File ID: PD091223.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 11/17/2025

Date Analyzed (1): 11/17/2025

Date Analyzed (2): 11/17/2025

Time Analyzed (1): 13:46

Time Analyzed (2): 13:46

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB170575BS	PB170575BS	PD091224.D	11/17/2025	11/17/2025
B5-0.0-0.5-20251114	Q3649-01	PD091237.D	11/17/2025	11/17/2025
DUPE-0.0-0.5-20251114	Q3649-03	PD091238.D	11/17/2025	11/17/2025
DUPE-0.0-0.5-20251114MS	Q3649-03MS	PD091239.D	11/17/2025	11/17/2025
DUPE-0.0-0.5-20251114MSD	Q3649-03MSD	PD091240.D	11/17/2025	11/17/2025
B4-0.0-0.5-20251114	Q3649-06	PD091241.D	11/17/2025	11/17/2025
B3-0.0-0.5-20251114	Q3649-08	PD091242.D	11/17/2025	11/17/2025

COMMENTS: _____



SAMPLE DATA



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	HDR, Inc.	Date Collected:	11/14/25
Project:	PVWC Linear Construction	Date Received:	11/14/25
Client Sample ID:	B5-0.0-0.5-20251114	SDG No.:	Q3649
Lab Sample ID:	Q3649-01	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	90.6
Sample Wt/Vol:	30.02 g	Final Vol:	10000 uL
Prep Method:	SW3541B	Test:	Pesticide-TCL
	Prep Date:	11/17/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.14	U	1	0.14	1.90	ug/kg	11/17/25 21:11	PB170575
319-85-7	beta-BHC	0.20	U	1	0.20	1.90	ug/kg	11/17/25 21:11	PB170575
319-86-8	delta-BHC	0.43	U	1	0.43	1.90	ug/kg	11/17/25 21:11	PB170575
58-89-9	gamma-BHC (Lindane)	0.15	U	1	0.15	1.90	ug/kg	11/17/25 21:11	PB170575
76-44-8	Heptachlor	0.13	U	1	0.13	1.90	ug/kg	11/17/25 21:11	PB170575
309-00-2	Aldrin	0.13	U	1	0.13	1.90	ug/kg	11/17/25 21:11	PB170575
1024-57-3	Heptachlor epoxide	0.21	U	1	0.21	1.90	ug/kg	11/17/25 21:11	PB170575
959-98-8	Endosulfan I	0.15	U	1	0.15	1.90	ug/kg	11/17/25 21:11	PB170575
60-57-1	Dieldrin	0.15	U	1	0.15	1.90	ug/kg	11/17/25 21:11	PB170575
72-55-9	4,4-DDE	0.15	U	1	0.15	1.90	ug/kg	11/17/25 21:11	PB170575
72-20-8	Endrin	0.15	U	1	0.15	1.90	ug/kg	11/17/25 21:11	PB170575
33213-65-9	Endosulfan II	0.32	U	1	0.32	1.90	ug/kg	11/17/25 21:11	PB170575
72-54-8	4,4-DDD	0.17	U	1	0.17	1.90	ug/kg	11/17/25 21:11	PB170575
1031-07-8	Endosulfan Sulfate	0.14	U	1	0.14	1.90	ug/kg	11/17/25 21:11	PB170575
50-29-3	4,4-DDT	0.15	U	1	0.15	1.90	ug/kg	11/17/25 21:11	PB170575
72-43-5	Methoxychlor	0.41	U	1	0.41	1.90	ug/kg	11/17/25 21:11	PB170575
53494-70-5	Endrin ketone	0.21	U	1	0.21	1.90	ug/kg	11/17/25 21:11	PB170575
7421-93-4	Endrin aldehyde	0.41	U	1	0.41	1.90	ug/kg	11/17/25 21:11	PB170575
5103-71-9	alpha-Chlordane	0.13	U	1	0.13	1.90	ug/kg	11/17/25 21:11	PB170575
5103-74-2	gamma-Chlordane	0.17	U	1	0.17	1.90	ug/kg	11/17/25 21:11	PB170575
8001-35-2	Toxaphene	6.00	U	1	6.00	36.4	ug/kg	11/17/25 21:11	PB170575
SURROGATES									
2051-24-3	Decachlorobiphenyl	16.7			30 (10) - 150 (143)	84%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	17.9			30 (10) - 150 (131)	89%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
 Data File : PD091237.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2025 21:11
 Operator : AR\AJ
 Sample : Q3649-01
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 B5-0.0-0.5-20251114

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 11/18/2025
 Supervised By :mohammad ahmed 11/19/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 01:11:31 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.540	2.870	57760958	654.1E6	16.749m	17.857m
28) SA Decachlor...	9.062	8.056	83360352	609.2E6	16.733	15.394m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091237.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 21:11
Operator : AR\AJ
Sample : Q3649-01
Misc :
ALS Vial : 18 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
B5-0.0-0.5-20251114


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 11/18/2025
Supervised By : mohammad ahmed 11/19/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:11:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.540 min
Delta R.T.: -0.004 min
Response: 57760958
Conc: 16.75 ng/ml

Instrument :
ECD_D
ClientSampleId :
B5-0.0-0.5-20251114


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 11/18/2025
Supervised By :mohammad ahmed 11/19/2025




#1 Tetrachloro-m-xylene

R.T.: 2.870 min
Delta R.T.: -0.004 min
Response: 654107954
Conc: 17.86 ng/ml m



#28 Decachlorobiphenyl

R.T.: 9.062 min
Delta R.T.: -0.001 min
Response: 83360352
Conc: 16.73 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.056 min
Delta R.T.: -0.003 min
Response: 609214233
Conc: 15.39 ng/ml m



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	HDR, Inc.		Date Collected:	11/14/25
Project:	PVWC Linear Construction		Date Received:	11/14/25
Client Sample ID:	DUPE-0.0-0.5-20251114		SDG No.:	Q3649
Lab Sample ID:	Q3649-03		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	86.5
Sample Wt/Vol:	30.09 g	Final Vol: 10000 uL	Test:	Pesticide-TCL
Prep Method:	SW3541B	Prep Date: 11/17/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.15	U	1	0.15	2.00	ug/kg	11/17/25 21:24	PB170575
319-85-7	beta-BHC	0.21	U	1	0.21	2.00	ug/kg	11/17/25 21:24	PB170575
319-86-8	delta-BHC	0.45	U	1	0.45	2.00	ug/kg	11/17/25 21:24	PB170575
58-89-9	gamma-BHC (Lindane)	0.16	U	1	0.16	2.00	ug/kg	11/17/25 21:24	PB170575
76-44-8	Heptachlor	0.14	U	1	0.14	2.00	ug/kg	11/17/25 21:24	PB170575
309-00-2	Aldrin	0.14	U	1	0.14	2.00	ug/kg	11/17/25 21:24	PB170575
1024-57-3	Heptachlor epoxide	0.22	U	1	0.22	2.00	ug/kg	11/17/25 21:24	PB170575
959-98-8	Endosulfan I	0.16	U	1	0.16	2.00	ug/kg	11/17/25 21:24	PB170575
60-57-1	Dieldrin	0.16	U	1	0.16	2.00	ug/kg	11/17/25 21:24	PB170575
72-55-9	4,4-DDE	0.16	U	1	0.16	2.00	ug/kg	11/17/25 21:24	PB170575
72-20-8	Endrin	0.16	U	1	0.16	2.00	ug/kg	11/17/25 21:24	PB170575
33213-65-9	Endosulfan II	0.33	U	1	0.33	2.00	ug/kg	11/17/25 21:24	PB170575
72-54-8	4,4-DDD	0.17	U	1	0.17	2.00	ug/kg	11/17/25 21:24	PB170575
1031-07-8	Endosulfan Sulfate	0.15	U	1	0.15	2.00	ug/kg	11/17/25 21:24	PB170575
50-29-3	4,4-DDT	2.60	P	1	0.16	2.00	ug/kg	11/17/25 21:24	PB170575
72-43-5	Methoxychlor	0.43	U	1	0.43	2.00	ug/kg	11/17/25 21:24	PB170575
53494-70-5	Endrin ketone	0.22	U	1	0.22	2.00	ug/kg	11/17/25 21:24	PB170575
7421-93-4	Endrin aldehyde	0.43	U	1	0.43	2.00	ug/kg	11/17/25 21:24	PB170575
5103-71-9	alpha-Chlordane	0.14	U	1	0.14	2.00	ug/kg	11/17/25 21:24	PB170575
5103-74-2	gamma-Chlordane	0.17	U	1	0.17	2.00	ug/kg	11/17/25 21:24	PB170575
8001-35-2	Toxaphene	6.20	U	1	6.20	38.0	ug/kg	11/17/25 21:24	PB170575
SURROGATES									
2051-24-3	Decachlorobiphenyl	14.2			30 (10) - 150 (143)	71%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	17.7			30 (10) - 150 (131)	88%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
 Data File : PD091238.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2025 21:24
 Operator : AR\AJ
 Sample : Q3649-03
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 DUPE-0.0-0.5-20251114

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 11/18/2025
 Supervised By :mohammad ahmed 11/19/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 08:51:21 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.540	2.887	60979394	609.7E6	17.682m	16.644m
28) SA Decachlor...	9.062	8.072	70508420	499.7E6	14.153	12.626m
Target Compounds						
17) MA 4,4'-DDT	7.013	6.184	16270083	265.4E6	3.934	6.782 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091238.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 21:24
Operator : AR\AJ
Sample : Q3649-03
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
DUPE-0.0-0.5-20251114


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 11/18/2025
Supervised By : mohammad ahmed 11/19/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 08:51:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.540 min
Delta R.T.: -0.004 min
Response: 60979394
Conc: 17.68 ng/ml

Instrument :
ECD_D
ClientSampleId :
DUPE-0.0-0.5-20251114


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 11/18/2025
Supervised By :mohammad ahmed 11/19/2025




#1 Tetrachloro-m-xylene

R.T.: 2.887 min
Delta R.T.: 0.013 min
Response: 609669024
Conc: 16.64 ng/ml m




#17 4,4'-DDT

R.T.: 7.013 min
Delta R.T.: -0.001 min
Response: 16270083
Conc: 3.93 ng/ml



#17 4,4'-DDT

R.T.: 6.184 min
Delta R.T.: 0.012 min
Response: 265395983
Conc: 6.78 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.062 min
Delta R.T.: 0.000 min
Response: 70508420
Conc: 14.15 ng/ml

Instrument :
ECD_D
ClientSampleId :
DUPE-0.0-0.5-20251114

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 11/18/2025
Supervised By :mohammad ahmed 11/19/2025



#28 Decachlorobiphenyl

R.T.: 8.072 min
Delta R.T.: 0.013 min
Response: 499670582
Conc: 12.63 ng/ml m



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	HDR, Inc.	Date Collected:	11/14/25
Project:	PVWC Linear Construction	Date Received:	11/14/25
Client Sample ID:	B4-0.0-0.5-20251114	SDG No.:	Q3649
Lab Sample ID:	Q3649-06	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	91.4
Sample Wt/Vol:	30.01 g	Final Vol:	10000 uL
Prep Method:	SW3541B	Test:	Pesticide-TCL
	Prep Date:		11/17/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.14	U	1	0.14	1.90	ug/kg	11/17/25 22:05	PB170575
319-85-7	beta-BHC	0.20	U	1	0.20	1.90	ug/kg	11/17/25 22:05	PB170575
319-86-8	delta-BHC	0.43	U	1	0.43	1.90	ug/kg	11/17/25 22:05	PB170575
58-89-9	gamma-BHC (Lindane)	0.15	U	1	0.15	1.90	ug/kg	11/17/25 22:05	PB170575
76-44-8	Heptachlor	0.13	U	1	0.13	1.90	ug/kg	11/17/25 22:05	PB170575
309-00-2	Aldrin	0.13	U	1	0.13	1.90	ug/kg	11/17/25 22:05	PB170575
1024-57-3	Heptachlor epoxide	0.21	U	1	0.21	1.90	ug/kg	11/17/25 22:05	PB170575
959-98-8	Endosulfan I	0.15	U	1	0.15	1.90	ug/kg	11/17/25 22:05	PB170575
60-57-1	Dieldrin	0.15	U	1	0.15	1.90	ug/kg	11/17/25 22:05	PB170575
72-55-9	4,4-DDE	0.15	U	1	0.15	1.90	ug/kg	11/17/25 22:05	PB170575
72-20-8	Endrin	0.15	U	1	0.15	1.90	ug/kg	11/17/25 22:05	PB170575
33213-65-9	Endosulfan II	0.32	U	1	0.32	1.90	ug/kg	11/17/25 22:05	PB170575
72-54-8	4,4-DDD	0.16	U	1	0.16	1.90	ug/kg	11/17/25 22:05	PB170575
1031-07-8	Endosulfan Sulfate	0.14	U	1	0.14	1.90	ug/kg	11/17/25 22:05	PB170575
50-29-3	4,4-DDT	0.15	U	1	0.15	1.90	ug/kg	11/17/25 22:05	PB170575
72-43-5	Methoxychlor	0.40	U	1	0.40	1.90	ug/kg	11/17/25 22:05	PB170575
53494-70-5	Endrin ketone	0.21	U	1	0.21	1.90	ug/kg	11/17/25 22:05	PB170575
7421-93-4	Endrin aldehyde	0.40	U	1	0.40	1.90	ug/kg	11/17/25 22:05	PB170575
5103-71-9	alpha-Chlordane	0.13	U	1	0.13	1.90	ug/kg	11/17/25 22:05	PB170575
5103-74-2	gamma-Chlordane	0.16	U	1	0.16	1.90	ug/kg	11/17/25 22:05	PB170575
8001-35-2	Toxaphene	5.90	U	1	5.90	36.1	ug/kg	11/17/25 22:05	PB170575
SURROGATES									
2051-24-3	Decachlorobiphenyl	13.5			30 (10) - 150 (143)	67%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	18.1			30 (10) - 150 (131)	91%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
 Data File : PD091241.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2025 22:05
 Operator : AR\AJ
 Sample : Q3649-06
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 B4-0.0-0.5-20251114

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 11/18/2025
 Supervised By :mohammad ahmed 11/19/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 01:13:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.540	2.872	58897174	663.1E6	17.078m	18.103
28)	SA Decachlor...	9.060	8.057	67181137	456.8E6	13.485m	11.542

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091241.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 22:05
Operator : AR\AJ
Sample : Q3649-06
Misc :
ALS Vial : 22 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
B4-0.0-0.5-20251114


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 11/18/2025
Supervised By :mohammad ahmed 11/19/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:13:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.540 min
Delta R.T.: -0.004 min
Response: 58897174
Conc: 17.08 ng/ml

Instrument :
ECD_D
ClientSampleId :
B4-0.0-0.5-20251114


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 11/18/2025
Supervised By :mohammad ahmed 11/19/2025




#1 Tetrachloro-m-xylene

R.T.: 2.872 min
Delta R.T.: -0.002 min
Response: 663113366
Conc: 18.10 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.060 min
Delta R.T.: -0.003 min
Response: 67181137
Conc: 13.49 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.057 min
Delta R.T.: -0.002 min
Response: 456777909
Conc: 11.54 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	HDR, Inc.	Date Collected:	11/14/25
Project:	PVWC Linear Construction	Date Received:	11/14/25
Client Sample ID:	B3-0.0-0.5-20251114	SDG No.:	Q3649
Lab Sample ID:	Q3649-08	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	86
Sample Wt/Vol:	30.05 g	Final Vol:	10000 uL
Prep Method:	SW3541B	Test:	Pesticide-TCL
	Prep Date:	11/17/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.15	U	1	0.15	2.00	ug/kg	11/17/25 22:19	PB170575
319-85-7	beta-BHC	0.21	U	1	0.21	2.00	ug/kg	11/17/25 22:19	PB170575
319-86-8	delta-BHC	0.45	U	1	0.45	2.00	ug/kg	11/17/25 22:19	PB170575
58-89-9	gamma-BHC (Lindane)	0.16	U	1	0.16	2.00	ug/kg	11/17/25 22:19	PB170575
76-44-8	Heptachlor	0.14	U	1	0.14	2.00	ug/kg	11/17/25 22:19	PB170575
309-00-2	Aldrin	0.14	U	1	0.14	2.00	ug/kg	11/17/25 22:19	PB170575
1024-57-3	Heptachlor epoxide	0.22	U	1	0.22	2.00	ug/kg	11/17/25 22:19	PB170575
959-98-8	Endosulfan I	0.16	U	1	0.16	2.00	ug/kg	11/17/25 22:19	PB170575
60-57-1	Dieldrin	0.16	U	1	0.16	2.00	ug/kg	11/17/25 22:19	PB170575
72-55-9	4,4-DDE	0.16	U	1	0.16	2.00	ug/kg	11/17/25 22:19	PB170575
72-20-8	Endrin	0.16	U	1	0.16	2.00	ug/kg	11/17/25 22:19	PB170575
33213-65-9	Endosulfan II	0.34	U	1	0.34	2.00	ug/kg	11/17/25 22:19	PB170575
72-54-8	4,4-DDD	0.17	U	1	0.17	2.00	ug/kg	11/17/25 22:19	PB170575
1031-07-8	Endosulfan Sulfate	0.15	U	1	0.15	2.00	ug/kg	11/17/25 22:19	PB170575
50-29-3	4,4-DDT	0.16	U	1	0.16	2.00	ug/kg	11/17/25 22:19	PB170575
72-43-5	Methoxychlor	0.43	U	1	0.43	2.00	ug/kg	11/17/25 22:19	PB170575
53494-70-5	Endrin ketone	0.22	U	1	0.22	2.00	ug/kg	11/17/25 22:19	PB170575
7421-93-4	Endrin aldehyde	0.43	U	1	0.43	2.00	ug/kg	11/17/25 22:19	PB170575
5103-71-9	alpha-Chlordane	0.14	U	1	0.14	2.00	ug/kg	11/17/25 22:19	PB170575
5103-74-2	gamma-Chlordane	0.17	U	1	0.17	2.00	ug/kg	11/17/25 22:19	PB170575
8001-35-2	Toxaphene	6.30	U	1	6.30	38.3	ug/kg	11/17/25 22:19	PB170575
SURROGATES									
2051-24-3	Decachlorobiphenyl	16.3			30 (10) - 150 (143)	81%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	19.2			30 (10) - 150 (131)	96%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
 Data File : PD091242.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2025 22:19
 Operator : AR\AJ
 Sample : Q3649-08
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 B3-0.0-0.5-20251114

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 11/18/2025
 Supervised By :mohammad ahmed 11/19/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 01:13:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.541	2.872	65977326	704.6E6	19.131m	19.237
28) SA Decachlor...	9.061	8.057	81091323	582.3E6	16.277	14.713

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091242.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 22:19
Operator : AR\AJ
Sample : Q3649-08
Misc :
ALS Vial : 23 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
B3-0.0-0.5-20251114


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 11/18/2025
Supervised By : mohammad ahmed 11/19/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:13:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.541 min
Delta R.T.: -0.003 min
Response: 65977326
Conc: 19.13 ng/ml

Instrument :
ECD_D
ClientSampleId :
B3-0.0-0.5-20251114


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 11/18/2025
Supervised By :mohammad ahmed 11/19/2025




#1 Tetrachloro-m-xylene

R.T.: 2.872 min
Delta R.T.: -0.002 min
Response: 704645244
Conc: 19.24 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.061 min
Delta R.T.: -0.002 min
Response: 81091323
Conc: 16.28 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.057 min
Delta R.T.: -0.002 min
Response: 582269683
Conc: 14.71 ng/ml



CALIBRATION SUMMARY

Calibration Times: **20:56** **21:51**

LAB FILE ID:	RT 100 =	<u>PD091196.D</u>	RT 075 =	<u>PD091197.D</u>
	RT 050 =	PD091198.D	RT 025 =	PD091199.D
			RT 005 =	PD091200.D

[illegible]

Calibration Times: **20:56** **21:51**

LAB FILE ID:	RT 100 =	<u>PD091196.D</u>	RT 075 =	<u>PD091197.D</u>
	RT 050 =	PD091198.D	RT 025 =	PD091199.D
			RT 005 =	PD091200.D

[illegible]



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_D

Contract: HDRI08
SDG NO.: Q3649

Calibration Date(s): 11/14/2025 11/14/2025
Calibration Times: 20:56 21:51

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD091196.D</u>	CF 075 =	<u>PD091197.D</u>		
CF 050 =		<u>PD091198.D</u>	CF 025 =	<u>PD091199.D</u>	CF 005 =	<u>PD091200.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	4592720000	4750410000	4706670000	4762630000	5745110000	4911510000	10
4,4'-DDE	5537410000	5724420000	5594750000	5596930000	6543500000	5799400000	7
4,4'-DDT	4033900000	4103640000	4023370000	3973800000	4542840000	4135510000	6
Aldrin	6913630000	7108360000	6994820000	7033330000	8231840000	7256400000	8
alpha-BHC	7991490000	8120730000	7870330000	7613270000	8381750000	7995510000	4
alpha-Chlordane	6168980000	6320470000	6262210000	6530520000	9131880000	6882810000	18
beta-BHC	2584470000	2701070000	2721620000	2850470000	3669860000	2905500000	15
Decachlorobiphenyl	4266550000	4520270000	4633000000	4951710000	6537810000	4981870000	18
delta-BHC	7343040000	7477970000	7279360000	7082560000	7977500000	7432080000	5
Dieldrin	6112350000	6323390000	6220240000	6264150000	7404460000	6464920000	8
Endosulfan I	5622190000	5847020000	5809210000	5991120000	7445430000	6142990000	12
Endosulfan II	5022110000	5221970000	5253890000	5419940000	6878740000	5559330000	14
Endosulfan sulfate	4635330000	4824080000	4867360000	5023730000	6373440000	5144790000	14
Endrin	4763010000	4861760000	4775160000	4762720000	5654790000	4963490000	8
Endrin aldehyde	3642590000	3820110000	3882150000	4087850000	5397480000	4166040000	17
Endrin ketone	5131250000	5317460000	5324710000	5507230000	6887120000	5633560000	13
gamma-BHC (Lindane)	7349570000	7492470000	7318000000	7162850000	8041460000	7472870000	5
gamma-Chlordane	6123460000	6343120000	6211230000	6313670000	7512390000	6500770000	9
Heptachlor	6873030000	7051630000	6922840000	6902860000	8058350000	7161740000	7
Heptachlor epoxide	6045880000	6198590000	6197710000	6327850000	7713630000	6496730000	11
Methoxychlor	1967810000	2033910000	2048060000	2143440000	2678460000	2174340000	13
Tetrachloro-m-xylene	3100550000	3224630000	3262160000	3454950000	4200860000	3448630000	13



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_D

Contract: HDRI08
SDG NO.: Q3649

Calibration Date(s): 11/14/2025 11/14/2025
Calibration Times: 20:56 21:51

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD091196.D</u>	CF 075 =	<u>PD091197.D</u>		
CF 050 =	<u>PD091198.D</u>	CF 025 =	<u>PD091199.D</u>	CF 005 =	<u>PD091200.D</u>		
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	42169800000	44652200000	46098100000	48914200000	62508300000	48868500000	16
4,4'-DDE	46784900000	49153600000	50413800000	53041500000	67816900000	53442100000	16
4,4'-DDT	36565400000	37844700000	38225900000	38748200000	44278500000	39132500000	8
Aldrin	49756900000	52224800000	53642000000	56183400000	70485000000	56458400000	14
alpha-BHC	57081100000	59401300000	60005600000	62092800000	77433200000	63202800000	13
alpha-Chlordane	46162300000	48496400000	49694800000	52042000000	65261400000	52331400000	14
beta-BHC	20855800000	21963600000	22629700000	23880400000	30805800000	24027100000	16
Decachlorobiphenyl	35391500000	36530600000	37125400000	38755300000	50071600000	39574900000	15
delta-BHC	51951800000	54520500000	55333700000	57523500000	73056200000	58477100000	14
Dieldrin	47342900000	49974500000	51581700000	54390600000	68265900000	54311100000	15
Endosulfan I	41219300000	43810000000	45334200000	48114600000	60781000000	47851800000	16
Endosulfan II	40233600000	42630400000	44155200000	46739900000	59639900000	46679800000	16
Endosulfan sulfate	38807500000	40991800000	42555300000	44718700000	57101400000	44834900000	16
Endrin	39723100000	41747800000	43010100000	44994600000	56487600000	45192700000	15
Endrin aldehyde	29904600000	31738200000	33159100000	35293100000	46350200000	35289100000	18
Endrin ketone	43743400000	46381200000	48213200000	51142400000	65299300000	50955900000	17
gamma-BHC (Lindane)	52036700000	54159800000	54955200000	56993600000	70961400000	57821300000	13
gamma-Chlordane	48541600000	50822000000	51970500000	54222400000	67580500000	54627400000	14
Heptachlor	48452800000	51096600000	52341200000	55061200000	69095300000	55209400000	15
Heptachlor epoxide	43794700000	46344600000	48042100000	50883900000	65728100000	50958700000	17
Methoxychlor	17875100000	18763500000	19368600000	20213700000	24036200000	20051400000	12
Tetrachloro-m-xylene	32274400000	33875900000	34577100000	36418900000	46004800000	36630200000	15



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** HDRI08
Lab Code: ACE **SDG NO.:** Q3649
Instrument ID: ECD_D **Date(s) Analyzed:** 11/14/2025 11/14/2025
GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.23	6.13	6.33	42812100
		2	6.44	6.34	6.54	68979500
		3	7.14	7.04	7.24	112624000
		4	7.56	7.46	7.66	130712000
		5	7.92	7.82	8.02	79291700



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** HDRI08
Lab Code: ACE **SDG NO.:** Q3649
Instrument ID: ECD_D **Date(s) Analyzed:** 11/14/2025 11/14/2025
GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.46	5.36	5.56	348471000
		2	5.64	5.54	5.74	231072000
		3	6.74	6.64	6.84	1091020000
		4	7.19	7.09	7.29	705058000
		5	7.32	7.22	7.42	442502000

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091196.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2025 20:56
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:13:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:12:56 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	310.1E6	3227.4E6	95.046	93.341
28)	SA Decachlor...	9.063	8.059	426.7E6	3539.1E6	92.090	95.330
Target Compounds							
2)	A alpha-BHC	3.993	3.385	799.1E6	5708.1E6	101.539	95.126
3)	MA gamma-BHC...	4.324	3.721	735.0E6	5203.7E6	100.431	94.689
4)	MA Heptachlor	4.921	4.073	687.3E6	4845.3E6	99.281	92.571
5)	MB Aldrin	5.262	4.358	691.4E6	4975.7E6	98.839	92.757
6)	B beta-BHC	4.511	4.017	258.4E6	2085.6E6	94.961	92.161
7)	B delta-BHC	4.759	4.253	734.3E6	5195.2E6	100.875	93.888
8)	B Heptachlo...	5.683	4.862	604.6E6	4379.5E6	97.550	91.159
9)	A Endosulfan I	6.066	5.236	562.2E6	4121.9E6	96.781	90.923
10)	B gamma-Chl...	5.938	5.114	612.3E6	4854.2E6	98.587	93.402
11)	B alpha-Chl...	6.019	5.179	616.9E6	4616.2E6	98.511	92.892
12)	B 4,4'-DDE	6.189	5.363	553.7E6	4678.5E6	98.975	92.802
13)	MA Dieldrin	6.339	5.501	611.2E6	4734.3E6	98.265	91.782
14)	MA Endrin	6.566	5.778	476.3E6	3972.3E6	99.746	92.358
15)	B Endosulfa...	6.779	6.069	502.2E6	4023.4E6	95.589	91.118
16)	A 4,4'-DDD	6.699	5.918	459.3E6	4217.0E6	97.579	91.478
17)	MA 4,4'-DDT	7.014	6.171	403.4E6	3656.5E6	100.262	95.656
18)	B Endrin al...	6.908	6.248	364.3E6	2990.5E6	93.829	90.185
19)	B Endosulfa...	7.143	6.471	463.5E6	3880.7E6	95.233	91.193
20)	A Methoxychlor	7.487	6.742	196.8E6	1787.5E6	96.081	92.289
21)	B Endrin ke...	7.624	6.980	513.1E6	4374.3E6	96.367	90.729
22)	Mirex	8.106	7.172	302.9E6	2859.2E6	92.228	91.815



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091196.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 20:56
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 54 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:13:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:12:56 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091197.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2025 21:10
 Operator : AR\AJ
 Sample : PSTDICC075
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC075

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:14:14 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:12:56 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	241.8E6	2540.7E6	74.137	73.479
28)	SA Decachlor...	9.064	8.059	339.0E6	2739.8E6	73.175	73.798
Target Compounds							
2)	A alpha-BHC	3.993	3.385	609.1E6	4455.1E6	77.386	74.245
3)	MA gamma-BHC...	4.324	3.721	561.9E6	4062.0E6	76.788	73.915
4)	MA Heptachlor	4.921	4.073	528.9E6	3832.2E6	76.395	73.217
5)	MB Aldrin	5.263	4.358	533.1E6	3916.9E6	76.217	73.019
6)	B beta-BHC	4.511	4.017	202.6E6	1647.3E6	74.433	72.793
7)	B delta-BHC	4.760	4.253	560.8E6	4089.0E6	77.046	73.898
8)	B Heptachlo...	5.683	4.862	464.9E6	3475.8E6	75.011	72.350
9)	A Endosulfan I	6.067	5.236	438.5E6	3285.7E6	75.488	72.478
10)	B gamma-Chl...	5.939	5.114	475.7E6	3811.6E6	76.593	73.343
11)	B alpha-Chl...	6.019	5.179	474.0E6	3637.2E6	75.698	73.191
12)	B 4,4'-DDE	6.189	5.364	429.3E6	3686.5E6	76.738	73.125
13)	MA Dieldrin	6.339	5.501	474.3E6	3748.1E6	76.244	72.663
14)	MA Endrin	6.567	5.778	364.6E6	3131.1E6	76.360	72.799
15)	B Endosulfa...	6.780	6.070	391.6E6	3197.3E6	74.544	72.410
16)	A 4,4'-DDD	6.699	5.918	356.3E6	3348.9E6	75.697	72.648
17)	MA 4,4'-DDT	7.015	6.172	307.8E6	2838.4E6	76.496	74.252
18)	B Endrin al...	6.909	6.248	286.5E6	2380.4E6	73.801	71.786
19)	B Endosulfa...	7.144	6.472	361.8E6	3074.4E6	74.333	72.245
20)	A Methoxychlor	7.488	6.742	152.5E6	1407.3E6	74.482	72.657
21)	B Endrin ke...	7.625	6.980	398.8E6	3478.6E6	74.898	72.150
22)	Mirex	8.107	7.173	239.3E6	2256.0E6	72.881	72.445



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091197.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 21:10
Operator : AR\AJ
Sample : PSTDICC075
Misc :
ALS Vial : 55 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:14:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:12:56 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091198.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2025 21:24
 Operator : AR\AJ
 Sample : PSTDICC050
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:14:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:12:56 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	163.1E6	1728.9E6	50.000	50.000
28)	SA Decachlor...	9.063	8.058	231.7E6	1856.3E6	50.000	50.000
Target Compounds							
2)	A alpha-BHC	3.993	3.386	393.5E6	3000.3E6	50.000	50.000
3)	MA gamma-BHC...	4.324	3.721	365.9E6	2747.8E6	50.000	50.000
4)	MA Heptachlor	4.921	4.073	346.1E6	2617.1E6	50.000	50.000
5)	MB Aldrin	5.263	4.358	349.7E6	2682.1E6	50.000	50.000
6)	B beta-BHC	4.511	4.018	136.1E6	1131.5E6	50.000	50.000
7)	B delta-BHC	4.760	4.253	364.0E6	2766.7E6	50.000	50.000
8)	B Heptachlo...	5.682	4.862	309.9E6	2402.1E6	50.000	50.000
9)	A Endosulfan I	6.066	5.235	290.5E6	2266.7E6	50.000	50.000
10)	B gamma-Chl...	5.938	5.114	310.6E6	2598.5E6	50.000	50.000
11)	B alpha-Chl...	6.019	5.179	313.1E6	2484.7E6	50.000	50.000
12)	B 4,4'-DDE	6.189	5.363	279.7E6	2520.7E6	50.000	50.000
13)	MA Dieldrin	6.339	5.501	311.0E6	2579.1E6	50.000	50.000
14)	MA Endrin	6.566	5.778	238.8E6	2150.5E6	50.000	50.000
15)	B Endosulfa...	6.779	6.069	262.7E6	2207.8E6	50.000	50.000
16)	A 4,4'-DDD	6.699	5.917	235.3E6	2304.9E6	50.000	50.000
17)	MA 4,4'-DDT	7.014	6.171	201.2E6	1911.3E6	50.000	50.000
18)	B Endrin al...	6.908	6.248	194.1E6	1658.0E6	50.000	50.000
19)	B Endosulfa...	7.143	6.471	243.4E6	2127.8E6	50.000	50.000
20)	A Methoxychlor	7.487	6.741	102.4E6	968.4E6	50.000	50.000
21)	B Endrin ke...	7.624	6.980	266.2E6	2410.7E6	50.000	50.000
22)	Mirex	8.106	7.173	164.2E6	1557.0E6	50.000	50.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091198.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 21:24
Operator : AR\AJ
Sample : PSTDICC050
Misc :
ALS Vial : 56 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:14:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:12:56 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091199.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2025 21:37
 Operator : AR\AJ
 Sample : PSTDICC025
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:15:02 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:12:56 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	86373657	910.5E6	26.477	26.332
28)	SA Decachlor...	9.065	8.060	123.8E6	968.9E6	26.720	26.098
Target Compounds							
2)	A alpha-BHC	3.993	3.385	190.3E6	1552.3E6	24.183	25.870
3)	MA gamma-BHC...	4.324	3.721	179.1E6	1424.8E6	24.470	25.927
4)	MA Heptachlor	4.921	4.073	172.6E6	1376.5E6	24.928	26.299
5)	MB Aldrin	5.263	4.358	175.8E6	1404.6E6	25.138	26.184
6)	B beta-BHC	4.511	4.018	71261837	597.0E6	26.184	26.382
7)	B delta-BHC	4.759	4.253	177.1E6	1438.1E6	24.324	25.989
8)	B Heptachlo...	5.683	4.862	158.2E6	1272.1E6	25.525	26.479
9)	A Endosulfan I	6.066	5.236	149.8E6	1202.9E6	25.783	26.533
10)	B gamma-Chl...	5.938	5.115	157.8E6	1355.6E6	25.412	26.083
11)	B alpha-Chl...	6.019	5.179	163.3E6	1301.1E6	26.071	26.181
12)	B 4,4'-DDE	6.189	5.363	139.9E6	1326.0E6	25.010	26.303
13)	MA Dieldrin	6.340	5.502	156.6E6	1359.8E6	25.176	26.361
14)	MA Endrin	6.567	5.778	119.1E6	1124.9E6	24.935	26.153
15)	B Endosulfa...	6.780	6.070	135.5E6	1168.5E6	25.790	26.463
16)	A 4,4'-DDD	6.699	5.918	119.1E6	1222.9E6	25.297	26.527
17)	MA 4,4'-DDT	7.015	6.172	99345097	968.7E6	24.692	25.342
18)	B Endrin al...	6.909	6.248	102.2E6	882.3E6	26.325	26.609
19)	B Endosulfa...	7.144	6.472	125.6E6	1118.0E6	25.803	26.271
20)	A Methoxychlor	7.488	6.743	53585924	505.3E6	26.164	26.091
21)	B Endrin ke...	7.624	6.981	137.7E6	1278.6E6	25.857	26.519
22)	Mirex	8.106	7.173	88650675	829.2E6	26.994	26.629



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091199.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 21:37
Operator : AR\AJ
Sample : PSTDICC025
Misc :
ALS Vial : 57 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:15:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:12:56 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091200.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2025 21:51
 Operator : AR\AJ
 Sample : PSTDICC005
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC005

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 11/17/2025
 Supervised By : Sohil Jodhani 11/17/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:15:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:12:56 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	2.872	21004307	230.0E6	6.439m	6.652m
28)	SA Decachlor...	9.063	8.059	32689037	250.4E6	7.056	6.744
Target Compounds							
2)	A alpha-BHC	3.992	3.385	41908732	387.2E6	5.325	6.452
3)	MA gamma-BHC...	4.323	3.721	40207275	354.8E6	5.494	6.456
4)	MA Heptachlor	4.919	4.073	40291748	345.5E6	5.820m	6.600
5)	MB Aldrin	5.262	4.358	41159200	352.4E6	5.884	6.570
6)	B beta-BHC	4.511	4.017	18349309	154.0E6	6.742	6.806
7)	B delta-BHC	4.757	4.253	39887495	365.3E6	5.480m	6.601
8)	B Heptachlo...	5.681	4.862	38568147	328.6E6	6.223m	6.841
9)	A Endosulfan I	6.067	5.236	37227156	303.9E6	6.408	6.704
10)	B gamma-Chl...	5.939	5.114	37561958	337.9E6	6.047	6.502
11)	B alpha-Chl...	6.019	5.179	45659405	326.3E6	7.291	6.566
12)	B 4,4'-DDE	6.188	5.363	32717510	339.1E6	5.848	6.726
13)	MA Dieldrin	6.339	5.502	37022298	341.3E6	5.952	6.617
14)	MA Endrin	6.566	5.778	28273962	282.4E6	5.921	6.567
15)	B Endosulfa...	6.780	6.070	34393712	298.2E6	6.546	6.753
16)	A 4,4'-DDD	6.699	5.918	28725548	312.5E6	6.103	6.780
17)	MA 4,4'-DDT	7.014	6.172	22714196	221.4E6	5.646	5.792
18)	B Endrin al...	6.908	6.248	26987398	231.8E6	6.952	6.989
19)	B Endosulfa...	7.143	6.471	31867215	285.5E6	6.547	6.709
20)	A Methoxychlor	7.487	6.742	13392299	120.2E6	6.539	6.205
21)	B Endrin ke...	7.624	6.980	34435592	326.5E6	6.467	6.772
22)	Mirex	8.105	7.173	23674832	213.6E6	7.209	6.859

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091200.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 21:51
Operator : AR\AJ
Sample : PSTDICC005
Misc :
ALS Vial : 58 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDICC005



Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 11/17/2025
Supervised By :Sohil Jodhani 11/17/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:15:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:12:56 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091203.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2025 22:32
 Operator : AR\AJ
 Sample : PCHLORICC500
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PCHLORICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:29:43 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:28:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.543	2.873	143.9E6	1902.7E6	50.000	50.000
28)	SA Decachlor...	9.063	8.058	209.3E6	1703.5E6	50.000	50.000
Target Compounds							
23)	Chlordane-1	4.707	3.896	113.3E6	788.6E6	500.000	500.000
24)	Chlordane-2	5.233	4.476	110.4E6	798.4E6	500.000	500.000
25)	Chlordane-3	5.939	5.114	448.7E6	2625.7E6	500.000	500.000
26)	Chlordane-4	6.024	5.178	539.1E6	2177.3E6	500.000	500.000
27)	Chlordane-5	6.864	6.078	88781325	938.4E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091203.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 22:32
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 61 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:29:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:28:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091208.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2025 23:40
 Operator : AR\AJ
 Sample : PTOXICC500
 Misc :
 ALS Vial : 66 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PTOXICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 05:12:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 05:11:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.543	2.874	163.5E6	1762.4E6	50.000	50.000
7)	SA Decachlor...	9.063	8.058	236.5E6	1944.6E6	50.000	50.000
Target Compounds							
2)	Toxaphene-1	6.234	5.463	21406071	174.2E6	500.000	500.000
3)	Toxaphene-2	6.435	5.635	34489739	115.5E6	500.000	500.000
4)	Toxaphene-3	7.141	6.743	56311870	545.5E6	500.000	500.000
5)	Toxaphene-4	7.556	7.186	65356075	352.5E6	500.000	500.000
6)	Toxaphene-5	7.922	7.317	39645850	221.3E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091208.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 23:40
Operator : AR\AJ
Sample : PTOXICC500
Misc :
ALS Vial : 66 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 05:12:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 05:11:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091211.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Nov 2025 00:21
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 69 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD111525

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:49:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	162.5E6	1718.4E6	47.132	46.911
28)	SA Decachlor...	9.063	8.058	232.3E6	1905.9E6	46.634	48.160
Target Compounds							
2)	A alpha-BHC	3.993	3.385	390.8E6	2984.9E6	48.872	47.228
3)	MA gamma-BHC...	4.324	3.721	361.9E6	2728.5E6	48.435	47.188
4)	MA Heptachlor	4.921	4.073	342.0E6	2581.9E6	47.755	46.765
5)	MB Aldrin	5.262	4.358	348.3E6	2666.1E6	47.995	47.222
6)	B beta-BHC	4.511	4.017	135.0E6	1124.4E6	46.461	46.799
7)	B delta-BHC	4.760	4.253	363.6E6	2741.9E6	48.917	46.888
8)	B Heptachlo...	5.683	4.862	313.5E6	2381.7E6	48.259	46.737
9)	A Endosulfan I	6.067	5.235	288.8E6	2250.9E6	47.010	47.039
10)	B gamma-Chl...	5.938	5.114	308.9E6	2585.4E6	47.516	47.328
11)	B alpha-Chl...	6.019	5.179	316.9E6	2460.3E6	46.047	47.013
12)	B 4,4'-DDE	6.188	5.363	276.9E6	2496.0E6	47.745	46.704
13)	MA Dieldrin	6.339	5.501	309.0E6	2552.1E6	47.803	46.991
14)	MA Endrin	6.566	5.778	234.4E6	2114.9E6	47.229	46.798
15)	B Endosulfa...	6.779	6.070	260.8E6	2189.8E6	46.913	46.910
16)	A 4,4'-DDD	6.698	5.917	235.1E6	2292.3E6	47.858	46.907
17)	MA 4,4'-DDT	7.014	6.171	197.9E6	1882.1E6	47.844	48.095
18)	B Endrin al...	6.908	6.248	193.5E6	1652.0E6	46.452	46.813
19)	B Endosulfa...	7.143	6.471	241.2E6	2113.4E6	46.888	47.137
20)	A Methoxychlor	7.487	6.742	101.6E6	965.3E6	46.720	48.141
21)	B Endrin ke...	7.623	6.980	268.5E6	2432.9E6	47.664	47.746
22)	Mirex	8.105	7.172	165.8E6	1568.2E6	46.599	47.322



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091211.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Nov 2025 00:21
Operator : AR\AJ
Sample : PSTDICV050
Misc :
ALS Vial : 69 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
ICVPD111525

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:49:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091212.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Nov 2025 00:49
 Operator : AR\AJ
 Sample : PCHLORICV050
 Misc :
 ALS Vial : 71 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD111525CHLOR

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:41:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:40:47 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	144.0E6	1897.0E6	45.917	45.899
28)	SA Decachlor...	9.063	8.059	208.1E6	1722.5E6	44.473	45.857
Target Compounds							
23)	Chlordane-1	4.707	3.896	114.2E6	783.3E6	464.144	455.309
24)	Chlordane-2	5.234	4.477	110.1E6	794.6E6	445.736	453.786
25)	Chlordane-3	5.939	5.115	447.1E6	2621.5E6	463.407	455.626
26)	Chlordane-4	6.024	5.179	540.3E6	2175.2E6	449.814	454.716
27)	Chlordane-5	6.863	6.079	88911954	936.5E6	450.023	461.495



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091212.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Nov 2025 00:49
Operator : AR\AJ
Sample : PCHLORICV050
Misc :
ALS Vial : 71 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
ICVPD111525CHLOR

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:41:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:40:47 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091213.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Nov 2025 01:30
 Operator : AR\AJ
 Sample : PTOXICV050
 Misc :
 ALS Vial : 74 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD111525TOX

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 05:21:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 05:17:12 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds						
1) SA Tetrachlo...	3.544	2.874	163.1E6	1761.6E6	49.939	50.116
7) SA Decachlor...	9.063	8.059	235.5E6	1961.7E6	49.446	50.407
Target Compounds						
2) Toxaphene-1	6.234	5.464	21366491	177.4E6	506.837	508.980
3) Toxaphene-2	6.435	5.635	34816331	114.1E6	493.754	505.708
4) Toxaphene-3	7.141	6.744	56004022	537.8E6	503.381	513.961
5) Toxaphene-4	7.557	7.186	66555590	352.7E6	512.569	521.193
6) Toxaphene-5	7.923	7.317	39466192	222.5E6	507.805	531.712

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091213.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Nov 2025 01:30
Operator : AR\AJ
Sample : PTOXICV050
Misc :
ALS Vial : 74 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 05:21:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 05:17:12 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3649

Continuing Calib Date: 11/17/2025

Initial Calibration Date(s): 11/14/2025

11/14/2025

Continuing Calib Time: 10:18

Initial Calibration Time(s): 20:56

21:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.07	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
alpha-BHC	3.99	3.99	3.89	4.09	0.00
beta-BHC	4.51	4.51	4.41	4.61	0.00
delta-BHC	4.76	4.76	4.66	4.86	0.00
gamma-BHC (Lindane)	4.32	4.32	4.22	4.42	0.00
Heptachlor	4.92	4.92	4.82	5.02	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.78	6.78	6.68	6.88	0.00
4,4'-DDD	6.70	6.70	6.60	6.80	0.00
Endosulfan sulfate	7.15	7.14	7.04	7.24	0.00
4,4'-DDT	7.02	7.01	6.91	7.11	-0.01
Methoxychlor	7.49	7.49	7.39	7.59	0.00
Endrin ketone	7.63	7.62	7.52	7.72	-0.01
Endrin aldehyde	6.91	6.91	6.81	7.01	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3649

Continuing Calib Date: 11/17/2025

Initial Calibration Date(s): 11/14/2025

11/14/2025

Continuing Calib Time: 10:18

Initial Calibration Time(s): 20:56

21:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.87	2.77	2.97	0.00
alpha-BHC	3.39	3.39	3.29	3.49	0.01
beta-BHC	4.02	4.02	3.92	4.12	0.00
delta-BHC	4.25	4.25	4.15	4.35	0.00
gamma-BHC (Lindane)	3.72	3.72	3.62	3.82	0.00
Heptachlor	4.07	4.07	3.97	4.17	0.00
Aldrin	4.36	4.36	4.26	4.46	0.00
Heptachlor epoxide	4.86	4.86	4.76	4.96	0.00
Endosulfan I	5.24	5.24	5.14	5.34	0.00
Dieldrin	5.50	5.50	5.40	5.60	0.00
4,4'-DDE	5.36	5.36	5.26	5.46	0.00
Endrin	5.78	5.78	5.68	5.88	0.00
Endosulfan II	6.07	6.07	5.97	6.17	0.00
4,4'-DDD	5.92	5.92	5.82	6.02	0.00
Endosulfan sulfate	6.47	6.47	6.37	6.57	0.00
4,4'-DDT	6.17	6.17	6.07	6.27	0.00
Methoxychlor	6.74	6.74	6.64	6.84	0.00
Endrin ketone	6.98	6.98	6.88	7.08	0.00
Endrin aldehyde	6.25	6.25	6.15	6.35	0.00
alpha-Chlordane	5.18	5.18	5.08	5.28	0.00
gamma-Chlordane	5.12	5.11	5.01	5.21	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** HDRI08
Lab Code: ACE **SDG NO.:** Q3649
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/17/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091220.D **Time Analyzed:** 10:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.700	6.599	6.799	46.860	50.000	-6.3
4,4'-DDE	6.189	6.089	6.289	48.040	50.000	-3.9
4,4'-DDT	7.016	6.914	7.114	48.850	50.000	-2.3
Aldrin	5.263	5.163	5.363	48.120	50.000	-3.8
alpha-BHC	3.993	3.893	4.093	49.860	50.000	-0.3
alpha-Chlordane	6.020	5.919	6.119	45.510	50.000	-9.0
beta-BHC	4.511	4.411	4.611	47.530	50.000	-4.9
Decachlorobiphenyl	9.065	8.963	9.163	45.920	50.000	-8.2
delta-BHC	4.760	4.660	4.860	49.340	50.000	-1.3
Dieldrin	6.340	6.239	6.439	47.780	50.000	-4.4
Endosulfan I	6.067	5.966	6.166	47.150	50.000	-5.7
Endosulfan II	6.781	6.679	6.879	46.250	50.000	-7.5
Endosulfan sulfate	7.145	7.043	7.243	46.450	50.000	-7.1
Endrin	6.567	6.466	6.666	47.230	50.000	-5.5
Endrin aldehyde	6.910	6.808	7.008	46.310	50.000	-7.4
Endrin ketone	7.626	7.524	7.724	47.510	50.000	-5.0
gamma-BHC (Lindane)	4.324	4.224	4.424	49.610	50.000	-0.8
gamma-Chlordane	5.939	5.838	6.038	47.710	50.000	-4.6
Heptachlor	4.922	4.821	5.021	48.940	50.000	-2.1
Heptachlor epoxide	5.684	5.582	5.782	47.620	50.000	-4.8
Methoxychlor	7.488	7.387	7.587	47.290	50.000	-5.4
Tetrachloro-m-xylene	3.544	3.444	3.644	48.000	50.000	-4.0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Contract: HDRI08

Lab Code: ACE SDG NO.: Q3649

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No.: CCAL01 Date Analyzed: 11/17/2025

Lab Sample No.: PSTDCCC050 Data File : PD091220.D Time Analyzed: 10:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.918	5.817	6.017	49.260	50.000	-1.5
4,4'-DDE	5.363	5.263	5.463	51.190	50.000	2.4
4,4'-DDT	6.172	6.071	6.271	52.440	50.000	4.9
Aldrin	4.358	4.258	4.458	52.180	50.000	4.4
alpha-BHC	3.385	3.286	3.486	52.350	50.000	4.7
alpha-Chlordane	5.179	5.079	5.279	51.710	50.000	3.4
beta-BHC	4.017	3.918	4.118	51.600	50.000	3.2
Decachlorobiphenyl	8.059	7.958	8.158	48.700	50.000	-2.6
delta-BHC	4.254	4.153	4.353	51.680	50.000	3.4
Dieldrin	5.502	5.401	5.601	51.270	50.000	2.5
Endosulfan I	5.236	5.135	5.335	51.230	50.000	2.5
Endosulfan II	6.070	5.969	6.169	50.200	50.000	0.4
Endosulfan sulfate	6.472	6.371	6.571	50.000	50.000	0.0
Endrin	5.779	5.678	5.878	50.140	50.000	0.3
Endrin aldehyde	6.249	6.148	6.348	49.880	50.000	-0.2
Endrin ketone	6.981	6.880	7.080	49.990	50.000	0.0
gamma-BHC (Lindane)	3.721	3.621	3.821	52.280	50.000	4.6
gamma-Chlordane	5.115	5.014	5.214	52.060	50.000	4.1
Heptachlor	4.073	3.973	4.173	51.470	50.000	2.9
Heptachlor epoxide	4.862	4.762	4.962	51.490	50.000	3.0
Methoxychlor	6.743	6.641	6.841	49.060	50.000	-1.9
Tetrachloro-m-xylene	2.873	2.774	2.974	51.820	50.000	3.6

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
 Data File : PD091220.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2025 10:18
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 01:08:05 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.873	165.5E6	1898.3E6	47.999	51.823
28)	SA Decachlor...	9.065	8.059	228.8E6	1927.4E6	45.917	48.702
Target Compounds							
2)	A alpha-BHC	3.993	3.385	398.7E6	3308.5E6	49.860	52.348
3)	MA gamma-BHC...	4.324	3.721	370.7E6	3023.0E6	49.608	52.282
4)	MA Heptachlor	4.922	4.073	350.5E6	2841.7E6	48.938	51.471
5)	MB Aldrin	5.263	4.358	349.2E6	2946.0E6	48.119	52.180
6)	B beta-BHC	4.511	4.017	138.1E6	1239.7E6	47.527	51.595
7)	B delta-BHC	4.760	4.254	366.7E6	3022.3E6	49.343	51.684
8)	B Heptachlo...	5.684	4.862	309.4E6	2623.9E6	47.620	51.491
9)	A Endosulfan I	6.067	5.236	289.7E6	2451.5E6	47.154	51.230
10)	B gamma-Chl...	5.939	5.115	310.1E6	2843.8E6	47.706	52.057
11)	B alpha-Chl...	6.020	5.179	313.2E6	2706.0E6	45.505	51.709
12)	B 4,4'-DDE	6.189	5.363	278.6E6	2735.9E6	48.036	51.194
13)	MA Dieldrin	6.340	5.502	308.9E6	2784.6E6	47.782	51.272
14)	MA Endrin	6.567	5.779	234.4E6	2265.9E6	47.232	50.139
15)	B Endosulfa...	6.781	6.070	257.1E6	2343.2E6	46.249	50.197
16)	A 4,4'-DDD	6.700	5.918	230.1E6	2407.4E6	46.858	49.263
17)	MA 4,4'-DDT	7.016	6.172	202.0E6	2052.3E6	48.845	52.445
18)	B Endrin al...	6.910	6.249	192.9E6	1760.0E6	46.315	49.875
19)	B Endosulfa...	7.145	6.472	239.0E6	2241.5E6	46.454	49.995
20)	A Methoxychlor	7.488	6.743	102.8E6	983.7E6	47.288	49.061
21)	B Endrin ke...	7.626	6.981	267.6E6	2547.5E6	47.505	49.994
22)	Mirex	8.108	7.173	165.1E6	1635.1E6	46.416	49.340



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091220.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 10:18
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:08:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3649

Continuing Calib Date: 11/17/2025

Initial Calibration Date(s): 11/14/2025

11/14/2025

Continuing Calib Time: 20:02

Initial Calibration Time(s): 20:56

21:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
alpha-BHC	3.99	3.99	3.89	4.09	0.00
beta-BHC	4.51	4.51	4.41	4.61	0.00
delta-BHC	4.76	4.76	4.66	4.86	0.00
gamma-BHC (Lindane)	4.32	4.32	4.22	4.42	0.00
Heptachlor	4.92	4.92	4.82	5.02	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.78	6.78	6.68	6.88	0.00
4,4'-DDD	6.70	6.70	6.60	6.80	0.00
Endosulfan sulfate	7.14	7.14	7.04	7.24	0.00
4,4'-DDT	7.01	7.01	6.91	7.11	0.00
Methoxychlor	7.49	7.49	7.39	7.59	0.00
Endrin ketone	7.62	7.62	7.52	7.72	0.00
Endrin aldehyde	6.91	6.91	6.81	7.01	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3649

Continuing Calib Date: 11/17/2025

Initial Calibration Date(s): 11/14/2025

11/14/2025

Continuing Calib Time: 20:02

Initial Calibration Time(s): 20:56

21:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.87	2.77	2.97	0.00
alpha-BHC	3.38	3.39	3.29	3.49	0.01
beta-BHC	4.02	4.02	3.92	4.12	0.01
delta-BHC	4.25	4.25	4.15	4.35	0.00
gamma-BHC (Lindane)	3.72	3.72	3.62	3.82	0.00
Heptachlor	4.07	4.07	3.97	4.17	0.00
Aldrin	4.36	4.36	4.26	4.46	0.00
Heptachlor epoxide	4.86	4.86	4.76	4.96	0.00
Endosulfan I	5.23	5.24	5.14	5.34	0.01
Dieldrin	5.50	5.50	5.40	5.60	0.00
4,4'-DDE	5.36	5.36	5.26	5.46	0.00
Endrin	5.78	5.78	5.68	5.88	0.00
Endosulfan II	6.07	6.07	5.97	6.17	0.00
4,4'-DDD	5.92	5.92	5.82	6.02	0.00
Endosulfan sulfate	6.47	6.47	6.37	6.57	0.00
4,4'-DDT	6.17	6.17	6.07	6.27	0.00
Methoxychlor	6.74	6.74	6.64	6.84	0.00
Endrin ketone	6.98	6.98	6.88	7.08	0.00
Endrin aldehyde	6.25	6.25	6.15	6.35	0.00
alpha-Chlordane	5.18	5.18	5.08	5.28	0.00
gamma-Chlordane	5.11	5.11	5.01	5.21	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Contract: HDRI08

Lab Code: ACE SDG NO.: Q3649

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No.: CCAL02 Date Analyzed: 11/17/2025

Lab Sample No.: PSTDCCC050 Data File : PD091236.D Time Analyzed: 20:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.697	6.599	6.799	44.500	50.000	-11.0
4,4'-DDE	6.187	6.089	6.289	49.510	50.000	-1.0
4,4'-DDT	7.013	6.914	7.114	58.790	50.000	17.6
Aldrin	5.261	5.163	5.363	48.730	50.000	-2.5
alpha-BHC	3.991	3.893	4.093	50.370	50.000	0.7
alpha-Chlordane	6.017	5.919	6.119	46.460	50.000	-7.1
beta-BHC	4.509	4.411	4.611	47.900	50.000	-4.2
Decachlorobiphenyl	9.062	8.963	9.163	46.500	50.000	-7.0
delta-BHC	4.757	4.660	4.860	45.560	50.000	-8.9
Dieldrin	6.337	6.239	6.439	48.980	50.000	-2.0
Endosulfan I	6.065	5.966	6.166	48.360	50.000	-3.3
Endosulfan II	6.778	6.679	6.879	49.660	50.000	-0.7
Endosulfan sulfate	7.142	7.043	7.243	47.340	50.000	-5.3
Endrin	6.565	6.466	6.666	54.020	50.000	8.0
Endrin aldehyde	6.907	6.808	7.008	46.460	50.000	-7.1
Endrin ketone	7.623	7.524	7.724	45.130	50.000	-9.7
gamma-BHC (Lindane)	4.322	4.224	4.424	50.160	50.000	0.3
gamma-Chlordane	5.937	5.838	6.038	48.750	50.000	-2.5
Heptachlor	4.919	4.821	5.021	51.520	50.000	3.0
Heptachlor epoxide	5.681	5.582	5.782	48.060	50.000	-3.9
Methoxychlor	7.486	7.387	7.587	55.420	50.000	10.8
Tetrachloro-m-xylene	3.542	3.444	3.644	48.160	50.000	-3.7



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** HDRI08
Lab Code: ACE **SDG NO.:** Q3649
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/17/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091236.D **Time Analyzed:** 20:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.916	5.817	6.017	45.100	50.000	-9.8
4,4'-DDE	5.362	5.263	5.463	49.970	50.000	-0.1
4,4'-DDT	6.169	6.071	6.271	59.320	50.000	18.6
Aldrin	4.356	4.258	4.458	49.970	50.000	-0.1
alpha-BHC	3.383	3.286	3.486	50.530	50.000	1.1
alpha-Chlordane	5.178	5.079	5.279	50.170	50.000	0.3
beta-BHC	4.015	3.918	4.118	48.950	50.000	-2.1
Decachlorobiphenyl	8.057	7.958	8.158	44.480	50.000	-11.0
delta-BHC	4.252	4.153	4.353	49.960	50.000	-0.1
Dieldrin	5.500	5.401	5.601	49.880	50.000	-0.2
Endosulfan I	5.234	5.135	5.335	49.550	50.000	-0.9
Endosulfan II	6.068	5.969	6.169	48.100	50.000	-3.8
Endosulfan sulfate	6.470	6.371	6.571	48.370	50.000	-3.3
Endrin	5.776	5.678	5.878	55.250	50.000	10.5
Endrin aldehyde	6.246	6.148	6.348	47.470	50.000	-5.1
Endrin ketone	6.978	6.880	7.080	45.020	50.000	-10.0
gamma-BHC (Lindane)	3.719	3.621	3.821	49.620	50.000	-0.8
gamma-Chlordane	5.113	5.014	5.214	50.420	50.000	0.8
Heptachlor	4.071	3.973	4.173	52.080	50.000	4.2
Heptachlor epoxide	4.860	4.762	4.962	49.300	50.000	-1.4
Methoxychlor	6.741	6.641	6.841	56.900	50.000	13.8
Tetrachloro-m-xylene	2.872	2.774	2.974	49.870	50.000	-0.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
 Data File : PD091236.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2025 20:02
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 01:11:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	2.872	166.1E6	1826.6E6	48.160	49.865
2)	SA Decachlor...	9.062	8.057	231.7E6	1760.3E6	46.499	44.481
Target Compounds							
2)	A alpha-BHC	3.991	3.383	402.8E6	3193.7E6	50.374	50.531
3)	MA gamma-BHC...	4.322	3.719	374.8E6	2869.0E6	50.158	49.618
4)	MA Heptachlor	4.919	4.071	369.0E6	2875.3E6	51.519	52.080
5)	MB Aldrin	5.261	4.356	353.6E6	2821.5E6	48.729	49.974
6)	B beta-BHC	4.509	4.015	139.2E6	1176.1E6	47.904	48.948
7)	B delta-BHC	4.757	4.252	338.6E6	2921.4E6	45.560	49.958
8)	B Heptachlo...	5.681	4.860	312.2E6	2512.4E6	48.057	49.303
9)	A Endosulfan I	6.065	5.234	297.1E6	2371.0E6	48.362	49.550
10)	B gamma-Chl...	5.937	5.113	316.9E6	2754.3E6	48.753	50.420
11)	B alpha-Chl...	6.017	5.178	319.8E6	2625.3E6	46.465	50.167
12)	B 4,4'-DDE	6.187	5.362	287.1E6	2670.7E6	49.505	49.973
13)	MA Dieldrin	6.337	5.500	316.6E6	2709.1E6	48.976	49.880
14)	MA Endrin	6.565	5.776	268.1E6	2497.0E6	54.017	55.253
15)	B Endosulfa...	6.778	6.068	276.1E6	2245.3E6	49.658	48.100
16)	A 4,4'-DDD	6.697	5.916	218.6E6	2203.8E6	44.504	45.096
17)	MA 4,4'-DDT	7.013	6.169	243.1E6	2321.3E6	58.785	59.318
18)	B Endrin al...	6.907	6.246	193.5E6	1675.1E6	46.458	47.468
19)	B Endosulfa...	7.142	6.470	243.6E6	2168.6E6	47.340	48.369
20)	A Methoxychlor	7.486	6.741	120.5E6	1140.9E6	55.416	56.901
21)	B Endrin ke...	7.623	6.978	254.2E6	2294.0E6	45.127	45.020
22)	Mirex	8.105	7.171	162.8E6	1485.3E6	45.776	44.818



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091236.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 20:02
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:11:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3649

Continuing Calib Date: 11/17/2025

Initial Calibration Date(s): 11/14/2025

11/14/2025

Continuing Calib Time: 23:00

Initial Calibration Time(s): 20:56

21:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
alpha-BHC	3.99	3.99	3.89	4.09	0.00
beta-BHC	4.51	4.51	4.41	4.61	0.00
delta-BHC	4.76	4.76	4.66	4.86	0.00
gamma-BHC (Lindane)	4.32	4.32	4.22	4.42	0.00
Heptachlor	4.92	4.92	4.82	5.02	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.78	6.78	6.68	6.88	0.00
4,4'-DDD	6.70	6.70	6.60	6.80	0.00
Endosulfan sulfate	7.14	7.14	7.04	7.24	0.00
4,4'-DDT	7.01	7.01	6.91	7.11	0.00
Methoxychlor	7.49	7.49	7.39	7.59	0.00
Endrin ketone	7.62	7.62	7.52	7.72	0.00
Endrin aldehyde	6.91	6.91	6.81	7.01	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3649

Continuing Calib Date: 11/17/2025

Initial Calibration Date(s): 11/14/2025

11/14/2025

Continuing Calib Time: 23:00

Initial Calibration Time(s): 20:56

21:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.87	2.77	2.97	0.00
alpha-BHC	3.38	3.39	3.29	3.49	0.01
beta-BHC	4.02	4.02	3.92	4.12	0.00
delta-BHC	4.25	4.25	4.15	4.35	0.00
gamma-BHC (Lindane)	3.72	3.72	3.62	3.82	0.00
Heptachlor	4.07	4.07	3.97	4.17	0.00
Aldrin	4.36	4.36	4.26	4.46	0.00
Heptachlor epoxide	4.86	4.86	4.76	4.96	0.00
Endosulfan I	5.23	5.24	5.14	5.34	0.01
Dieldrin	5.50	5.50	5.40	5.60	0.00
4,4'-DDE	5.36	5.36	5.26	5.46	0.00
Endrin	5.78	5.78	5.68	5.88	0.00
Endosulfan II	6.07	6.07	5.97	6.17	0.00
4,4'-DDD	5.92	5.92	5.82	6.02	0.00
Endosulfan sulfate	6.47	6.47	6.37	6.57	0.00
4,4'-DDT	6.17	6.17	6.07	6.27	0.00
Methoxychlor	6.74	6.74	6.64	6.84	0.00
Endrin ketone	6.98	6.98	6.88	7.08	0.00
Endrin aldehyde	6.25	6.25	6.15	6.35	0.00
alpha-Chlordane	5.18	5.18	5.08	5.28	0.00
gamma-Chlordane	5.11	5.11	5.01	5.21	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Contract: HDRI08

Lab Code: ACE SDG NO.: Q3649

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No.: CCAL03 Date Analyzed: 11/17/2025

Lab Sample No.: PSTDCCC050 Data File : PD091244.D Time Analyzed: 23:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.697	6.599	6.799	42.880	50.000	-14.2
4,4'-DDE	6.187	6.089	6.289	46.130	50.000	-7.7
4,4'-DDT	7.013	6.914	7.114	49.770	50.000	-0.5
Aldrin	5.261	5.163	5.363	46.240	50.000	-7.5
alpha-BHC	3.991	3.893	4.093	48.140	50.000	-3.7
alpha-Chlordane	6.017	5.919	6.119	43.510	50.000	-13.0
beta-BHC	4.509	4.411	4.611	45.990	50.000	-8.0
Decachlorobiphenyl	9.062	8.963	9.163	43.560	50.000	-12.9
delta-BHC	4.758	4.660	4.860	46.150	50.000	-7.7
Dieldrin	6.337	6.239	6.439	45.910	50.000	-8.2
Endosulfan I	6.065	5.966	6.166	45.230	50.000	-9.5
Endosulfan II	6.778	6.679	6.879	47.520	50.000	-5.0
Endosulfan sulfate	7.142	7.043	7.243	44.020	50.000	-12.0
Endrin	6.565	6.466	6.666	50.530	50.000	1.1
Endrin aldehyde	6.907	6.808	7.008	43.780	50.000	-12.4
Endrin ketone	7.622	7.524	7.724	41.250	50.000	-17.5
gamma-BHC (Lindane)	4.322	4.224	4.424	47.370	50.000	-5.3
gamma-Chlordane	5.937	5.838	6.038	46.090	50.000	-7.8
Heptachlor	4.919	4.821	5.021	45.720	50.000	-8.6
Heptachlor epoxide	5.681	5.582	5.782	46.110	50.000	-7.8
Methoxychlor	7.486	7.387	7.587	46.530	50.000	-6.9
Tetrachloro-m-xylene	3.542	3.444	3.644	46.160	50.000	-7.7



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** HDRI08
Lab Code: ACE **SDG NO.:** Q3649
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/17/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091244.D **Time Analyzed:** 23:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.916	5.817	6.017	43.200	50.000	-13.6
4,4'-DDE	5.361	5.263	5.463	47.610	50.000	-4.8
4,4'-DDT	6.170	6.071	6.271	50.400	50.000	0.8
Aldrin	4.356	4.258	4.458	47.980	50.000	-4.0
alpha-BHC	3.384	3.286	3.486	48.310	50.000	-3.4
alpha-Chlordane	5.177	5.079	5.279	47.460	50.000	-5.1
beta-BHC	4.016	3.918	4.118	47.640	50.000	-4.7
Decachlorobiphenyl	8.057	7.958	8.158	40.840	50.000	-18.3
delta-BHC	4.251	4.153	4.353	48.090	50.000	-3.8
Dieldrin	5.500	5.401	5.601	47.090	50.000	-5.8
Endosulfan I	5.234	5.135	5.335	46.000	50.000	-8.0
Endosulfan II	6.068	5.969	6.169	45.850	50.000	-8.3
Endosulfan sulfate	6.470	6.371	6.571	45.380	50.000	-9.2
Endrin	5.776	5.678	5.878	51.450	50.000	2.9
Endrin aldehyde	6.246	6.148	6.348	44.780	50.000	-10.4
Endrin ketone	6.979	6.880	7.080	40.500	50.000	-19.0
gamma-BHC (Lindane)	3.719	3.621	3.821	47.730	50.000	-4.5
gamma-Chlordane	5.113	5.014	5.214	47.910	50.000	-4.2
Heptachlor	4.071	3.973	4.173	46.880	50.000	-6.2
Heptachlor epoxide	4.860	4.762	4.962	47.080	50.000	-5.8
Methoxychlor	6.741	6.641	6.841	47.560	50.000	-4.9
Tetrachloro-m-xylene	2.872	2.774	2.974	48.030	50.000	-3.9

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
 Data File : PD091244.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2025 23:00
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 01:14:15 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	2.872	159.2E6	1759.2E6	46.162	48.026
2)	SA Decachlor...	9.062	8.057	217.0E6	1616.4E6	43.564	40.844
Target Compounds							
2)	A alpha-BHC	3.991	3.384	384.9E6	3053.6E6	48.136	48.314
3)	MA gamma-BHC...	4.322	3.719	354.0E6	2759.9E6	47.367	47.732
4)	MA Heptachlor	4.919	4.071	327.4E6	2588.2E6	45.721	46.879
5)	MB Aldrin	5.261	4.356	335.6E6	2708.8E6	46.243	47.978
6)	B beta-BHC	4.509	4.016	133.6E6	1144.7E6	45.992	47.640
7)	B delta-BHC	4.758	4.251	343.0E6	2812.4E6	46.147	48.094
8)	B Heptachlo...	5.681	4.860	299.6E6	2399.3E6	46.110	47.082
9)	A Endosulfan I	6.065	5.234	277.8E6	2201.2E6	45.229	45.999
10)	B gamma-Chl...	5.937	5.113	299.6E6	2617.1E6	46.087	47.909
11)	B alpha-Chl...	6.017	5.177	299.5E6	2483.7E6	43.509	47.460
12)	B 4,4'-DDE	6.187	5.361	267.5E6	2544.1E6	46.130	47.605
13)	MA Dieldrin	6.337	5.500	296.8E6	2557.4E6	45.907	47.089
14)	MA Endrin	6.565	5.776	250.8E6	2325.3E6	50.533	51.452
15)	B Endosulfa...	6.778	6.068	264.2E6	2140.2E6	47.525	45.849
16)	A 4,4'-DDD	6.697	5.916	210.6E6	2111.2E6	42.882	43.202
17)	MA 4,4'-DDT	7.013	6.170	205.8E6	1972.1E6	49.767	50.397
18)	B Endrin al...	6.907	6.246	182.4E6	1580.1E6	43.783	44.777
19)	B Endosulfa...	7.142	6.470	226.5E6	2034.7E6	44.016	45.382
20)	A Methoxychlor	7.486	6.741	101.2E6	953.6E6	46.528	47.556
21)	B Endrin ke...	7.622	6.979	232.4E6	2063.7E6	41.251	40.501
22)	Mirex	8.104	7.171	147.8E6	1388.3E6	41.559	41.891



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091244.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 23:00
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:14:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3649

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No. (PEM): PEM - PD091194.D Date Analyzed: 11/14/2025

Lab Sample No.(PEM): PEM Time Analyzed: 20:15

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	9.064	8.960	9.160	20.170	20.000	0.9
Tetrachloro-m-xylene	3.544	3.490	3.590	19.500	20.000	-2.5
alpha-BHC	3.993	3.940	4.040	8.250	10.000	-17.5
beta-BHC	4.511	4.460	4.560	9.240	10.000	-7.6
gamma-BHC (Lindane)	4.324	4.270	4.370	8.450	10.000	-15.5
Endrin	6.567	6.500	6.640	44.390	50.000	-11.2
4,4'-DDT	7.015	6.940	7.090	88.290	100.000	-11.7
Methoxychlor	7.487	7.420	7.560	207.040	250.000	-17.2

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No. (PEM): PEM - PD091194.D Date Analyzed: 11/14/2025

Lab Sample No.(PEM): PEM Time Analyzed: 20:15

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	8.059	7.960	8.160	19.620	20.000	-1.9
Tetrachloro-m-xylene	2.874	2.820	2.920	19.370	20.000	-3.2
alpha-BHC	3.385	3.330	3.440	9.380	10.000	-6.2
beta-BHC	4.018	3.970	4.070	9.530	10.000	-4.7
gamma-BHC (Lindane)	3.721	3.670	3.770	9.460	10.000	-5.4
Endrin	5.779	5.710	5.850	45.450	50.000	-9.1
4,4'-DDT	6.172	6.100	6.240	89.340	100.000	-10.7
Methoxychlor	6.742	6.670	6.810	186.630	250.000	-25.3

Data File: PEM
 PD091194.D Date Acquired 11/14/2025 20:15
 Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down Down
Endrin	6.57	220342338.4	243232929.5	22890591.1	9.41
Endrin aldehyde	6.91	5179237.826			
Endrin ketone	7.62	17711353.28			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.78	2053826744	2298027068	244200325	10.63
Endrin aldehyde #2	6.25	61443615.46			
Endrin ketone #2	6.98	182756709.2			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	7.01	365114982.7	402071544.1	36956561.3	9.19
4,4'-DDE	6.19	1068437.233			
4,4'-DDD	6.70	35888124.09			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.17	3496021709	3894922521	398900812	10.24
4,4'-DDE #2	5.36	8351296.931			
4,4'-DDD #2	5.92	390549515.4			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091194.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2025 20:15
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 11/17/2025
 Supervised By :Sohil Jodhani 11/17/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:48:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	67260764	709.4E6	19.504	19.367
28)	SA Decachlor...	9.064	8.059	100.5E6	776.5E6	20.172	19.621
Target Compounds							
2)	A alpha-BHC	3.993	3.385	65992882	592.9E6	8.254	9.381
3)	MA gamma-BHC...	4.324	3.721	63139958	546.9E6	8.449	9.458
6)	B beta-BHC	4.511	4.018	26858838	229.1E6	9.244	9.533
12)	B 4,4'-DDE	6.188	5.362	1068437	8351297	0.184	0.156m
14)	MA Endrin	6.567	5.779	220.3E6	2053.8E6	44.393	45.446
16)	A 4,4'-DDD	6.698	5.918	35888124	390.5E6	7.307	7.992
17)	MA 4,4'-DDT	7.015	6.172	365.1E6	3496.0E6	88.288	89.338
18)	B Endrin al...	6.908	6.246	5179238	61443615	1.243m	1.741m#
20)	A Methoxychlor	7.487	6.742	450.2E6	3742.1E6	207.044	186.626
21)	B Endrin ke...	7.624	6.980	17711353	182.8E6	3.144	3.587

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091194.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 20:15
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 11/17/2025
Supervised By : Sohil Jodhani 11/17/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:48:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3649

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No. (PEM): PEM - PD091219.D Date Analyzed: 11/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:04

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.067	8.970	9.170	20.390	20.000	2.0
Tetrachloro-m-xylene	3.544	3.490	3.590	21.040	20.000	5.2
alpha-BHC	3.993	3.940	4.040	8.920	10.000	-10.8
beta-BHC	4.511	4.460	4.560	9.940	10.000	-0.6
gamma-BHC (Lindane)	4.324	4.270	4.370	9.120	10.000	-8.8
Endrin	6.567	6.500	6.640	44.720	50.000	-10.6
4,4'-DDT	7.016	6.950	7.090	91.520	100.000	-8.5
Methoxychlor	7.489	7.420	7.560	211.580	250.000	-15.4

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No. (PEM): PEM - PD091219.D Date Analyzed: 11/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:04

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.060	7.960	8.160	20.690	20.000	3.5
Tetrachloro-m-xylene	2.873	2.820	2.920	22.610	20.000	13.1
alpha-BHC	3.385	3.330	3.440	11.010	10.000	10.1
beta-BHC	4.017	3.970	4.070	11.240	10.000	12.4
gamma-BHC (Lindane)	3.721	3.670	3.770	11.030	10.000	10.3
Endrin	5.777	5.710	5.850	47.210	50.000	-5.6
4,4'-DDT	6.171	6.100	6.240	98.490	100.000	-1.5
Methoxychlor	6.741	6.670	6.810	191.760	250.000	-23.3

PEM
 Data File: PD091219.D Date Acquired 11/17/2025 10:04
 Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin	6.57	221986255.2	249073127.5	27086872.3	10.88
Endrin aldehyde	6.91	7568925.994			
Endrin ketone	7.63	19517946.28			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.78	2133505788	2452875834	319370046	13.02
Endrin aldehyde #2	6.25	84362913.82			
Endrin ketone #2	6.98	235007132.6			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	7.02	378465526.5	414877012.8	36411486.4	8.78
4,4'-DDE	6.19	1587124.769			
4,4'-DDD	6.70	34824361.6			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.17	3854065509	4284063898	429998389	10.04
4,4'-DDE #2	5.36	18405683.02			
4,4'-DDD #2	5.92	411592706.3			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
 Data File : PD091219.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2025 10:04
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 01:07:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.873	72576102	828.3E6	21.045	22.611
28)	SA Decachlor...	9.067	8.060	101.6E6	818.9E6	20.391	20.693
Target Compounds							
2)	A alpha-BHC	3.993	3.385	71357424	696.0E6	8.925	11.012
3)	MA gamma-BHC...	4.324	3.721	68153737	637.8E6	9.120	11.031
6)	B beta-BHC	4.511	4.017	28881769	270.2E6	9.940	11.245
12)	B 4,4'-DDE	6.189	5.364	1587125	18405683	0.274	0.344 #
14)	MA Endrin	6.567	5.777	222.0E6	2133.5E6	44.724	47.209
16)	A 4,4'-DDD	6.700	5.918	34824362	411.6E6	7.090	8.422
17)	MA 4,4'-DDT	7.016	6.171	378.5E6	3854.1E6	91.516	98.487
18)	B Endrin al...	6.910	6.247	7568926	84362914	1.817	2.391 #
20)	A Methoxychlor	7.489	6.741	460.1E6	3845.0E6	211.583	191.757
21)	B Endrin ke...	7.626	6.980	19517946	235.0E6	3.465	4.612 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091219.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 10:04
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PEM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:07:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3649

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No. (PEM): PEM - PD091235.D Date Analyzed: 11/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 19:35

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.062	8.960	9.160	16.950	20.000	-15.3
Tetrachloro-m-xylene	3.542	3.490	3.590	17.020	20.000	-14.9
alpha-BHC	3.991	3.940	4.040	7.060	10.000	-29.4
beta-BHC	4.509	4.460	4.560	8.040	10.000	-19.6
gamma-BHC (Lindane)	4.321	4.270	4.370	7.320	10.000	-26.8
Endrin	6.565	6.490	6.640	41.850	50.000	-16.3
4,4'-DDT	7.013	6.940	7.080	90.270	100.000	-9.7
Methoxychlor	7.486	7.420	7.560	208.670	250.000	-16.5

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/14/2025 11/14/2025

Client Sample No. (PEM): PEM - PD091235.D Date Analyzed: 11/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 19:35

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.058	7.960	8.160	15.820	20.000	-20.9
Tetrachloro-m-xylene	2.872	2.820	2.920	17.620	20.000	-11.9
alpha-BHC	3.383	3.330	3.430	8.510	10.000	-14.9
beta-BHC	4.016	3.970	4.070	8.510	10.000	-14.9
gamma-BHC (Lindane)	3.720	3.670	3.770	8.370	10.000	-16.3
Endrin	5.777	5.710	5.850	43.790	50.000	-12.4
4,4'-DDT	6.171	6.100	6.240	92.490	100.000	-7.5
Methoxychlor	6.741	6.670	6.810	190.580	250.000	-23.8

Data File: PEM
 PD091235.D **Date Acquired** 11/17/2025 19:35
Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down Down
Endrin	6.57	207733142.9	210297653.3	2564510.4	1.22
Endrin aldehyde	6.91	351396.179			
Endrin ketone	7.62	2213114.222			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.78	1978977858	2030162859	51185001	2.52
Endrin aldehyde #2	6.25	25683721.57			
Endrin ketone #2	6.98	25501279.45			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	7.01	373317035.2	375897812.8	2580777.6	0.69
4,4'-DDE	0.00	0			
4,4'-DDD	6.70	2580777.601			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.17	3619447981	3644691080	25243098.2	0.69
4,4'-DDE #2	0.00	0			
4,4'-DDD #2	5.92	25243098.15			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
 Data File : PD091235.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2025 19:35
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 11/18/2025
 Supervised By :mohammad ahmed 11/19/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 01:11:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	2.872	58679095	645.5E6	17.015	17.622
28)	SA Decachlor...	9.062	8.058	84440758	626.0E6	16.950	15.818
Target Compounds							
2)	A alpha-BHC	3.991	3.383	56484336	537.8E6	7.065	8.509
3)	MA gamma-BHC...	4.321	3.720	54679034	483.9E6	7.317	8.369
6)	B beta-BHC	4.509	4.016	23362439	204.4E6	8.041	8.509
14)	MA Endrin	6.565	5.777	207.7E6	1979.0E6	41.852	43.790
16)	A 4,4'-DDD	6.698	5.917	2580778	25243098	0.525	0.517m
17)	MA 4,4'-DDT	7.013	6.171	373.3E6	3619.4E6	90.271	92.492
18)	B Endrin al...	6.911	6.245	351396	25683722	0.084m	0.728 #
20)	A Methoxychlor	7.486	6.741	453.7E6	3821.3E6	208.670	190.576
21)	B Endrin ke...	7.623	6.977	2213114	25501279	0.393	0.500m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091235.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 19:35
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 11/18/2025
Supervised By :mohammad ahmed 11/19/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:11:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091195.D
Acq On : 14 Nov 2025 20:43
Operator : AR\AJ
Sample : RESCHK
Misc :
ALS Vial : 51 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e

Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Title : GC Extractables
Last Update : Sat Nov 15 04:46:13 2025
Integrator: ChemStation

RT#1	RT#2	Resolution
3.544	5.939	100.00%
5.939	6.067	100.00%
6.067	6.189	100.00%
6.189	6.339	100.00%
6.339	7.143	100.00%
7.143	7.487	100.00%
7.487	7.624	100.00%
7.624	9.063	100.00%

Signal #2

2.874	5.114	100.00%
5.114	5.235	100.00%
5.235	5.363	100.00%
5.363	5.501	100.00%
5.501	6.471	100.00%
6.471	6.742	100.00%
6.742	6.980	100.00%
6.980	8.059	100.00%

PD111525.M Sat Nov 15 05:00:31 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
 Data File : PD091195.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2025 20:43
 Operator : AR\AJ
 Sample : RESCHK
 Misc :
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 RESCHK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 04:48:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	59572378	631.9E6	17.274	17.250
28)	SA Decachlor...	9.063	8.059	86552368	667.6E6	17.373	16.869
Target Compounds							
9)	A Endosulfan I	6.067	5.235	45261732	379.7E6	7.368	7.934
10)	B gamma-Chl...	5.939	5.114	49362700	455.8E6	7.593	8.344
12)	B 4,4'-DDE	6.189	5.363	87544052	881.3E6	15.095	16.491
13)	MA Dieldrin	6.339	5.501	96362551	874.4E6	14.905	16.100
19)	B Endosulfa...	7.143	6.471	78876117	723.4E6	15.331	16.134
20)	A Methoxychlor	7.487	6.742	154.5E6	1426.0E6	71.049	71.120
21)	B Endrin ke...	7.624	6.980	80568543	766.5E6	14.302	15.043



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091195.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 20:43
Operator : AR\AJ
Sample : RESCHK
Misc :
ALS Vial : 51 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
RESCHK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:48:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Analytical Sequence

Client: HDR, Inc.

SDG No.: Q3649

Project: PVWC Linear Construction

Instrument ID: ECD_D

GC Column: ZB-MR1

ID: 0.32 (mm)

Inst. Calib. Date(s): 11/14/2025 11/14/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	11/14/2025	20:01	PD091193.D	9.06	3.54
PEM	PEM	11/14/2025	20:15	PD091194.D	9.06	3.54
RESCHK	RESCHK	11/14/2025	20:43	PD091195.D	9.06	3.54
PSTDICC100	PSTDICC100	11/14/2025	20:56	PD091196.D	9.06	3.54
PSTDICC075	PSTDICC075	11/14/2025	21:10	PD091197.D	9.06	3.54
PSTDICC050	PSTDICC050	11/14/2025	21:24	PD091198.D	9.06	3.54
PSTDICC025	PSTDICC025	11/14/2025	21:37	PD091199.D	9.07	3.54
PSTDICC005	PSTDICC005	11/14/2025	21:51	PD091200.D	9.06	3.54
PCHLORICC500	PCHLORICC500	11/14/2025	22:32	PD091203.D	9.06	3.54
PTOXICC500	PTOXICC500	11/14/2025	23:40	PD091208.D	9.06	3.54
IBLK	IBLK	11/17/2025	09:51	PD091218.D	9.07	3.55
PEM	PEM	11/17/2025	10:04	PD091219.D	9.07	3.54
PSTDCCC050	PSTDCCC050	11/17/2025	10:18	PD091220.D	9.07	3.54
PB170575BL	PB170575BL	11/17/2025	13:46	PD091223.D	9.07	3.54
PB170575BS	PB170575BS	11/17/2025	14:00	PD091224.D	9.07	3.54
IBLK	IBLK	11/17/2025	19:21	PD091234.D	9.06	3.54
PEM	PEM	11/17/2025	19:35	PD091235.D	9.06	3.54
PSTDCCC050	PSTDCCC050	11/17/2025	20:02	PD091236.D	9.06	3.54
B5-0.0-0.5-20251114	Q3649-01	11/17/2025	21:11	PD091237.D	9.06	3.54
DUPE-0.0-0.5-20251114	Q3649-03	11/17/2025	21:24	PD091238.D	9.06	3.54
DUPE-0.0-0.5-20251114MS	Q3649-03MS	11/17/2025	21:38	PD091239.D	9.06	3.54
DUPE-0.0-0.5-20251114MSD	Q3649-03MSD	11/17/2025	21:52	PD091240.D	9.06	3.54
B4-0.0-0.5-20251114	Q3649-06	11/17/2025	22:05	PD091241.D	9.06	3.54
B3-0.0-0.5-20251114	Q3649-08	11/17/2025	22:19	PD091242.D	9.06	3.54
IBLK	IBLK	11/17/2025	22:33	PD091243.D	9.06	3.54
PSTDCCC050	PSTDCCC050	11/17/2025	23:00	PD091244.D	9.06	3.54

Analytical Sequence

Client: HDR, Inc.

SDG No.: Q3649

Project: PVWC Linear Construction

Instrument ID: ECD_D

GC Column: ZB-MR2

ID: 0.32 (mm)

Inst. Calib. Date(s): 11/14/2025

11/14/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
LBLK	LBLK	11/14/2025	20:01	PD091193.D	8.06	2.88
PEM	PEM	11/14/2025	20:15	PD091194.D	8.06	2.87
RESCHK	RESCHK	11/14/2025	20:43	PD091195.D	8.06	2.87
PSTDICC100	PSTDICC100	11/14/2025	20:56	PD091196.D	8.06	2.87
PSTDICC075	PSTDICC075	11/14/2025	21:10	PD091197.D	8.06	2.87
PSTDICC050	PSTDICC050	11/14/2025	21:24	PD091198.D	8.06	2.87
PSTDICC025	PSTDICC025	11/14/2025	21:37	PD091199.D	8.06	2.87
PSTDICC005	PSTDICC005	11/14/2025	21:51	PD091200.D	8.06	2.87
PCHLORICC500	PCHLORICC500	11/14/2025	22:32	PD091203.D	8.06	2.87
PTOXICC500	PTOXICC500	11/14/2025	23:40	PD091208.D	8.06	2.87
LBLK	LBLK	11/17/2025	09:51	PD091218.D	8.06	2.87
PEM	PEM	11/17/2025	10:04	PD091219.D	8.06	2.87
PSTDCCC050	PSTDCCC050	11/17/2025	10:18	PD091220.D	8.06	2.87
PB170575BL	PB170575BL	11/17/2025	13:46	PD091223.D	8.06	2.87
PB170575BS	PB170575BS	11/17/2025	14:00	PD091224.D	8.06	2.87
LBLK	LBLK	11/17/2025	19:21	PD091234.D	8.06	2.87
PEM	PEM	11/17/2025	19:35	PD091235.D	8.06	2.87
PSTDCCC050	PSTDCCC050	11/17/2025	20:02	PD091236.D	8.06	2.87
B5-0.0-0.5-20251114	Q3649-01	11/17/2025	21:11	PD091237.D	8.06	2.87
DUPE-0.0-0.5-20251114	Q3649-03	11/17/2025	21:24	PD091238.D	8.07	2.89
DUPE-0.0-0.5-20251114MS	Q3649-03MS	11/17/2025	21:38	PD091239.D	8.06	2.87
DUPE-0.0-0.5-20251114MSD	Q3649-03MSD	11/17/2025	21:52	PD091240.D	8.06	2.87
B4-0.0-0.5-20251114	Q3649-06	11/17/2025	22:05	PD091241.D	8.06	2.87
B3-0.0-0.5-20251114	Q3649-08	11/17/2025	22:19	PD091242.D	8.06	2.87
LBLK	LBLK	11/17/2025	22:33	PD091243.D	8.06	2.87
PSTDCCC050	PSTDCCC050	11/17/2025	23:00	PD091244.D	8.06	2.87

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

DUPE-0.0-0.5-20251114

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3649

Lab Sample ID: Q3649-03

Date(s) Analyzed: 11/17/2025 11/17/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDT	1	7.01	6.96	7.06	1.50	53.7
	2	6.18	6.13	6.23	2.60	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

DUPE-0.0-0.5-20251114MS

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3649

Lab Sample ID: Q3649-03MS

Date(s) Analyzed: 11/17/2025 11/17/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan II	1	6.78	6.73	6.83	16.9	6.1
	2	6.07	6.02	6.12	15.9	
4,4'-DDD	1	6.70	6.65	6.75	15.8	22.5
	2	5.92	5.87	5.97	12.6	
4,4'-DDT	1	7.01	6.96	7.06	21.4	7.8
	2	6.17	6.12	6.22	19.8	
Endrin aldehyde	1	6.91	6.86	6.96	16.7	10.1
	2	6.25	6.20	6.30	15.1	
Endosulfan sulfate	1	7.14	7.09	7.19	15.9	7.2
	2	6.47	6.42	6.52	14.8	
Methoxychlor	1	7.49	7.44	7.54	18.9	20.4
	2	6.74	6.69	6.79	15.4	
Endrin ketone	1	7.62	7.57	7.67	15.5	27.9
	2	6.98	6.93	7.03	11.7	
alpha-BHC	1	3.99	3.94	4.04	17.5	0
	2	3.38	3.33	3.43	17.5	
gamma-BHC (Lindane)	1	4.32	4.27	4.37	17.5	2.9
	2	3.72	3.67	3.77	17.0	
Heptachlor	1	4.92	4.87	4.97	18.2	4.5
	2	4.07	4.02	4.12	17.4	
Aldrin	1	5.26	5.21	5.31	16.7	0
	2	4.36	4.31	4.41	16.7	
beta-BHC	1	4.51	4.46	4.56	16.4	10.4
	2	4.02	3.97	4.07	18.2	
delta-BHC	1	4.76	4.71	4.81	16.7	0.6
	2	4.25	4.20	4.30	16.6	
Heptachlor epoxide	1	5.68	5.63	5.73	16.1	1.8
	2	4.86	4.81	4.91	16.4	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

DUPE-0.0-0.5-20251114MS

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3649

Lab Sample ID: Q3649-03MS

Date(s) Analyzed: 11/17/2025 11/17/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan I	1	6.07	6.02	6.12	16.9	15.3
	2	5.23	5.18	5.28	14.5	
gamma-Chlordane	1	5.94	5.89	5.99	16.2	2.4
	2	5.11	5.06	5.16	16.6	
alpha-Chlordane	1	6.02	5.97	6.07	16.2	4.8
	2	5.18	5.13	5.23	17.0	
4,4'-DDE	1	6.19	6.14	6.24	15.8	1.3
	2	5.36	5.31	5.41	16.0	
Dieldrin	1	6.34	6.29	6.39	24.0	3.8
	2	5.50	5.45	5.55	23.1	
Endrin	1	6.57	6.52	6.62	18.7	7.8
	2	5.78	5.73	5.83	17.3	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

DUPE-0.0-0.5-20251114MSD

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3649

Lab Sample ID: Q3649-03MSD

Date(s) Analyzed: 11/17/2025 11/17/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan II	1	6.78	6.73	6.83	16.3	6.3
	2	6.07	6.02	6.12	15.3	
4,4'-DDD	1	6.70	6.65	6.75	15.3	22.5
	2	5.92	5.87	5.97	12.2	
4,4'-DDT	1	7.01	6.96	7.06	20.0	10
	2	6.17	6.12	6.22	18.1	
Endrin aldehyde	1	6.91	6.86	6.96	15.8	10
	2	6.25	6.20	6.30	14.3	
Endosulfan sulfate	1	7.14	7.09	7.19	15.2	8.2
	2	6.47	6.42	6.52	14.0	
Methoxychlor	1	7.49	7.44	7.54	17.2	19.8
	2	6.74	6.69	6.79	14.1	
Endrin ketone	1	7.62	7.57	7.67	14.8	30.4
	2	6.98	6.93	7.03	10.9	
alpha-BHC	1	3.99	3.94	4.04	16.9	1.8
	2	3.38	3.33	3.43	17.2	
gamma-BHC (Lindane)	1	4.32	4.27	4.37	16.9	1.8
	2	3.72	3.67	3.77	16.6	
Heptachlor	1	4.92	4.87	4.97	16.9	3.6
	2	4.07	4.02	4.12	16.3	
Aldrin	1	5.26	5.21	5.31	16.2	0.6
	2	4.36	4.31	4.41	16.3	
beta-BHC	1	4.51	4.46	4.56	16.5	7.6
	2	4.02	3.97	4.07	17.8	
delta-BHC	1	4.76	4.71	4.81	16.4	1.2
	2	4.25	4.20	4.30	16.2	
Heptachlor epoxide	1	5.68	5.63	5.73	15.7	1.9
	2	4.86	4.81	4.91	16.0	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

DUPE-0.0-0.5-20251114MSD

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3649

Lab Sample ID: Q3649-03MSD

Date(s) Analyzed: 11/17/2025 11/17/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan I	1	6.06	6.01	6.11	16.2	13.9
	2	5.24	5.19	5.29	14.1	
gamma-Chlordane	1	5.94	5.89	5.99	15.7	9.7
	2	5.11	5.06	5.16	17.3	
alpha-Chlordane	1	6.02	5.97	6.07	15.4	6.9
	2	5.18	5.13	5.23	16.5	
4,4'-DDE	1	6.19	6.14	6.24	15.2	1.3
	2	5.36	5.31	5.41	15.4	
Dieldrin	1	6.34	6.29	6.39	23.1	2.6
	2	5.50	5.45	5.55	22.5	
Endrin	1	6.56	6.51	6.61	18.0	8.1
	2	5.78	5.73	5.83	16.6	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB170575BS

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3649

Lab Sample ID: PB170575BS

Date(s) Analyzed: 11/17/2025

11/17/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan I	1	6.07	6.02	6.12	16.0	6.6
	2	5.24	5.19	5.29	17.1	
alpha-Chlordane	1	6.02	5.97	6.07	15.3	12.3
	2	5.18	5.13	5.23	17.3	
4,4'-DDE	1	6.19	6.14	6.24	16.3	4.8
	2	5.36	5.31	5.41	17.1	
Dieldrin	1	6.34	6.29	6.39	16.3	4.8
	2	5.50	5.45	5.55	17.1	
Endrin	1	6.57	6.52	6.62	15.0	3.9
	2	5.78	5.73	5.83	15.6	
Methoxychlor	1	7.49	7.44	7.54	15.6	6.2
	2	6.74	6.69	6.79	16.6	
Endrin ketone	1	7.63	7.58	7.68	16.5	7
	2	6.98	6.93	7.03	17.7	
gamma-BHC (Lindane)	1	4.32	4.27	4.37	16.3	4.8
	2	3.72	3.67	3.77	17.1	
Heptachlor	1	4.92	4.87	4.97	16.3	3.6
	2	4.07	4.02	4.12	16.9	
Heptachlor epoxide	1	5.68	5.63	5.73	16.0	6.1
	2	4.86	4.81	4.91	17.0	
gamma-Chlordane	1	5.94	5.89	5.99	16.3	6
	2	5.12	5.07	5.17	17.3	
Endosulfan II	1	6.78	6.73	6.83	16.1	5.4
	2	6.07	6.02	6.12	17.0	
4,4'-DDD	1	6.70	6.65	6.75	16.0	4.3
	2	5.92	5.87	5.97	16.7	
4,4'-DDT	1	7.02	6.97	7.07	16.5	7.6
	2	6.17	6.12	6.22	17.8	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB170575BS

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3649

Lab Sample ID: PB170575BS

Date(s) Analyzed: 11/17/2025 11/17/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endrin aldehyde	1	6.91	6.86	6.96	16.5	6.5
	2	6.25	6.20	6.30	17.6	
Endosulfan sulfate	1	7.14	7.09	7.19	15.9	6.7
	2	6.47	6.42	6.52	17.0	
alpha-BHC	1	3.99	3.94	4.04	16.3	4.8
	2	3.39	3.34	3.44	17.1	
Aldrin	1	5.26	5.21	5.31	16.2	6
	2	4.36	4.31	4.41	17.2	
beta-BHC	1	4.51	4.46	4.56	15.6	8
	2	4.02	3.97	4.07	16.9	
delta-BHC	1	4.76	4.71	4.81	16.5	3.6
	2	4.25	4.20	4.30	17.1	



QC SAMPLE DATA



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	HDR, Inc.		Date Collected:	
Project:	PVWC Linear Construction		Date Received:	
Client Sample ID:	PB170575BL		SDG No.:	Q3649
Lab Sample ID:	PB170575BL		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	100
Sample Wt/Vol:	30.01 g	Final Vol: 10000 uL	Test:	Pesticide-TCL
Prep Method:	SW3541B	Prep Date: 11/17/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.13	U	1	0.13	1.70	ug/kg	11/17/25 13:46	PB170575
319-85-7	beta-BHC	0.18	U	1	0.18	1.70	ug/kg	11/17/25 13:46	PB170575
319-86-8	delta-BHC	0.39	U	1	0.39	1.70	ug/kg	11/17/25 13:46	PB170575
58-89-9	gamma-BHC (Lindane)	0.14	U	1	0.14	1.70	ug/kg	11/17/25 13:46	PB170575
76-44-8	Heptachlor	0.12	U	1	0.12	1.70	ug/kg	11/17/25 13:46	PB170575
309-00-2	Aldrin	0.12	U	1	0.12	1.70	ug/kg	11/17/25 13:46	PB170575
1024-57-3	Heptachlor epoxide	0.19	U	1	0.19	1.70	ug/kg	11/17/25 13:46	PB170575
959-98-8	Endosulfan I	0.14	U	1	0.14	1.70	ug/kg	11/17/25 13:46	PB170575
60-57-1	Dieldrin	0.14	U	1	0.14	1.70	ug/kg	11/17/25 13:46	PB170575
72-55-9	4,4-DDE	0.14	U	1	0.14	1.70	ug/kg	11/17/25 13:46	PB170575
72-20-8	Endrin	0.14	U	1	0.14	1.70	ug/kg	11/17/25 13:46	PB170575
33213-65-9	Endosulfan II	0.29	U	1	0.29	1.70	ug/kg	11/17/25 13:46	PB170575
72-54-8	4,4-DDD	0.15	U	1	0.15	1.70	ug/kg	11/17/25 13:46	PB170575
1031-07-8	Endosulfan Sulfate	0.13	U	1	0.13	1.70	ug/kg	11/17/25 13:46	PB170575
50-29-3	4,4-DDT	0.14	U	1	0.14	1.70	ug/kg	11/17/25 13:46	PB170575
72-43-5	Methoxychlor	0.37	U	1	0.37	1.70	ug/kg	11/17/25 13:46	PB170575
53494-70-5	Endrin ketone	0.19	U	1	0.19	1.70	ug/kg	11/17/25 13:46	PB170575
7421-93-4	Endrin aldehyde	0.37	U	1	0.37	1.70	ug/kg	11/17/25 13:46	PB170575
5103-71-9	alpha-Chlordane	0.12	U	1	0.12	1.70	ug/kg	11/17/25 13:46	PB170575
5103-74-2	gamma-Chlordane	0.15	U	1	0.15	1.70	ug/kg	11/17/25 13:46	PB170575
8001-35-2	Toxaphene	5.40	U	1	5.40	33.0	ug/kg	11/17/25 13:46	PB170575
SURROGATES									
2051-24-3	Decachlorobiphenyl	19.9			30 (10) - 150 (143)	100%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	19.7			30 (10) - 150 (131)	98%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
 Data File : PD091223.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2025 13:46
 Operator : AR\AJ
 Sample : PB170575BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB170575BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 01:08:44 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	62697450	721.2E6	18.180	19.690
28)	SA Decachlor...	9.067	8.060	95177735	788.8E6	19.105	19.933

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091223.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 13:46
Operator : AR\AJ
Sample : PB170575BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB170575BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:08:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.544 min
Delta R.T.: 0.000 min
Response: 62697450
Conc: 18.18 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB170575BL




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 721236048
Conc: 19.69 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.067 min
Delta R.T.: 0.004 min
Response: 95177735
Conc: 19.10 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.060 min
Delta R.T.: 0.001 min
Response: 788845597
Conc: 19.93 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	HDR, Inc.		Date Collected:	11/14/25
Project:	PVWC Linear Construction		Date Received:	11/14/25
Client Sample ID:	PIBLK-PD091193.D		SDG No.:	Q3649
Lab Sample ID:	I.BLK-PD091193.D		Matrix:	WATER
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	Pesticide-TCL
Prep Method:	3510C	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.0039	U	1	0.0039	0.050	ug/L	11/14/25	PD111525
319-85-7	beta-BHC	0.0049	U	1	0.0049	0.050	ug/L	11/14/25	PD111525
319-86-8	delta-BHC	0.011	U	1	0.011	0.050	ug/L	11/14/25	PD111525
58-89-9	gamma-BHC (Lindane)	0.0037	U	1	0.0037	0.050	ug/L	11/14/25	PD111525
76-44-8	Heptachlor	0.0027	U	1	0.0027	0.050	ug/L	11/14/25	PD111525
309-00-2	Aldrin	0.0036	U	1	0.0036	0.050	ug/L	11/14/25	PD111525
1024-57-3	Heptachlor epoxide	0.0096	U	1	0.0096	0.050	ug/L	11/14/25	PD111525
959-98-8	Endosulfan I	0.0031	U	1	0.0031	0.050	ug/L	11/14/25	PD111525
60-57-1	Dieldrin	0.0036	U	1	0.0036	0.050	ug/L	11/14/25	PD111525
72-55-9	4,4-DDE	0.0037	U	1	0.0037	0.050	ug/L	11/14/25	PD111525
72-20-8	Endrin	0.0032	U	1	0.0032	0.050	ug/L	11/14/25	PD111525
33213-65-9	Endosulfan II	0.0079	U	1	0.0079	0.050	ug/L	11/14/25	PD111525
72-54-8	4,4-DDD	0.0071	U	1	0.0071	0.050	ug/L	11/14/25	PD111525
1031-07-8	Endosulfan Sulfate	0.0037	U	1	0.0037	0.050	ug/L	11/14/25	PD111525
50-29-3	4,4-DDT	0.0035	U	1	0.0035	0.050	ug/L	11/14/25	PD111525
72-43-5	Methoxychlor	0.011	U	1	0.011	0.050	ug/L	11/14/25	PD111525
53494-70-5	Endrin ketone	0.0093	U	1	0.0093	0.050	ug/L	11/14/25	PD111525
7421-93-4	Endrin aldehyde	0.011	U	1	0.011	0.050	ug/L	11/14/25	PD111525
5103-71-9	alpha-Chlordane	0.0035	U	1	0.0035	0.050	ug/L	11/14/25	PD111525
5103-74-2	gamma-Chlordane	0.0039	U	1	0.0039	0.050	ug/L	11/14/25	PD111525
8001-35-2	Toxaphene	0.17	U	1	0.17	1.00	ug/L	11/14/25	PD111525
SURROGATES									
2051-24-3	Decachlorobiphenyl	20.8			30 (31) - 150 (149)	104%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.0			30 (41) - 150 (134)	100%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091193.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 20:01
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:47:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.875	67367446	734.0E6	19.535	20.038
28)	SA Decachlor...	9.064	8.059	103.5E6	793.9E6	20.767	20.060

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111525\
Data File : PD091193.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2025 20:01
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 04:47:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.544 min
Delta R.T.: 0.000 min
Response: 67367446
Conc: 19.53 ng/ml

Instrument :
ECD_D
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.875 min
Delta R.T.: 0.000 min
Response: 734006061
Conc: 20.04 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.064 min
Delta R.T.: 0.000 min
Response: 103459386
Conc: 20.77 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.059 min
Delta R.T.: 0.000 min
Response: 793870994
Conc: 20.06 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	HDR, Inc.		Date Collected:	11/17/25
Project:	PVWC Linear Construction		Date Received:	11/17/25
Client Sample ID:	PIBLK-PD091218.D		SDG No.:	Q3649
Lab Sample ID:	I.BLK-PD091218.D		Matrix:	WATER
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	Pesticide-TCL
Prep Method:	3510C	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.0039	U	1	0.0039	0.050	ug/L	11/17/25	pd111725
319-85-7	beta-BHC	0.0049	U	1	0.0049	0.050	ug/L	11/17/25	pd111725
319-86-8	delta-BHC	0.011	U	1	0.011	0.050	ug/L	11/17/25	pd111725
58-89-9	gamma-BHC (Lindane)	0.0037	U	1	0.0037	0.050	ug/L	11/17/25	pd111725
76-44-8	Heptachlor	0.0027	U	1	0.0027	0.050	ug/L	11/17/25	pd111725
309-00-2	Aldrin	0.0036	U	1	0.0036	0.050	ug/L	11/17/25	pd111725
1024-57-3	Heptachlor epoxide	0.0096	U	1	0.0096	0.050	ug/L	11/17/25	pd111725
959-98-8	Endosulfan I	0.0031	U	1	0.0031	0.050	ug/L	11/17/25	pd111725
60-57-1	Dieldrin	0.0036	U	1	0.0036	0.050	ug/L	11/17/25	pd111725
72-55-9	4,4-DDE	0.0037	U	1	0.0037	0.050	ug/L	11/17/25	pd111725
72-20-8	Endrin	0.0032	U	1	0.0032	0.050	ug/L	11/17/25	pd111725
33213-65-9	Endosulfan II	0.0079	U	1	0.0079	0.050	ug/L	11/17/25	pd111725
72-54-8	4,4-DDD	0.0071	U	1	0.0071	0.050	ug/L	11/17/25	pd111725
1031-07-8	Endosulfan Sulfate	0.0037	U	1	0.0037	0.050	ug/L	11/17/25	pd111725
50-29-3	4,4-DDT	0.0035	U	1	0.0035	0.050	ug/L	11/17/25	pd111725
72-43-5	Methoxychlor	0.011	U	1	0.011	0.050	ug/L	11/17/25	pd111725
53494-70-5	Endrin ketone	0.0093	U	1	0.0093	0.050	ug/L	11/17/25	pd111725
7421-93-4	Endrin aldehyde	0.011	U	1	0.011	0.050	ug/L	11/17/25	pd111725
5103-71-9	alpha-Chlordane	0.0035	U	1	0.0035	0.050	ug/L	11/17/25	pd111725
5103-74-2	gamma-Chlordane	0.0039	U	1	0.0039	0.050	ug/L	11/17/25	pd111725
8001-35-2	Toxaphene	0.17	U	1	0.17	1.00	ug/L	11/17/25	pd111725
SURROGATES									
2051-24-3	Decachlorobiphenyl	20.0			30 (31) - 150 (149)	100%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	22.6			30 (41) - 150 (134)	113%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091218.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 09:51
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:07:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.548	2.873	67500173	827.2E6	19.573	22.584
28)	SA Decachlor...	9.071	8.061	99597433	763.2E6	19.992	19.286

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091218.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 09:51
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:07:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.548 min
Delta R.T.: 0.004 min
Response: 67500173
Conc: 19.57 ng/ml

Instrument :
ECD_D
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.873 min
Delta R.T.: 0.000 min
Response: 827248235
Conc: 22.58 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.071 min
Delta R.T.: 0.007 min
Response: 99597433
Conc: 19.99 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.061 min
Delta R.T.: 0.003 min
Response: 763248619
Conc: 19.29 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	HDR, Inc.		Date Collected:	11/17/25
Project:	PVWC Linear Construction		Date Received:	11/17/25
Client Sample ID:	PIBLK-PD091234.D		SDG No.:	Q3649
Lab Sample ID:	I.BLK-PD091234.D		Matrix:	WATER
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	Pesticide-TCL
Prep Method:	3510C	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.0039	U	1	0.0039	0.050	ug/L	11/17/25	pd111725
319-85-7	beta-BHC	0.0049	U	1	0.0049	0.050	ug/L	11/17/25	pd111725
319-86-8	delta-BHC	0.011	U	1	0.011	0.050	ug/L	11/17/25	pd111725
58-89-9	gamma-BHC (Lindane)	0.0037	U	1	0.0037	0.050	ug/L	11/17/25	pd111725
76-44-8	Heptachlor	0.0027	U	1	0.0027	0.050	ug/L	11/17/25	pd111725
309-00-2	Aldrin	0.0036	U	1	0.0036	0.050	ug/L	11/17/25	pd111725
1024-57-3	Heptachlor epoxide	0.0096	U	1	0.0096	0.050	ug/L	11/17/25	pd111725
959-98-8	Endosulfan I	0.0031	U	1	0.0031	0.050	ug/L	11/17/25	pd111725
60-57-1	Dieldrin	0.0036	U	1	0.0036	0.050	ug/L	11/17/25	pd111725
72-55-9	4,4-DDE	0.0037	U	1	0.0037	0.050	ug/L	11/17/25	pd111725
72-20-8	Endrin	0.0032	U	1	0.0032	0.050	ug/L	11/17/25	pd111725
33213-65-9	Endosulfan II	0.0079	U	1	0.0079	0.050	ug/L	11/17/25	pd111725
72-54-8	4,4-DDD	0.0071	U	1	0.0071	0.050	ug/L	11/17/25	pd111725
1031-07-8	Endosulfan Sulfate	0.0037	U	1	0.0037	0.050	ug/L	11/17/25	pd111725
50-29-3	4,4-DDT	0.0035	U	1	0.0035	0.050	ug/L	11/17/25	pd111725
72-43-5	Methoxychlor	0.011	U	1	0.011	0.050	ug/L	11/17/25	pd111725
53494-70-5	Endrin ketone	0.0093	U	1	0.0093	0.050	ug/L	11/17/25	pd111725
7421-93-4	Endrin aldehyde	0.011	U	1	0.011	0.050	ug/L	11/17/25	pd111725
5103-71-9	alpha-Chlordane	0.0035	U	1	0.0035	0.050	ug/L	11/17/25	pd111725
5103-74-2	gamma-Chlordane	0.0039	U	1	0.0039	0.050	ug/L	11/17/25	pd111725
8001-35-2	Toxaphene	0.17	U	1	0.17	1.00	ug/L	11/17/25	pd111725
SURROGATES									
2051-24-3	Decachlorobiphenyl	20.8			30 (31) - 150 (149)	104%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	21.4			30 (41) - 150 (134)	107%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091234.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 19:21
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:11:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	2.872	68845015	784.5E6	19.963	21.417
28)	SA Decachlor...	9.062	8.057	103.4E6	746.6E6	20.751	18.866

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091234.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 19:21
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:11:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.542 min
Delta R.T.: -0.002 min
Response: 68845015
Conc: 19.96 ng/ml

Instrument :
ECD_D
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.872 min
Delta R.T.: -0.002 min
Response: 784503654
Conc: 21.42 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.062 min
Delta R.T.: -0.001 min
Response: 103377420
Conc: 20.75 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.057 min
Delta R.T.: -0.001 min
Response: 746618568
Conc: 18.87 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	HDR, Inc.		Date Collected:	11/17/25
Project:	PVWC Linear Construction		Date Received:	11/17/25
Client Sample ID:	PIBLK-PD091243.D		SDG No.:	Q3649
Lab Sample ID:	I.BLK-PD091243.D		Matrix:	WATER
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	Pesticide-TCL
Prep Method:	3510C	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	0.0039	U	1	0.0039	0.050	ug/L	11/17/25	pd111725
319-85-7	beta-BHC	0.0049	U	1	0.0049	0.050	ug/L	11/17/25	pd111725
319-86-8	delta-BHC	0.011	U	1	0.011	0.050	ug/L	11/17/25	pd111725
58-89-9	gamma-BHC (Lindane)	0.0037	U	1	0.0037	0.050	ug/L	11/17/25	pd111725
76-44-8	Heptachlor	0.0027	U	1	0.0027	0.050	ug/L	11/17/25	pd111725
309-00-2	Aldrin	0.0036	U	1	0.0036	0.050	ug/L	11/17/25	pd111725
1024-57-3	Heptachlor epoxide	0.0096	U	1	0.0096	0.050	ug/L	11/17/25	pd111725
959-98-8	Endosulfan I	0.0031	U	1	0.0031	0.050	ug/L	11/17/25	pd111725
60-57-1	Dieldrin	0.0036	U	1	0.0036	0.050	ug/L	11/17/25	pd111725
72-55-9	4,4-DDE	0.0037	U	1	0.0037	0.050	ug/L	11/17/25	pd111725
72-20-8	Endrin	0.0032	U	1	0.0032	0.050	ug/L	11/17/25	pd111725
33213-65-9	Endosulfan II	0.0079	U	1	0.0079	0.050	ug/L	11/17/25	pd111725
72-54-8	4,4-DDD	0.0071	U	1	0.0071	0.050	ug/L	11/17/25	pd111725
1031-07-8	Endosulfan Sulfate	0.0037	U	1	0.0037	0.050	ug/L	11/17/25	pd111725
50-29-3	4,4-DDT	0.0035	U	1	0.0035	0.050	ug/L	11/17/25	pd111725
72-43-5	Methoxychlor	0.011	U	1	0.011	0.050	ug/L	11/17/25	pd111725
53494-70-5	Endrin ketone	0.0093	U	1	0.0093	0.050	ug/L	11/17/25	pd111725
7421-93-4	Endrin aldehyde	0.011	U	1	0.011	0.050	ug/L	11/17/25	pd111725
5103-71-9	alpha-Chlordane	0.0035	U	1	0.0035	0.050	ug/L	11/17/25	pd111725
5103-74-2	gamma-Chlordane	0.0039	U	1	0.0039	0.050	ug/L	11/17/25	pd111725
8001-35-2	Toxaphene	0.17	U	1	0.17	1.00	ug/L	11/17/25	pd111725
SURROGATES									
2051-24-3	Decachlorobiphenyl	20.3			30 (31) - 150 (149)	102%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	21.2			30 (41) - 150 (134)	106%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091243.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 22:33
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:13:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	2.873	68184897	777.1E6	19.772	21.215
28)	SA Decachlor...	9.062	8.057	101.2E6	694.2E6	20.308	17.540

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091243.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 22:33
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:13:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.542 min
Delta R.T.: -0.002 min
Response: 68184897
Conc: 19.77 ng/ml

Instrument :
ECD_D
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.873 min
Delta R.T.: -0.001 min
Response: 777103026
Conc: 21.21 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.062 min
Delta R.T.: -0.002 min
Response: 101174102
Conc: 20.31 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.057 min
Delta R.T.: -0.002 min
Response: 694160383
Conc: 17.54 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	HDR, Inc.		Date Collected:	
Project:	PVWC Linear Construction		Date Received:	
Client Sample ID:	PB170575BS		SDG No.:	Q3649
Lab Sample ID:	PB170575BS		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	100
Sample Wt/Vol:	30.02 g	Final Vol: 10000 uL	Test:	Pesticide-TCL
Prep Method:	SW3541B	Prep Date: 11/17/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	17.1		1	0.13	1.70	ug/kg	11/17/25 14:00	PB170575
319-85-7	beta-BHC	16.9		1	0.18	1.70	ug/kg	11/17/25 14:00	PB170575
319-86-8	delta-BHC	17.1		1	0.39	1.70	ug/kg	11/17/25 14:00	PB170575
58-89-9	gamma-BHC (Lindane)	17.1		1	0.14	1.70	ug/kg	11/17/25 14:00	PB170575
76-44-8	Heptachlor	16.9		1	0.12	1.70	ug/kg	11/17/25 14:00	PB170575
309-00-2	Aldrin	17.2		1	0.12	1.70	ug/kg	11/17/25 14:00	PB170575
1024-57-3	Heptachlor epoxide	17.0		1	0.19	1.70	ug/kg	11/17/25 14:00	PB170575
959-98-8	Endosulfan I	17.1		1	0.14	1.70	ug/kg	11/17/25 14:00	PB170575
60-57-1	Dieldrin	17.1		1	0.14	1.70	ug/kg	11/17/25 14:00	PB170575
72-55-9	4,4-DDE	17.1		1	0.14	1.70	ug/kg	11/17/25 14:00	PB170575
72-20-8	Endrin	15.6		1	0.14	1.70	ug/kg	11/17/25 14:00	PB170575
33213-65-9	Endosulfan II	17.0		1	0.29	1.70	ug/kg	11/17/25 14:00	PB170575
72-54-8	4,4-DDD	16.7		1	0.15	1.70	ug/kg	11/17/25 14:00	PB170575
1031-07-8	Endosulfan Sulfate	17.0		1	0.13	1.70	ug/kg	11/17/25 14:00	PB170575
50-29-3	4,4-DDT	17.8		1	0.14	1.70	ug/kg	11/17/25 14:00	PB170575
72-43-5	Methoxychlor	16.6		1	0.37	1.70	ug/kg	11/17/25 14:00	PB170575
53494-70-5	Endrin ketone	17.7		1	0.19	1.70	ug/kg	11/17/25 14:00	PB170575
7421-93-4	Endrin aldehyde	17.6		1	0.37	1.70	ug/kg	11/17/25 14:00	PB170575
5103-71-9	alpha-Chlordane	17.3		1	0.12	1.70	ug/kg	11/17/25 14:00	PB170575
5103-74-2	gamma-Chlordane	17.3		1	0.15	1.70	ug/kg	11/17/25 14:00	PB170575
8001-35-2	Toxaphene	5.40	U	1	5.40	33.0	ug/kg	11/17/25 14:00	PB170575
SURROGATES									
2051-24-3	Decachlorobiphenyl	25.0			30 (10) - 150 (143)	125%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	24.0			30 (10) - 150 (131)	120%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
 Data File : PD091224.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2025 14:00
 Operator : AR\AJ
 Sample : PB170575BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB170575BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 01:09:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	75425658	879.1E6	21.871	24.000
2)	SA Decachlor...	9.067	8.060	117.6E6	988.2E6	23.614	24.971
Target Compounds							
2)	A alpha-BHC	3.994	3.385	390.7E6	3248.1E6	48.869	51.392
3)	MA gamma-BHC...	4.324	3.721	365.6E6	2972.1E6	48.919	51.401
4)	MA Heptachlor	4.922	4.073	350.0E6	2808.8E6	48.865	50.876
5)	MB Aldrin	5.263	4.358	352.9E6	2920.3E6	48.627	51.724
6)	B beta-BHC	4.512	4.017	136.3E6	1221.9E6	46.908	50.856
7)	B delta-BHC	4.760	4.253	367.2E6	2995.2E6	49.401	51.221
8)	B Heptachlo...	5.684	4.862	312.4E6	2607.4E6	48.085	51.167
9)	A Endosulfan I	6.068	5.236	296.0E6	2462.6E6	48.178	51.463
10)	B gamma-Chl...	5.939	5.115	317.7E6	2838.5E6	48.866	51.961
11)	B alpha-Chl...	6.020	5.179	316.9E6	2711.4E6	46.041	51.812
12)	B 4,4'-DDE	6.189	5.364	284.5E6	2744.8E6	49.059	51.360
13)	MA Dieldrin	6.340	5.502	316.0E6	2783.2E6	48.879	51.245
14)	MA Endrin	6.568	5.778	223.9E6	2121.1E6	45.101	46.935
15)	B Endosulfa...	6.781	6.070	268.3E6	2387.5E6	48.268	51.146
16)	A 4,4'-DDD	6.700	5.918	236.4E6	2450.3E6	48.133	50.141
17)	MA 4,4'-DDT	7.016	6.172	205.3E6	2088.7E6	49.644	53.375
18)	B Endrin al...	6.910	6.248	205.9E6	1865.6E6	49.420	52.865
19)	B Endosulfa...	7.144	6.472	246.0E6	2285.8E6	47.810	50.982
20)	A Methoxychlor	7.489	6.742	101.7E6	1000.8E6	46.754	49.911
21)	B Endrin ke...	7.625	6.981	279.7E6	2704.1E6	49.652	53.068
22)	Mirex	8.108	7.173	167.0E6	1669.4E6	46.962	50.373



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091224.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 14:00
Operator : AR\AJ
Sample : PB170575BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB170575BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:09:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	HDR, Inc.		Date Collected:	11/14/25
Project:	PVWC Linear Construction		Date Received:	11/14/25
Client Sample ID:	DUPE-0.0-0.5-20251114MS		SDG No.:	Q3649
Lab Sample ID:	Q3649-03MS		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	86.5
Sample Wt/Vol:	30.04 g	Final Vol: 10000 uL	Test:	Pesticide-TCL
Prep Method:	SW3541B	Prep Date: 11/17/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	17.5		1	0.15	2.00	ug/kg	11/17/25 21:38	PB170575
319-85-7	beta-BHC	18.2		1	0.21	2.00	ug/kg	11/17/25 21:38	PB170575
319-86-8	delta-BHC	16.7		1	0.45	2.00	ug/kg	11/17/25 21:38	PB170575
58-89-9	gamma-BHC (Lindane)	17.5		1	0.16	2.00	ug/kg	11/17/25 21:38	PB170575
76-44-8	Heptachlor	18.2		1	0.14	2.00	ug/kg	11/17/25 21:38	PB170575
309-00-2	Aldrin	16.7		1	0.14	2.00	ug/kg	11/17/25 21:38	PB170575
1024-57-3	Heptachlor epoxide	16.4		1	0.22	2.00	ug/kg	11/17/25 21:38	PB170575
959-98-8	Endosulfan I	16.9		1	0.16	2.00	ug/kg	11/17/25 21:38	PB170575
60-57-1	Dieldrin	24.0		1	0.16	2.00	ug/kg	11/17/25 21:38	PB170575
72-55-9	4,4-DDE	16.0		1	0.16	2.00	ug/kg	11/17/25 21:38	PB170575
72-20-8	Endrin	18.7		1	0.16	2.00	ug/kg	11/17/25 21:38	PB170575
33213-65-9	Endosulfan II	16.9		1	0.33	2.00	ug/kg	11/17/25 21:38	PB170575
72-54-8	4,4-DDD	15.8		1	0.17	2.00	ug/kg	11/17/25 21:38	PB170575
1031-07-8	Endosulfan Sulfate	15.9		1	0.15	2.00	ug/kg	11/17/25 21:38	PB170575
50-29-3	4,4-DDT	21.4		1	0.16	2.00	ug/kg	11/17/25 21:38	PB170575
72-43-5	Methoxychlor	18.9		1	0.43	2.00	ug/kg	11/17/25 21:38	PB170575
53494-70-5	Endrin ketone	15.5		1	0.22	2.00	ug/kg	11/17/25 21:38	PB170575
7421-93-4	Endrin aldehyde	16.7		1	0.43	2.00	ug/kg	11/17/25 21:38	PB170575
5103-71-9	alpha-Chlordane	17.0		1	0.14	2.00	ug/kg	11/17/25 21:38	PB170575
5103-74-2	gamma-Chlordane	16.6		1	0.17	2.00	ug/kg	11/17/25 21:38	PB170575
8001-35-2	Toxaphene	6.20	U	1	6.20	38.1	ug/kg	11/17/25 21:38	PB170575
SURROGATES									
2051-24-3	Decachlorobiphenyl	14.4			30 (10) - 150 (143)	72%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	17.4			30 (10) - 150 (131)	87%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
 Data File : PD091239.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2025 21:38
 Operator : AR\AJ
 Sample : Q3649-03MS
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 DUPE-0.0-0.5-20251114MS

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 11/18/2025
 Supervised By :mohammad ahmed 11/19/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 01:13:06 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.541	2.872	57912282	635.9E6	16.793m	17.359
28)	SA Decachlor...	9.063	8.056	71586916	478.7E6	14.369	12.097
Target Compounds							
2)	A alpha-BHC	3.992	3.383	364.1E6	2873.2E6	45.540	45.460
3)	MA gamma-BHC...	4.321	3.719	339.0E6	2552.0E6	45.365m	44.136
4)	MA Heptachlor	4.919	4.071	337.9E6	2498.2E6	47.187	45.249
5)	MB Aldrin	5.261	4.356	314.2E6	2450.9E6	43.303	43.410
6)	B beta-BHC	4.510	4.015	123.5E6	1134.9E6	42.505	47.232
7)	B delta-BHC	4.758	4.251	323.0E6	2524.8E6	43.455	43.175
8)	B Heptachlo...	5.682	4.859	271.4E6	2174.4E6	41.768	42.671
9)	A Endosulfan I	6.065	5.234	269.4E6	1800.6E6	43.849	37.629
10)	B gamma-Chl...	5.935	5.111	273.6E6	2363.0E6	42.094m	43.256m
11)	B alpha-Chl...	6.018	5.177	289.1E6	2317.7E6	42.000	44.289
12)	B 4,4'-DDE	6.188	5.361	238.6E6	2216.9E6	41.145	41.482
13)	MA Dieldrin	6.338	5.500	403.0E6	3265.2E6	62.343	60.121
14)	MA Endrin	6.565	5.777	241.3E6	2030.5E6	48.624	44.930
15)	B Endosulfa...	6.778	6.068	243.5E6	1934.0E6	43.797	41.430
16)	A 4,4'-DDD	6.697	5.916	201.8E6	1598.2E6	41.083	32.704
17)	MA 4,4'-DDT	7.013	6.170	230.3E6	2009.2E6	55.683	51.345
18)	B Endrin al...	6.907	6.246	180.3E6	1384.2E6	43.274	39.224
19)	B Endosulfa...	7.142	6.470	212.4E6	1729.0E6	41.286	38.565
20)	A Methoxychlor	7.486	6.740	106.7E6	804.5E6	49.093	40.123
21)	B Endrin ke...	7.621	6.979	226.6E6	1551.5E6	40.229m	30.448
22)	Mirex	8.104	7.171	142.5E6	1070.3E6	40.066m	32.297

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091239.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 21:38
Operator : AR\AJ
Sample : Q3649-03MS
Misc :
ALS Vial : 20 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
DUPE-0.0-0.5-20251114MS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 11/18/2025
Supervised By :mohammad ahmed 11/19/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:13:06 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	HDR, Inc.	Date Collected:	11/14/25
Project:	PVWC Linear Construction	Date Received:	11/14/25
Client Sample ID:	DUPE-0.0-0.5-20251114MSD	SDG No.:	Q3649
Lab Sample ID:	Q3649-03MSD	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	86.5
Sample Wt/Vol:	30.07 g	Final Vol:	10000 uL
Prep Method:	SW3541B	Test:	Pesticide-TCL
	Prep Date:		11/17/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
319-84-6	alpha-BHC	17.2		1	0.15	2.00	ug/kg	11/17/25 21:52	PB170575
319-85-7	beta-BHC	17.8		1	0.21	2.00	ug/kg	11/17/25 21:52	PB170575
319-86-8	delta-BHC	16.4		1	0.45	2.00	ug/kg	11/17/25 21:52	PB170575
58-89-9	gamma-BHC (Lindane)	16.9		1	0.16	2.00	ug/kg	11/17/25 21:52	PB170575
76-44-8	Heptachlor	16.9		1	0.14	2.00	ug/kg	11/17/25 21:52	PB170575
309-00-2	Aldrin	16.3		1	0.14	2.00	ug/kg	11/17/25 21:52	PB170575
1024-57-3	Heptachlor epoxide	16.0		1	0.22	2.00	ug/kg	11/17/25 21:52	PB170575
959-98-8	Endosulfan I	16.2		1	0.16	2.00	ug/kg	11/17/25 21:52	PB170575
60-57-1	Dieldrin	23.1		1	0.16	2.00	ug/kg	11/17/25 21:52	PB170575
72-55-9	4,4-DDE	15.4		1	0.16	2.00	ug/kg	11/17/25 21:52	PB170575
72-20-8	Endrin	18.0		1	0.16	2.00	ug/kg	11/17/25 21:52	PB170575
33213-65-9	Endosulfan II	16.3		1	0.33	2.00	ug/kg	11/17/25 21:52	PB170575
72-54-8	4,4-DDD	15.3		1	0.17	2.00	ug/kg	11/17/25 21:52	PB170575
1031-07-8	Endosulfan Sulfate	15.2		1	0.15	2.00	ug/kg	11/17/25 21:52	PB170575
50-29-3	4,4-DDT	20.0		1	0.16	2.00	ug/kg	11/17/25 21:52	PB170575
72-43-5	Methoxychlor	17.2		1	0.43	2.00	ug/kg	11/17/25 21:52	PB170575
53494-70-5	Endrin ketone	14.8		1	0.22	2.00	ug/kg	11/17/25 21:52	PB170575
7421-93-4	Endrin aldehyde	15.8		1	0.43	2.00	ug/kg	11/17/25 21:52	PB170575
5103-71-9	alpha-Chlordane	16.5		1	0.14	2.00	ug/kg	11/17/25 21:52	PB170575
5103-74-2	gamma-Chlordane	17.3		1	0.17	2.00	ug/kg	11/17/25 21:52	PB170575
8001-35-2	Toxaphene	6.20	U	1	6.20	38.1	ug/kg	11/17/25 21:52	PB170575
SURROGATES									
2051-24-3	Decachlorobiphenyl	13.1			30 (10) - 150 (143)	65%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	17.1			30 (10) - 150 (131)	86%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
 Data File : PD091240.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2025 21:52
 Operator : AR\AJ
 Sample : Q3649-03MSD
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 DUPE-0.0-0.5-20251114MSD

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 11/18/2025
 Supervised By :mohammad ahmed 11/19/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 01:13:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.540	2.872	56160904	628.1E6	16.285m	17.148
28)	SA Decachlor...	9.062	8.057	65039166	450.6E6	13.055	11.385
Target Compounds							
2)	A alpha-BHC	3.991	3.383	352.1E6	2828.3E6	44.040	44.750
3)	MA gamma-BHC...	4.320	3.719	329.1E6	2492.1E6	44.037m	43.101
4)	MA Heptachlor	4.919	4.071	315.6E6	2338.1E6	44.062	42.350
5)	MB Aldrin	5.260	4.356	306.4E6	2391.1E6	42.225	42.352
6)	B beta-BHC	4.509	4.016	124.6E6	1109.7E6	42.895	46.187
7)	B delta-BHC	4.757	4.251	317.5E6	2461.7E6	42.726	42.097
8)	B Heptachlo...	5.680	4.860	265.3E6	2116.3E6	40.835	41.530
9)	A Endosulfan I	6.064	5.235	259.5E6	1751.6E6	42.248	36.604
10)	B gamma-Chl...	5.935	5.113	265.2E6	2455.4E6	40.799m	44.948
11)	B alpha-Chl...	6.017	5.177	275.6E6	2248.9E6	40.042	42.975
12)	B 4,4'-DDE	6.186	5.362	230.0E6	2145.5E6	39.655	40.146
13)	MA Dieldrin	6.337	5.499	388.8E6	3175.0E6	60.133	58.460
14)	MA Endrin	6.564	5.776	232.5E6	1951.2E6	46.847	43.176
15)	B Endosulfa...	6.777	6.067	236.0E6	1858.7E6	42.455	39.818
16)	A 4,4'-DDD	6.696	5.916	195.2E6	1555.3E6	39.735	31.826
17)	MA 4,4'-DDT	7.012	6.170	214.8E6	1847.1E6	51.952	47.202
18)	B Endrin al...	6.906	6.246	170.7E6	1309.6E6	40.971	37.109
19)	B Endosulfa...	7.141	6.469	203.5E6	1631.4E6	39.556	36.388
20)	A Methoxychlor	7.485	6.740	97413476	737.8E6	44.801	36.796
21)	B Endrin ke...	7.620	6.978	217.4E6	1442.2E6	38.595m	28.303 #
22)	Mirex	8.102	7.171	135.8E6	1007.3E6	38.181m	30.396

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111725\
Data File : PD091240.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2025 21:52
Operator : AR\AJ
Sample : Q3649-03MSD
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
DUPE-0.0-0.5-20251114MSD

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 11/18/2025
Supervised By :mohammad ahmed 11/19/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 01:13:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integration Report

Sequence:	PD111525	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PD091194.D	4,4"-DDE #2	Abdul	11/17/2025 8:53:25 AM	Sohil	11/17/2025 2:13:45	Peak Integrated by Software
PEM	PD091194.D	Endrin aldehyde	Abdul	11/17/2025 8:53:25 AM	Sohil	11/17/2025 2:13:45	Peak Integrated by Software
PEM	PD091194.D	Endrin aldehyde #2	Abdul	11/17/2025 8:53:25 AM	Sohil	11/17/2025 2:13:45	Peak Integrated by Software
PSTDICC005	PD091200.D	delta-BHC	Abdul	11/17/2025 8:54:07 AM	Sohil	11/17/2025 2:13:50	Peak Integrated by Software
PSTDICC005	PD091200.D	Heptachlor	Abdul	11/17/2025 8:54:07 AM	Sohil	11/17/2025 2:13:50	Peak Integrated by Software
PSTDICC005	PD091200.D	Heptachlor epoxide	Abdul	11/17/2025 8:54:07 AM	Sohil	11/17/2025 2:13:50	Peak Integrated by Software
PSTDICC005	PD091200.D	Tetrachloro-m-xylene	Abdul	11/17/2025 8:54:07 AM	Sohil	11/17/2025 2:13:50	Peak Integrated by Software
PSTDICC005	PD091200.D	Tetrachloro-m-xylene #2	Abdul	11/17/2025 8:54:07 AM	Sohil	11/17/2025 2:13:50	Peak Integrated by Software
PEM	PD091215.D	4,4"-DDE #2	Abdul	11/17/2025 8:53:18 AM	Sohil	11/17/2025 2:14:03	Peak Integrated by Software

Manual Integration Report

Sequence:	pd111725	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PD091235.D	4,4"-DDD #2	Abdul	11/18/2025 11:03:29 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
PEM	PD091235.D	Endrin aldehyde	Abdul	11/18/2025 11:03:29 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
PEM	PD091235.D	Endrin ketone #2	Abdul	11/18/2025 11:03:29 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
Q3649-01	PD091237.D	Decachlorobiphenyl #2	Abdul	11/18/2025 11:03:25 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
Q3649-01	PD091237.D	Tetrachloro-m-xylene	Abdul	11/18/2025 11:03:25 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
Q3649-01	PD091237.D	Tetrachloro-m-xylene #2	Abdul	11/18/2025 11:03:25 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
Q3649-03	PD091238.D	Decachlorobiphenyl #2	Abdul	11/18/2025 11:03:23 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
Q3649-03	PD091238.D	Tetrachloro-m-xylene	Abdul	11/18/2025 11:03:23 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
Q3649-03	PD091238.D	Tetrachloro-m-xylene #2	Abdul	11/18/2025 11:03:23 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
Q3649-03MS	PD091239.D	Endrin ketone	Abdul	11/18/2025 11:03:19 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
Q3649-03MS	PD091239.D	gamma-BHC (Lindane)	Abdul	11/18/2025 11:03:19 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
Q3649-03MS	PD091239.D	gamma-Chlordane	Abdul	11/18/2025 11:03:19 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
Q3649-03MS	PD091239.D	gamma-Chlordane #2	Abdul	11/18/2025 11:03:19 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software

Manual Integration Report

Sequence:	pd111725	Instrument	ECD_d
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q3649-03MS	PD091239.D	Mirex	Abdul	11/18/2025 11:03:19 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
Q3649-03MS	PD091239.D	Tetrachloro-m-xylene	Abdul	11/18/2025 11:03:19 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
Q3649-03MSD	PD091240.D	Endrin ketone	Abdul	11/18/2025 11:03:16 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
Q3649-03MSD	PD091240.D	gamma-BHC (Lindane)	Abdul	11/18/2025 11:03:16 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
Q3649-03MSD	PD091240.D	gamma-Chlordane	Abdul	11/18/2025 11:03:16 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
Q3649-03MSD	PD091240.D	Mirex	Abdul	11/18/2025 11:03:16 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
Q3649-03MSD	PD091240.D	Tetrachloro-m-xylene	Abdul	11/18/2025 11:03:16 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
Q3649-06	PD091241.D	Decachlorobiphenyl	Abdul	11/18/2025 11:03:13 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
Q3649-06	PD091241.D	Tetrachloro-m-xylene	Abdul	11/18/2025 11:03:13 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software
Q3649-08	PD091242.D	Tetrachloro-m-xylene	Abdul	11/18/2025 11:03:10 AM	mohammad	11/19/2025 12:23:26	Peak Integrated by Software

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111525

Review By	rahul	Review On	11/14/2025 3:12:58 PM
Supervise By	Sohil	Supervise On	11/17/2025 2:14:18 PM
SubDirectory	PD111525	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	I.BLK	PD091193.D	14 Nov 2025 20:01	AR\AJ	Ok
2	PEM	PD091194.D	14 Nov 2025 20:15	AR\AJ	Ok,M
3	RESCHK	PD091195.D	14 Nov 2025 20:43	AR\AJ	Ok
4	PSTDICC100	PD091196.D	14 Nov 2025 20:56	AR\AJ	Ok
5	PSTDICC075	PD091197.D	14 Nov 2025 21:10	AR\AJ	Ok
6	PSTDICC050	PD091198.D	14 Nov 2025 21:24	AR\AJ	Ok
7	PSTDICC025	PD091199.D	14 Nov 2025 21:37	AR\AJ	Ok
8	PSTDICC005	PD091200.D	14 Nov 2025 21:51	AR\AJ	Ok,M
9	PCHLORICC1000	PD091201.D	14 Nov 2025 22:04	AR\AJ	Ok
10	PCHLORICC750	PD091202.D	14 Nov 2025 22:18	AR\AJ	Ok
11	PCHLORICC500	PD091203.D	14 Nov 2025 22:32	AR\AJ	Ok
12	PCHLORICC250	PD091204.D	14 Nov 2025 22:46	AR\AJ	Ok,M
13	PCHLORICC50	PD091205.D	14 Nov 2025 22:59	AR\AJ	Ok,M
14	PTOXICC1000	PD091206.D	14 Nov 2025 23:13	AR\AJ	Ok
15	PTOXICC750	PD091207.D	14 Nov 2025 23:27	AR\AJ	Ok
16	PTOXICC500	PD091208.D	14 Nov 2025 23:40	AR\AJ	Ok
17	PTOXICC250	PD091209.D	14 Nov 2025 23:54	AR\AJ	Ok
18	PTOXICC100	PD091210.D	15 Nov 2025 00:08	AR\AJ	Ok
19	PSTDICV050	PD091211.D	15 Nov 2025 00:21	AR\AJ	Ok
20	PCHLORICV050	PD091212.D	15 Nov 2025 00:49	AR\AJ	Ok
21	PTOXICV050	PD091213.D	15 Nov 2025 01:30	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111525

Review By	rahul	Review On	11/14/2025 3:12:58 PM
Supervise By	Sohil	Supervise On	11/17/2025 2:14:18 PM
SubDirectory	PD111525	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	I.BLK	PD091214.D	15 Nov 2025 01:44	AR\AJ	Ok
23	PEM	PD091215.D	15 Nov 2025 01:57	AR\AJ	Ok,M
24	PSTDCCC050	PD091216.D	15 Nov 2025 02:11	AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111725

Review By	Abdul	Review On	11/18/2025 11:05:20 AM
Supervise By	mohammad	Supervise On	11/19/2025 12:23:26 AM
SubDirectory	PD111725	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD091217.D	17 Nov 2025 09:35	AR\AJ	Ok
2	I.BLK	PD091218.D	17 Nov 2025 09:51	AR\AJ	Ok
3	PEM	PD091219.D	17 Nov 2025 10:04	AR\AJ	Ok
4	PSTDCCC050	PD091220.D	17 Nov 2025 10:18	AR\AJ	Ok
5	PB170527BS	PD091221.D	17 Nov 2025 12:54	AR\AJ	Ok
6	PB170506BS	PD091222.D	17 Nov 2025 13:33	AR\AJ	Ok
7	PB170575BL	PD091223.D	17 Nov 2025 13:46	AR\AJ	Ok
8	PB170575BS	PD091224.D	17 Nov 2025 14:00	AR\AJ	Ok
9	Q3638-01	PD091225.D	17 Nov 2025 14:14	AR\AJ	Ok,M
10	Q3638-03	PD091226.D	17 Nov 2025 14:27	AR\AJ	Ok,M
11	Q3640-02	PD091227.D	17 Nov 2025 14:41	AR\AJ	Ok,M
12	Q3640-03	PD091228.D	17 Nov 2025 14:55	AR\AJ	Ok,M
13	Q3641-01	PD091229.D	17 Nov 2025 15:08	AR\AJ	Ok,M
14	Q3645-01	PD091230.D	17 Nov 2025 15:49	AR\AJ	Ok
15	Q3646-01	PD091231.D	17 Nov 2025 16:02	AR\AJ	Ok,M
16	Q3647-01	PD091232.D	17 Nov 2025 16:16	AR\AJ	Ok,M
17	Q3648-01	PD091233.D	17 Nov 2025 16:30	AR\AJ	Ok,M
18	I.BLK	PD091234.D	17 Nov 2025 19:21	AR\AJ	Ok
19	PEM	PD091235.D	17 Nov 2025 19:35	AR\AJ	Ok,M
20	PSTDCCC050	PD091236.D	17 Nov 2025 20:02	AR\AJ	Ok
21	Q3649-01	PD091237.D	17 Nov 2025 21:11	AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111725

Review By	Abdul	Review On	11/18/2025 11:05:20 AM
Supervise By	mohammad	Supervise On	11/19/2025 12:23:26 AM
SubDirectory	PD111725	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q3649-03	PD091238.D	17 Nov 2025 21:24	AR\AJ	Ok,M
23	Q3649-03MS	PD091239.D	17 Nov 2025 21:38	AR\AJ	Ok,M
24	Q3649-03MSD	PD091240.D	17 Nov 2025 21:52	AR\AJ	Ok,M
25	Q3649-06	PD091241.D	17 Nov 2025 22:05	AR\AJ	Ok,M
26	Q3649-08	PD091242.D	17 Nov 2025 22:19	AR\AJ	Ok,M
27	I.BLK	PD091243.D	17 Nov 2025 22:33	AR\AJ	Ok
28	PSTDCCC050	PD091244.D	17 Nov 2025 23:00	AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111525

Review By	rahul	Review On	11/14/2025 3:12:58 PM
Supervise By	Sohil	Supervise On	11/17/2025 2:14:18 PM
SubDirectory	PD111525	HP Acquire Method	HP Processing Method pd111525 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24942,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	I.BLK	I.BLK	PD091193.D	14 Nov 2025 20:01		AR\AJ	Ok
2	PEM	PEM	PD091194.D	14 Nov 2025 20:15		AR\AJ	Ok,M
3	RESCHK	RESCHK	PD091195.D	14 Nov 2025 20:43		AR\AJ	Ok
4	PSTDICC100	PSTDICC100	PD091196.D	14 Nov 2025 20:56		AR\AJ	Ok
5	PSTDICC075	PSTDICC075	PD091197.D	14 Nov 2025 21:10		AR\AJ	Ok
6	PSTDICC050	PSTDICC050	PD091198.D	14 Nov 2025 21:24		AR\AJ	Ok
7	PSTDICC025	PSTDICC025	PD091199.D	14 Nov 2025 21:37		AR\AJ	Ok
8	PSTDICC005	PSTDICC005	PD091200.D	14 Nov 2025 21:51		AR\AJ	Ok,M
9	PCHLORICC1000	PCHLORICC1000	PD091201.D	14 Nov 2025 22:04		AR\AJ	Ok
10	PCHLORICC750	PCHLORICC750	PD091202.D	14 Nov 2025 22:18		AR\AJ	Ok
11	PCHLORICC500	PCHLORICC500	PD091203.D	14 Nov 2025 22:32		AR\AJ	Ok
12	PCHLORICC250	PCHLORICC250	PD091204.D	14 Nov 2025 22:46		AR\AJ	Ok,M
13	PCHLORICC50	PCHLORICC50	PD091205.D	14 Nov 2025 22:59		AR\AJ	Ok,M
14	PTOXICC1000	PTOXICC1000	PD091206.D	14 Nov 2025 23:13		AR\AJ	Ok
15	PTOXICC750	PTOXICC750	PD091207.D	14 Nov 2025 23:27		AR\AJ	Ok
16	PTOXICC500	PTOXICC500	PD091208.D	14 Nov 2025 23:40		AR\AJ	Ok
17	PTOXICC250	PTOXICC250	PD091209.D	14 Nov 2025 23:54		AR\AJ	Ok
18	PTOXICC100	PTOXICC100	PD091210.D	15 Nov 2025 00:08		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111525

Review By	rahul	Review On	11/14/2025 3:12:58 PM
Supervise By	Sohil	Supervise On	11/17/2025 2:14:18 PM
SubDirectory	PD111525	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PSTDICV050	ICVPD111525	PD091211.D	15 Nov 2025 00:21		AR\AJ	Ok
20	PCHLORICV050	ICVPD111525CHLOR	PD091212.D	15 Nov 2025 00:49		AR\AJ	Ok
21	PTOXICV050	ICVPD111525TOX	PD091213.D	15 Nov 2025 01:30		AR\AJ	Ok
22	I.BLK	I.BLK	PD091214.D	15 Nov 2025 01:44		AR\AJ	Ok
23	PEM	PEM	PD091215.D	15 Nov 2025 01:57		AR\AJ	Ok,M
24	PSTDCCC050	PSTDCCC050	PD091216.D	15 Nov 2025 02:11		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111725

Review By	Abdul	Review On	11/18/2025 11:05:20 AM
Supervise By	mohammad	Supervise On	11/19/2025 12:23:26 AM
SubDirectory	PD111725	HP Acquire Method	HP Processing Method pd111525 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24942,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD091217.D	17 Nov 2025 09:35		AR\AJ	Ok
2	I.BLK	I.BLK	PD091218.D	17 Nov 2025 09:51		AR\AJ	Ok
3	PEM	PEM	PD091219.D	17 Nov 2025 10:04		AR\AJ	Ok
4	PSTDCCC050	PSTDCCC050	PD091220.D	17 Nov 2025 10:18		AR\AJ	Ok
5	PB170527BS	PB170527BS	PD091221.D	17 Nov 2025 12:54		AR\AJ	Ok
6	PB170506BS	PB170506BS	PD091222.D	17 Nov 2025 13:33		AR\AJ	Ok
7	PB170575BL	PB170575BL	PD091223.D	17 Nov 2025 13:46		AR\AJ	Ok
8	PB170575BS	PB170575BS	PD091224.D	17 Nov 2025 14:00		AR\AJ	Ok
9	Q3638-01	MOO-25-0327	PD091225.D	17 Nov 2025 14:14		AR\AJ	Ok,M
10	Q3638-03	MOO-25-0328	PD091226.D	17 Nov 2025 14:27		AR\AJ	Ok,M
11	Q3640-02	LAW-25-0179-0180	PD091227.D	17 Nov 2025 14:41		AR\AJ	Ok,M
12	Q3640-03	LAW-25-0182	PD091228.D	17 Nov 2025 14:55	TCMX low in 2nd column	AR\AJ	Ok,M
13	Q3641-01	TR-05-111325	PD091229.D	17 Nov 2025 15:08		AR\AJ	Ok,M
14	Q3645-01	HD-02-11142025	PD091230.D	17 Nov 2025 15:49		AR\AJ	Ok
15	Q3646-01	SU-03-11142025	PD091231.D	17 Nov 2025 16:02		AR\AJ	Ok,M
16	Q3647-01	VNJ-212	PD091232.D	17 Nov 2025 16:16		AR\AJ	Ok,M
17	Q3648-01	OR-02-111425	PD091233.D	17 Nov 2025 16:30		AR\AJ	Ok,M
18	I.BLK	I.BLK	PD091234.D	17 Nov 2025 19:21		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD111725

Review By	Abdul	Review On	11/18/2025 11:05:20 AM
Supervise By	mohammad	Supervise On	11/19/2025 12:23:26 AM
SubDirectory	PD111725	HP Acquire Method	HP Processing Method pd111525 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PEM	PEM	PD091235.D	17 Nov 2025 19:35		AR\AJ	Ok,M
20	PSTDCCC050	PSTDCCC050	PD091236.D	17 Nov 2025 20:02		AR\AJ	Ok
21	Q3649-01	B5-0.0-0.5-20251114	PD091237.D	17 Nov 2025 21:11		AR\AJ	Ok,M
22	Q3649-03	DUPE-0.0-0.5-20251114	PD091238.D	17 Nov 2025 21:24		AR\AJ	Ok,M
23	Q3649-03MS	DUPE-0.0-0.5-20251114	PD091239.D	17 Nov 2025 21:38		AR\AJ	Ok,M
24	Q3649-03MSD	DUPE-0.0-0.5-20251114	PD091240.D	17 Nov 2025 21:52		AR\AJ	Ok,M
25	Q3649-06	B4-0.0-0.5-20251114	PD091241.D	17 Nov 2025 22:05		AR\AJ	Ok,M
26	Q3649-08	B3-0.0-0.5-20251114	PD091242.D	17 Nov 2025 22:19		AR\AJ	Ok,M
27	I.BLK	I.BLK	PD091243.D	17 Nov 2025 22:33		AR\AJ	Ok
28	PSTDCCC050	PSTDCCC050	PD091244.D	17 Nov 2025 23:00		AR\AJ	Ok

M : Manual Integration

PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 11/17/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 11/14/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:27
Out Date: 11/15/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB137900

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3635-01	40635	1	1.00	1.00	2.00	2.00	100.0	debris
Q3635-02	40617	2	1.00	1.00	2.00	2.00	100.0	oil sample
Q3636-01	AUD-25-0114	3	1.15	10.21	11.36	10.28	89.4	
Q3637-01	BUR-25-0100	4	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3637-02	BUR-25-0100	5	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3638-01	MOO-25-0327	6	1.17	10.24	11.41	6.36	50.7	
Q3638-03	MOO-25-0328	7	1.16	10.37	11.53	10.94	94.3	
Q3639-01	L66A	8	1.00	1.00	2.00	2.00	100.0	PILC
Q3639-02	L66B	9	1.00	1.00	2.00	2.00	100.0	PILC
Q3639-03	L67A	10	1.00	1.00	2.00	2.00	100.0	PILC
Q3639-04	L67B	11	1.00	1.00	2.00	2.00	100.0	PILC
Q3639-05	L68A	12	1.00	1.00	2.00	2.00	100.0	PILC
Q3639-06	L68B	13	1.00	1.00	2.00	2.00	100.0	PILC
Q3639-07	L69A	14	1.00	1.00	2.00	2.00	100.0	PILC
Q3639-08	L69B	15	1.00	1.00	2.00	2.00	100.0	PILC
Q3639-09	L70A	16	1.00	1.00	2.00	2.00	100.0	PILC
Q3639-10	L70B	17	1.00	1.00	2.00	2.00	100.0	PILC
Q3639-11	L71A	18	1.00	1.00	2.00	2.00	100.0	PILC
Q3639-12	L71B	19	1.00	1.00	2.00	2.00	100.0	PILC
Q3639-13	L72A	20	1.00	1.00	2.00	2.00	100.0	PILC
Q3639-14	L72B	21	1.00	1.00	2.00	2.00	100.0	PILC
Q3639-15	STG15-1-1	22	1.00	1.00	2.00	2.00	100.0	PILC
Q3639-16	STG15-1-2	23	1.00	1.00	2.00	2.00	100.0	PILC
Q3640-01	LAW-25-0179	24	1.14	10.16	11.3	9.4	81.3	
Q3640-02	LAW-25-0179-0180	25	1.19	10.47	11.66	10.81	91.9	
Q3640-03	LAW-25-0182	26	1.18	10.35	11.53	10.62	91.2	
Q3641-01	TR-05-111325	27	1.11	10.23	11.34	10.74	94.1	
Q3641-01D	TR-05-111325DUP	28	1.14	10.21	11.35	10.75	94.1	



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 11/17/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 11/14/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:27
Out Date: 11/15/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB137900

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3641-02	TR-05-111325-E2	29	1.14	10.73	11.87	11.44	96.0	
Q3643-05	SVOC-GPC-BLANK	37	1.00	1.00	2.00	2.00	100.0	
Q3643-06	PEST-GPC-BLANK	38	1.00	1.00	2.00	2.00	100.0	
Q3643-07	PEST-GPC-BLANK-SPIKE	39	1.00	1.00	2.00	2.00	100.0	
Q3643-08	SVOC-GPC2-BLANK	40	1.00	1.00	2.00	2.00	100.0	
Q3643-09	PEST-GPC2-BLANK	41	1.00	1.00	2.00	2.00	100.0	
Q3643-10	PEST -GPC2-BLANK-SPIKE	42	1.00	1.00	2.00	2.00	100.0	
Q3645-01	HD-02-11142025	30	1.18	10.50	11.68	11.02	93.7	
Q3645-02	HD-02-11142025-E2	31	1.19	10.43	11.62	10.75	91.7	
Q3646-01	SU-03-11142025	32	1.11	10.26	11.37	10.38	90.4	
Q3646-02	SU-03-11142025-E2	33	1.18	11.13	12.31	11.35	91.4	
Q3647-01	VNJ-212	34	1.18	10.39	11.57	11.07	95.2	
Q3648-01	OR-02-111425	35	1.17	10.59	11.76	10.75	90.5	
Q3648-02	OR-02-111425-E2	36	1.14	10.78	11.92	10.96	91.1	
Q3649-01	B5-0.0-0.5-20251114	43	1.14	10.38	11.52	10.54	90.6	
Q3649-02	B5-0.5-1.0-20251114	44	1.16	10.61	11.77	10.8	90.9	
Q3649-03	DUPE-0.0-0.5-20251114	45	1.19	10.18	11.37	10.00	86.5	
Q3649-04	DUPE-0.5-1.0-20251114	46	1.11	10.07	11.18	10.04	88.7	
Q3649-05	B4-0.05-1.0-20251114	47	1.19	10.57	11.76	10.7	90.0	
Q3649-06	B4-0.0-0.5-20251114	48	1.14	10.54	11.68	10.77	91.4	
Q3649-07	B4-0.0-0.5-20251114	49	1.14	10.54	11.68	10.77	91.4	
Q3649-08	B3-0.0-0.5-20251114	50	1.16	10.61	11.77	10.28	86.0	
Q3649-09	B3-0.5-1.0-20251114	51	1.13	10.45	11.58	10.45	89.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

137900

WorkList Name : %1-111425 WorkList ID : 193119 Department : Wet-Chemistry Date : 11-14-2025 09:33:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3635-01	40635	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3635-02	40617	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3636-01	AUD-25-0114	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3637-01	BUR-25-0100	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3637-02	BUR-25-0100	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3638-01	MOO-25-0327	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3638-03	MOO-25-0328	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3639-01	L66A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3639-02	L66B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3639-03	L67A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3639-04	L67B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3639-05	L68A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3639-06	L68B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3639-07	L69A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3639-08	L69B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3639-09	L70A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3639-10	L70B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3639-11	L71A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3639-12	L71B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3639-13	L72A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3639-14	L72B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO

Date/Time: 11-14-25 15:25
 Raw Sample Received by: SAUCI
 Raw Sample Relinquished by: [Signature]
 Date/Time: 11-14-25 15:40
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

127900

WorkList Name : %1-111425

WorkList ID : 193119

Department : Wet-Chemistry

Date : 11-14-2025 09:33:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3639-15	STG15-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3639-16	STG15-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3640-01	LAW-25-0179	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3640-02	LAW-25-0179-0180	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3640-03	LAW-25-0182	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/13/2025	Chemtech -SO
Q3641-01	TR-05-111325	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	11/13/2025	Chemtech -SO
Q3641-02	TR-05-111325-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	11/13/2025	Chemtech -SO
Q3643-05	SVOC-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	A12	11/07/2025	Chemtech -SO
Q3643-06	PEST-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	A12	11/07/2025	Chemtech -SO
Q3643-07	PEST-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	A12	11/07/2025	Chemtech -SO
Q3643-08	SVOC-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	A12	11/07/2025	Chemtech -SO
Q3643-09	PEST-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	A12	11/07/2025	Chemtech -SO
Q3643-10	PEST -GPC2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	A12	11/07/2025	Chemtech -SO
Q3645-01	HD-02-11142025	Solid	Percent Solids	Cool 4 deg C	PSEG05	D51	11/14/2025	Chemtech -SO
Q3645-02	HD-02-11142025-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D51	11/14/2025	Chemtech -SO
Q3646-01	SU-03-11142025	Solid	Percent Solids	Cool 4 deg C	PSEG05	D31	11/14/2025	Chemtech -SO
Q3646-02	SU-03-11142025-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D31	11/14/2025	Chemtech -SO
Q3647-01	VNJ-212	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	11/14/2025	Chemtech -SO
Q3648-01	OR-02-111425	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	11/14/2025	Chemtech -SO
Q3648-02	OR-02-111425-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	11/14/2025	Chemtech -SO
Q3649-01	B5-0-0-0.5-20251114	Solid	Percent Solids	Cool 4 deg C	HDR108	D41	11/14/2025	Chemtech -SO

Date/Time 11-14-25 15:25
 Raw Sample Received by: SO WOLF SM
 Raw Sample Relinquished by: SO WOLF SM

Date/Time 11-14-25 17:40
 Raw Sample Received by: SO WOLF SM
 Raw Sample Relinquished by: SO WOLF SM

WORKLIST(Hardcopy Internal Chain)

JB 137900

WorkList Name : %1-111425

WorkList ID : 193119

Department : Wet-Chemistry

Date : 11-14-2025 09:33:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3649-02	B5-0.5-1.0-20251114	Solid	Percent Solids	Cool 4 deg C	HDR108	D41	11/14/2025	Chemtech -SO
Q3649-03	DUPE-0.0-0.5-20251114	Solid	Percent Solids	Cool 4 deg C	HDR108	D41	11/14/2025	Chemtech -SO
Q3649-04	DUPE-0.5-1.0-20251114	Solid	Percent Solids	Cool 4 deg C	HDR108	D41	11/14/2025	Chemtech -SO
Q3649-05	B4-0.05-1.0-20251114	Solid	Percent Solids	Cool 4 deg C	HDR108	D41	11/14/2025	Chemtech -SO
Q3649-06	B4-0.0-0.5-20251114	Solid	Percent Solids	Cool 4 deg C	HDR108	D41	11/14/2025	Chemtech -SO
Q3649-07	B4-0.0-0.5-20251114	Solid	Percent Solids	Cool 4 deg C	HDR108	D41	11/14/2025	Chemtech -SO
Q3649-08	B3-0.0-0.5-20251114	Solid	Percent Solids	Cool 4 deg C	HDR108	D41	11/14/2025	Chemtech -SO
Q3649-09	B3-0.5-1.0-20251114	Solid	Percent Solids	Cool 4 deg C	HDR108	D41	11/14/2025	Chemtech -SO

Date/Time 11-14-25 15:25
 Raw Sample Received by: JB wgy
 Raw Sample Relinquished by: JB wgy

Date/Time 11-14-25 17:40
 Raw Sample Received by: JB wgy
 Raw Sample Relinquished by: JB wgy

SOP ID: M3541-ASE Extraction-15

Clean Up SOP #: Florisil

Extraction Start Date : 11/17/2025

Matrix : Solid

Extraction Start Time : 09:06

Weigh By: EH

Extraction By: RJ

Extraction End Date : 11/17/2025

Balance check: EH

Filter By: RJ

Extraction End Time : 12:15

Balance ID: EX-SC-2

pH Meter ID: N/A

Concentration By: EH

pH Strip Lot#: N/A

Hood ID: 3,7

Supervisor By : RUPESH

Extraction Method: ☐ Separatory Funnel

☐ Continous Liquid/Liquid

☐ Sonication

☐ Waste Dilution

☒ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	500 PPB	PP24766
Surrogate	1.0ML	200 PPB	PP24663
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2660
Baked Na2SO4	N/A	EP2655
Sand	N/A	E3951
Hexane	N/A	E3984
Florisil	N/A	E3976
9:1 Hexane:Acetone Mixture	N/A	EP2644
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS721.

KD Bath ID: N/A

Envap ID: N-EVAP-02

KD Bath Temperature: N/A


Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/17/25	RS (Bath Lab)	RS (Bath Lab)
12:20	Preparation Group	Analysis Group

Analytical Method: M3541-ASE Extraction-15

Concentration Date: 11/14/2025

Sample ID	Client Sample ID	Test	g/mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB170575BL	PBLK575	Pesticide-TCL	30.01	N/A	ritesh	Evelyn	10			U1-1
PB170575BS	PLCS575	Pesticide-TCL	30.02	N/A	ritesh	Evelyn	10			2
Q3638-01	MOO-25-0327	Pesticide-TCL	30.05	N/A	ritesh	Evelyn	10	E	Small Particles	3
Q3638-03	MOO-25-0328	Pesticide-TCL	30.04	N/A	ritesh	Evelyn	10	E	Small Particles	4
Q3640-02	LAW-25-0179-0180	Pesticide-TCL	30.06	N/A	ritesh	Evelyn	10	A		5
Q3640-03	LAW-25-0182	Pesticide-TCL	30.02	N/A	ritesh	Evelyn	10	E	Small Particles	6
Q3641-01	TR-05-111325	Pesticide-TCL	30.03	N/A	ritesh	Evelyn	10	E		U2-1
Q3645-01	HD-02-11142025	Pesticide-TCL	30.04	N/A	ritesh	Evelyn	10	E		2
Q3646-01	SU-03-11142025	Pesticide-TCL	30.08	N/A	ritesh	Evelyn	10	E		3
Q3647-01	VNJ-212	Pesticide-TCL	30.03	N/A	ritesh	Evelyn	10	E	Concrete	4
Q3648-01	OR-02-111425	Pesticide-TCL	30.07	N/A	ritesh	Evelyn	10	E		5
Q3649-01	B5-0.0-0.5-20251114	Pesticide-TCL	30.02	N/A	ritesh	Evelyn	10	B		6
Q3649-03	DUPE-0.0-0.5-20251114	Pesticide-TCL	30.09	N/A	ritesh	Evelyn	10	B		U3-1
Q3649-03MS	DUPE-0.0-0.5-20251114 MS	Pesticide-TCL	30.04	N/A	ritesh	Evelyn	10	B		2
Q3649-03MS D	DUPE-0.0-0.5-20251114 MSD	Pesticide-TCL	30.07	N/A	ritesh	Evelyn	10	B		3
Q3649-06	B4-0.0-0.5-20251114	Pesticide-TCL	30.01	N/A	ritesh	Evelyn	10	B		4
Q3649-08	B3-0.0-0.5-20251114	Pesticide-TCL	30.05	N/A	ritesh	Evelyn	10	B		5



RS
11/14

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q3638 **WorkList ID :** 193146 **Department :** Extraction **Date :** 11-17-2025 08:59:04

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3638-01	MOO-25-0327	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	D41	11/13/2025	8081B
Q3638-03	MOO-25-0328	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	D41	11/13/2025	8081B
Q3640-02	LAW-25-0179-0180	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	D41	11/13/2025	8081B
Q3640-03	LAW-25-0182	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	D41	11/13/2025	8081B
Q3641-01	TR-05-111325	Solid	Pesticide-TCL	Cool 4 deg C	PSEG05	D41	11/13/2025	8081B
Q3645-01	HD-02-11142025	Solid	Pesticide-TCL	Cool 4 deg C	PSEG05	D51	11/14/2025	8081B
Q3646-01	SU-03-11142025	Solid	Pesticide-TCL	Cool 4 deg C	PSEG05	D31	11/14/2025	8081B
Q3647-01	VNJ-212	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	D41	11/14/2025	8081B
Q3648-01	OR-02-111425	Solid	Pesticide-TCL	Cool 4 deg C	PSEG05	D41	11/14/2025	8081B
Q3649-01	B5-0.0-0.5-20251114	Solid	Pesticide-TCL	Cool 4 deg C	HDR108	D41	11/14/2025	8081B
Q3649-03	DUPE-0.0-0.5-20251114	Solid	Pesticide-TCL	Cool 4 deg C	HDR108	D41	11/14/2025	8081B
Q3649-06	B4-0.0-0.5-20251114	Solid	Pesticide-TCL	Cool 4 deg C	HDR108	D41	11/14/2025	8081B
Q3649-08	B3-0.0-0.5-20251114	Solid	Pesticide-TCL	Cool 4 deg C	HDR108	D41	11/14/2025	8081B

Date/Time 11/17/25 9:01
Raw Sample Received by: RJ (Ext 106)
Raw Sample Relinquished by: CAP

Date/Time 11/17/25 9:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: RJ (Ext 106)



SHIPPING DOCUMENTS

CLIENT INFORMATION

COMPANY: HDR REPORT TO BE SENT TO:
ADDRESS: 50 Tice Blvd Suite 210
CITY: Woodcliff Lake STATE: NJ ZIP: 07677
ATTENTION: Andrew Wadden
PHONE: 201-312-0752 FAX: N/A

CLIENT PROJECT INFORMATION

PROJECT NAME: PVWL
PROJECT NO.: 10367568 LOCATION: Cliffon, NJ
PROJECT MANAGER: Andrew Wadden
e-mail: Andrew.Wadden@HDRInc.com
PHONE: 201-312-0752 FAX: N/A

CLIENT BILLING INFORMATION

BILL TO: HDR PO#:
ADDRESS: 50 Tice Blvd Suite 210
CITY: Woodcliff Lake STATE: NJ ZIP: 07677
ATTENTION: Andrew Wadden PHONE: 201-312-0752

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): Standard 1AT DAYS*
EDD: _____ DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☒ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other _____
☐ EDD FORMAT

1. TCL VOCs
2. TCLSVOCs
3. Petroleum Hydrocarbons
4. PAHs
5. PCBs
6. PCBs for PCBs
7. PCBs for PCBs
8. PCBs for PCBs
9. PCBs for PCBs

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS	
			COMP	GRAB	DATE	TIME		E	F	E	E	E	E	E	E	E	← Specify Preservatives	
1.	B5-0.0-0.5-20251114	S		X	11/14/25	1030	3		X	X	X						A-HCl	D-NaOH
2.	B5-0.5-1.0-20251114			X		1035	4	X									B-HNO3	E-ICE
3.	Pure-0.0-0.5-20251114			X		1230	3		X	X	X						C-H2SO4	F-OTHER
4.	Pure-0.5-1.0-20251114			X		1220	4	X										
5.	B4-0.05-1.0-20251114			X		1215	4	X										
6.	B4-0.0-0.5-20251114			X		1225	6		X	X	X	X	X	X	X	X		
7.	B3-0.0-0.5-20251114			X		1400	3		X	X	X							
8.	B3-0.5-1.0-20251114			X		1355	4	X										
9.				X														
10.				X														

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1420 RECEIVED BY: 1420
1. 11/14/25 11/14/25
RELINQUISHED BY SAMPLER: _____ RECEIVED BY: _____
2. _____
RELINQUISHED BY SAMPLER: 1500 RECEIVED BY: 1500
3. 11-14-25 11-14-25

Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 3.9 °C
Comments: _____

Page _____ of _____ CLIENT: ☐ Hand Delivered ☐ Other _____

Shipment Complete
☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q3649	HDRIO8	Order Date : 11/14/2025 3:02:13 PM	Project Mgr :
Client Name : HDR, Inc.		Project Name : PVWC Linear Construction	Report Type : NJ Reduced
Client Contact : Andrew Wadden		Receive DateTime : 11/14/2025 3:00:00 PM	EDD Type : Excel NJ
Invoice Name : HDR, Inc.		Purchase Order :	Hard Copy Date :
Invoice Contact : Andrew Wadden			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q3649-02	B5-0.5-1.0-20251114	Solid	11/14/2025	10:35	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q3649-04	DUPE-0.5-1.0-20251114	Solid	11/14/2025	12:20	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q3649-05	B4-0.05-1.0-20251114	Solid	11/14/2025	12:15	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q3649-09	B3-0.5-1.0-20251114	Solid	11/14/2025	13:55	VOC-TCLVOA-10		8260D	10 Bus. Days	

Technotes

Relinquished By :

Date / Time :

[Signature]
11-14-25 1550

Received By :

Date / Time :

[Signature]
11-14-25

15:50

Storage Area : VOA Refridgerator Room

*Stored in rva
E8202 & ref # 06 (1)*