

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111725\
 Data File : BG064784.D
 Acq On : 17 Nov 2025 13:23
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 17 14:01:23 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 12 17:10:56 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	132	0.00
2	1,4-Dioxane	0.443	0.432	2.5	122	0.00
3	Pyridine	1.321	1.383	-4.7	134	0.00
4	n-Nitrosodimethylamine	0.672	0.661	1.6	127	0.00
5 S	2-Fluorophenol	1.145	1.171	-2.3	132	0.00
6	Aniline	2.030	2.131	-5.0	133	0.00
7 S	Phenol-d6	1.523	1.626	-6.8	136	0.00
8	2-Chlorophenol	1.258	1.329	-5.6	135	0.00
9	Benzaldehyde	0.997	0.958	3.9	123	0.00
10 C	Phenol	1.657	1.757	-6.0	135	0.00
11	bis(2-Chloroethyl)ether	1.179	1.221	-3.6	130	0.00
12	1,3-Dichlorobenzene	1.447	1.498	-3.5	132	0.00
13 C	1,4-Dichlorobenzene	1.474	1.526	-3.5	134	0.00
14	1,2-Dichlorobenzene	1.429	1.454	-1.7	130	0.00
15	Benzyl Alcohol	1.250	1.326	-6.1	134	0.00
16	2,2'-oxybis(1-Chloropropane	2.792	2.755	1.3	128	0.00
17	2-Methylphenol	1.173	1.235	-5.3	134	0.00
18	Hexachloroethane	0.558	0.554	0.7	130	0.00
19 P	n-Nitroso-di-n-propylamine	1.030	1.098	-6.6	132	0.00
20	3+4-Methylphenols	1.607	1.688	-5.0	135	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	132	0.00
22	Acetophenone	0.545	0.564	-3.5	132	0.00
23 S	Nitrobenzene-d5	0.405	0.419	-3.5	133	0.00
24	Nitrobenzene	0.409	0.429	-4.9	136	0.00
25	Isophorone	0.770	0.777	-0.9	133	0.00
26 C	2-Nitrophenol	0.184	0.199	-8.2	136	0.00
27	2,4-Dimethylphenol	0.327	0.342	-4.6	136	0.00
28	bis(2-Chloroethoxy)methane	0.423	0.446	-5.4	138	0.00
29 C	2,4-Dichlorophenol	0.351	0.375	-6.8	136	0.00
30	1,2,4-Trichlorobenzene	0.423	0.423	0.0	131	0.00
31	Naphthalene	1.134	1.150	-1.4	133	0.00
32	Benzoic acid	0.205	0.232	-13.2	150	0.01
33	4-Chloroaniline	0.454	0.480	-5.7	134	0.00
34 C	Hexachlorobutadiene	0.363	0.356	1.9	128	0.00
35	Caprolactam	0.105	0.116	-10.5	148	0.00
36 C	4-Chloro-3-methylphenol	0.390	0.410	-5.1	139	0.00
37	2-Methylnaphthalene	0.848	0.857	-1.1	134	0.00
38	1-Methylnaphthalene	0.843	0.847	-0.5	134	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	134	0.00
40	1,2,4,5-Tetrachlorobenzene	0.778	0.767	1.4	131	0.00
41 P	Hexachlorocyclopentadiene	0.558	0.514	7.9	123	0.00
42 S	2,4,6-Tribromophenol	0.390	0.407	-4.4	141	0.00
43 C	2,4,6-Trichlorophenol	0.457	0.473	-3.5	136	0.00
44	2,4,5-Trichlorophenol	0.512	0.536	-4.7	138	0.00
45 S	2-Fluorobiphenyl	1.433	1.411	1.5	133	0.00
46	1,1'-Biphenyl	1.436	1.424	0.8	133	0.00
47	2-Chloronaphthalene	1.132	1.136	-0.4	136	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111725\
 Data File : BG064784.D
 Acq On : 17 Nov 2025 13:23
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 17 14:01:23 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 12 17:10:56 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.364	0.392	-7.7	141	0.00
49	Acenaphthylene	1.933	1.969	-1.9	137	0.00
50	Dimethylphthalate	1.580	1.594	-0.9	138	0.00
51	2,6-Dinitrotoluene	0.304	0.322	-5.9	141	0.00
52 C	Acenaphthene	1.159	1.215	-4.8	140	0.00
53	3-Nitroaniline	0.300	0.334	-11.3	146	0.00
54 P	2,4-Dinitrophenol	0.158	0.188	-19.0	149	0.00
55	Dibenzofuran	1.916	1.950	-1.8	140	0.00
56 P	4-Nitrophenol	0.238	0.275	-15.5	153#	0.00
57	2,4-Dinitrotoluene	0.456	0.508	-11.4	147	0.00
58	Fluorene	1.503	1.520	-1.1	138	0.00
59	2,3,4,6-Tetrachlorophenol	0.506	0.543	-7.3	145	0.00
60	Diethylphthalate	1.513	1.563	-3.3	140	0.00
61	4-Chlorophenyl-phenylether	0.902	0.910	-0.9	138	0.00
62	4-Nitroaniline	0.311	0.352	-13.2	147	0.00
63	Azobenzene	1.259	1.293	-2.7	140	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	145	0.00
65	4,6-Dinitro-2-methylphenol	0.111	0.119	-7.2	148	0.00
66 c	n-Nitrosodiphenylamine	0.524	0.519	1.0	140	0.00
67	4-Bromophenyl-phenylether	0.263	0.254	3.4	137	0.00
68	Hexachlorobenzene	0.318	0.306	3.8	139	0.00
69	Atrazine	0.253	0.248	2.0	142	0.00
70 C	Pentachlorophenol	0.157	0.178	-13.4	161#	0.00
71	Phenanthrene	1.094	1.079	1.4	143	0.00
72	Anthracene	1.116	1.114	0.2	145	0.00
73	Carbazole	0.959	0.965	-0.6	144	0.00
74	Di-n-butylphthalate	1.070	1.054	1.5	142	0.00
75 C	Fluoranthene	1.496	1.468	1.9	143	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	131	0.00
77	Benzidine	0.599	0.514	14.2	120	0.00
78	Pyrene	1.121	1.199	-7.0	142	0.00
79 S	Terphenyl-d14	1.017	1.055	-3.7	137	0.00
80	Butylbenzylphthalate	0.394	0.412	-4.6	138	0.00
81	Benzo(a)anthracene	1.360	1.351	0.7	129	0.00
82	3,3'-Dichlorobenzidine	0.501	0.464	7.4	118	0.00
83	Chrysene	1.282	1.288	-0.5	129	0.00
84	Bis(2-ethylhexyl)phthalate	0.576	0.568	1.4	128	0.00
85 c	Di-n-octyl phthalate	0.996	0.953	4.3	120	0.00
86 I	Perylene-d12	1.000	1.000	0.0	115	0.00
87	Indeno(1,2,3-cd)pyrene	1.374	1.407	-2.4	116	0.01
88	Benzo(b)fluoranthene	1.273	1.244	2.3	111	0.00
89	Benzo(k)fluoranthene	1.279	1.279	0.0	118	0.00
90 C	Benzo(a)pyrene	1.181	1.179	0.2	114	0.00
91	Dibenzo(a,h)anthracene	1.126	1.153	-2.4	115	0.00
92	Benzo(g,h,i)perylene	1.083	1.120	-3.4	117	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG111725\
 Data File : BG064784.D
 Acq On : 17 Nov 2025 13:23
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
LabSampleId :
 SSTDCCC040

Quant Time: Nov 17 14:01:23 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 12 17:10:56 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0