

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>Alliance</u>	Contract:	<u>HDRI08</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3649</u>
Instrument ID:	<u>MSVOA_W</u>	Calibration Date(s):	<u>11/10/2025</u> <u>11/10/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Calibration Time(s):	<u>15:11</u> <u>17:00</u>
GC Column:	<u>RXI-624</u> ID: <u>0.25</u> (mm)		

LAB FILE ID:		RRF005 = VW032406.D		RRF010 = VW032407.D		RRF020 = VW032408.D		
		RRF050 = VW032409.D		RRF100 = VW032410.D		RRF150 = VW032411.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Dichlorodifluoromethane	0.288	0.282	0.231	0.248	0.243	0.230	0.254	10
Chloromethane	0.476	0.439	0.408	0.408	0.438	0.440	0.435	5.8
Vinyl Chloride	0.591	0.567	0.525	0.538	0.543	0.528	0.549	4.6
Bromomethane	0.402	0.371	0.350	0.358	0.364	0.358	0.367	5.1
Chloroethane	0.382	0.361	0.340	0.355	0.362	0.354	0.359	3.9
Trichlorofluoromethane	0.420	0.416	0.392	0.459	0.442	0.452	0.430	5.9
1,1,2-Trichlorotrifluoroethane	0.506	0.495	0.438	0.467	0.458	0.449	0.469	5.6
1,1-Dichloroethene	0.576	0.541	0.506	0.540	0.539	0.523	0.537	4.3
Acetone	0.216	0.194	0.183	0.186	0.190	0.188	0.193	6.3
Carbon Disulfide	1.604	1.574	1.497	1.608	1.644	1.603	1.588	3.1
Methyl tert-butyl Ether	1.186	1.163	1.107	1.139	1.126	1.076	1.133	3.5
Methyl Acetate	0.486	0.494	0.459	0.463	0.476	0.487	0.478	2.9
Methylene Chloride	0.789	0.741	0.682	0.640	0.658	0.629	0.690	9.1
trans-1,2-Dichloroethene	0.646	0.619	0.596	0.603	0.624	0.598	0.615	3.1
1,1-Dichloroethane	1.217	1.156	1.091	1.148	1.168	1.138	1.153	3.5
Cyclohexane	1.236	1.141	1.016	1.027	1.003	0.976	1.066	9.5
2-Butanone	0.292	0.288	0.282	0.280	0.284	0.278	0.284	1.9
Carbon Tetrachloride	0.392	0.398	0.386	0.399	0.386	0.391	0.392	1.4
cis-1,2-Dichloroethene	0.764	0.727	0.713	0.729	0.748	0.723	0.734	2.5
Bromochloromethane	0.606	0.563	0.533	0.502	0.538	0.527	0.545	6.6
Chloroform	1.164	1.165	1.089	1.109	1.141	1.103	1.129	2.9
1,1,1-Trichloroethane	0.831	0.820	0.776	0.817	0.817	0.777	0.806	2.9
Methylcyclohexane	0.560	0.575	0.562	0.598	0.562	0.572	0.571	2.5
Benzene	1.410	1.473	1.410	1.437	1.413	1.380	1.420	2.2
1,2-Dichloroethane	0.425	0.423	0.416	0.421	0.409	0.409	0.417	1.7
Trichloroethene	0.336	0.340	0.327	0.335	0.323	0.323	0.331	2.2
1,2-Dichloropropane	0.356	0.360	0.354	0.356	0.350	0.346	0.354	1.4
Bromodichloromethane	0.460	0.469	0.456	0.482	0.481	0.482	0.471	2.5
4-Methyl-2-Pentanone	0.304	0.329	0.330	0.336	0.317	0.311	0.321	3.9
Toluene	0.875	0.896	0.870	0.894	0.870	0.859	0.877	1.7

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG No.: Q3649

Instrument ID: MSVOA_W

Calibration Date(s): 11/10/2025 11/10/2025

Heated Purge: (Y/N) Y

Calibration Time(s): 15:11 17:00

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VW032406.D		RRF010 = VW032407.D		RRF020 = VW032408.D			
		RRF050 = VW032409.D		RRF100 = VW032410.D		RRF150 = VW032411.D			
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
t-1,3-Dichloropropene	0.445	0.481	0.469	0.504	0.508	0.507	0.486	5.2	
cis-1,3-Dichloropropene	0.531	0.553	0.556	0.579	0.576	0.574	0.562	3.2	
1,1,2-Trichloroethane	0.288	0.283	0.282	0.293	0.278	0.277	0.284	2.1	
2-Hexanone	0.223	0.245	0.241	0.244	0.231	0.226	0.235	4	
Dibromochloromethane	0.306	0.317	0.314	0.320	0.320	0.322	0.317	1.9	
1,2-Dibromoethane	0.261	0.292	0.279	0.281	0.272	0.272	0.276	3.8	
Tetrachloroethene	0.298	0.304	0.285	0.300	0.303	0.287	0.296	2.8	
Chlorobenzene	1.087	1.088	1.042	1.133	1.093	1.023	1.078	3.7	
Ethyl Benzene	1.875	1.853	1.822	1.944	1.915	1.779	1.865	3.2	
m/p-Xylenes	0.692	0.713	0.679	0.732	0.708	0.677	0.700	3	
o-Xylene	0.668	0.659	0.662	0.682	0.699	0.653	0.670	2.5	
Styrene	1.104	1.146	1.132	1.208	1.212	1.118	1.153	4	
Bromoform	0.187	0.201	0.193	0.203	0.205	0.196	0.198	3.5	
Isopropylbenzene	3.523	3.562	3.511	3.968	3.781	3.689	3.672	4.9	
1,1,2,2-Tetrachloroethane	0.886	0.869	0.845	0.927	0.871	0.831	0.871	3.8	
1,3-Dichlorobenzene	1.665	1.630	1.620	1.721	1.659	1.596	1.648	2.7	
1,4-Dichlorobenzene	1.776	1.680	1.644	1.774	1.658	1.602	1.689	4.2	
1,2-Dichlorobenzene	1.547	1.528	1.464	1.545	1.509	1.488	1.513	2.2	
1,2-Dibromo-3-Chloropropane	0.145	0.146	0.145	0.159	0.152	0.154	0.150	3.8	
1,2,4-Trichlorobenzene	0.943	0.917	0.923	1.032	0.972	0.992	0.963	4.6	
1,2,3-Trichlorobenzene	0.892	0.831	0.836	0.901	0.880	0.906	0.874	3.8	
1,2-Dichloroethane-d4	0.691	0.654	0.647	0.605	0.640	0.630	0.645	4.4	
Dibromofluoromethane	0.314	0.302	0.305	0.287	0.297	0.291	0.299	3.3	
Toluene-d8	1.171	1.176	1.152	1.130	1.136	1.125	1.148	1.9	
4-Bromofluorobenzene	0.444	0.444	0.454	0.415	0.426	0.421	0.434	3.6	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.