

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: HDRI08
Lab Code: ACE SDG NO.: Q3649
Lab File ID: VW032468.D Date Analyzed: 11/18/2025
Instrument ID: MSVOA_W Time Analyzed: 09:46
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	361815	7.95	661363	8.85	600329	11.63
UPPER LIMIT	723630	8.453	1322730	9.349	1200660	12.129
LOWER LIMIT	180908	7.453	330682	8.349	300165	11.129
EPA SAMPLE NO.						
B4-0.05-1.0-20251114	232615	7.96	543396	8.85	601253	11.63
B3-0.5-1.0-20251114	212864	7.97	490545	8.85	510548	11.63
VW1118SBL01	212726	7.96	511741	8.85	533709	11.63
VW1118SBS01	368383	7.96	685400	8.85	608119	11.63
VW1118SBSD01	367695	7.96	692034	8.85	614192	11.63

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.