

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>HDRI08</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3649</u>
Lab File ID:	<u>VW032490.D</u>	Date Analyzed:	<u>11/19/2025</u>
Instrument ID:	<u>MSVOA_W</u>	Time Analyzed:	<u>10:01</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>Y</u>

	IS4 AREA #	RT #				
12 HOUR STD	269243	13.55				
UPPER LIMIT	538486	14.05				
LOWER LIMIT	134622	13.05				
EPA SAMPLE NO.						
B5-0.5-1.0-20251114	242262	13.56				
DUPE-0.5-1.0-20251114	249401	13.56				
VW1119SBL01	260882	13.56				
VW1119SBS01	285477	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.