

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043211.D
 Acq On : 18 Nov 2025 13:56
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL111825

Quant Time: Nov 19 01:34:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	1.410	1.363	3.3	95	0.00
3	Chlorodifluoromethane	1.387	1.276	8.0	90	0.00
4	Chloromethane	0.527	0.479	9.1	90	0.00
5 T	Vinyl Chloride	0.577	0.490	15.1	94	0.00
6 T	Bromomethane	0.280	0.227	18.9	84	0.00
7	Chloroethane	0.225	0.203	9.8	97	0.00
8 T	Dichlorotetrafluoroethane	1.173	1.130	3.7	96	0.00
9 T	Propene	0.589	0.552	6.3	103	0.00
10 T	Heptane	1.753	1.525	13.0	88	0.00
11 T	Trichlorofluoromethane	1.367	1.377	-0.7	103	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.067	1.032	3.3	97	0.00
13	Ethanol	0.207	0.157	24.2	88	0.00
14 T	Bromoethene	0.409	0.378	7.6	97	0.00
15 T	Acetone	1.389	1.071	22.9	97	0.00
16 T	1,3-Butadiene	0.619	0.560	9.5	95	0.00
17	tert-Butyl alcohol	1.824	1.686	7.6	92	0.00
18 T	1,1-Dichloroethene	0.497	0.448	9.9	95	0.00
19 T	Isopropyl Alcohol	0.785	0.662	15.7	90	0.00
20 T	Methylene Chloride	0.600	0.422	29.7	99	0.00
21 T	Allyl Chloride	0.935	0.772	17.4	85	0.00
22 T	trans-1,2-Dichloroethene	0.568	0.513	9.7	96	0.00
23 T	Vinyl Acetate	1.677	1.432	14.6	82	0.00
24 T	1,1-Dichloroethane	1.096	1.054	3.8	96	0.00
25 T	Ethyl Acetate	2.996	2.643	11.8	89	0.00
26 T	Hexane	1.314	1.168	11.1	88	0.00
27 T	Carbon Disulfide	1.341	1.278	4.7	95	0.00
28 T	Methyl tert-Butyl Ether	0.645	0.656	-1.7	103	0.00
29 T	Chloroform	1.867	1.778	4.8	93	0.00
30 T	Cyclohexane	1.101	1.008	8.4	92	0.00
31 T	cis-1,2-Dichloroethene	1.306	1.225	6.2	92	0.00
32 T	1,1,1-Trichloroethane	1.990	1.863	6.4	92	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
34 T	2-Butanone	0.706	0.615	12.9	88	0.00
35 T	Carbon Tetrachloride	0.759	0.698	8.0	91	0.00
36 T	Benzene	0.958	0.890	7.1	91	0.00
37 T	1,2-Dichloroethane	0.560	0.547	2.3	95	0.00
38 T	Trichloroethene	0.375	0.347	7.5	91	0.00
39 T	1,2-Dichloropropane	0.351	0.318	9.4	88	0.00
40 T	1,4-Dioxane	0.158	0.123	22.2	79	0.00
41 T	Tetrahydrofuran	0.390	0.347	11.0	87	0.00
42 T	Bromodichloromethane	0.773	0.746	3.5	92	0.00
43	Methyl Methacrylate	0.508	0.444	12.6	89	0.00
44 T	2,2,4-Trimethylpentane	1.645	1.502	8.7	89	0.00
45 T	t-1,3-Dichloropropene	0.474	0.373	21.3	76	0.00
46 T	cis-1,3-Dichloropropene	0.556	0.441	20.7	76	0.00
47 T	1,1,2-Trichloroethane	0.372	0.338	9.1	91	0.01
48 T	Dibromochloromethane	0.640	0.608	5.0	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043211.D
 Acq On : 18 Nov 2025 13:56
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL111825

Quant Time: Nov 19 01:34:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.566	0.515	9.0	86	0.00
50 T	4-Methyl-2-Pentanone	0.994	0.858	13.7	84	0.00
51 T	2-Hexanone	0.772	0.636	17.6	80	0.00
52 T	Tetrachloroethene	0.356	0.316	11.2	89	0.00
53 T	Toluene	1.134	1.043	8.0	90	0.00
54 T	1,2-Dibromoethane	0.547	0.477	12.8	84	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
56	1,1,1,2-Tetrachloroethane	0.735	0.659	10.3	90	0.00
57 T	Chlorobenzene	1.027	0.922	10.2	88	0.00
58 T	Ethyl Benzene	1.925	1.764	8.4	89	0.00
59 T	m/p-Xylene	1.543	1.397	9.5	90	0.00
60 T	o-Xylene	1.542	1.368	11.3	89	0.00
61 T	Styrene	0.675	0.588	12.9	84	0.00
62	Isopropylbenzene	2.758	2.431	11.9	88	0.00
63 T	1,1,2,2-Tetrachloroethane	0.981	0.826	15.8	85	0.00
64	n-propylbenzene	0.709	0.609	14.1	85	0.00
65	tert-Butylbenzene	2.463	2.085	15.3	86	0.00
66 T	Benzyl Chloride	0.250	0.129	48.4#	49	0.00
67	sec-Butylbenzene	3.438	2.884	16.1	85	0.00
68 S	1-Bromo-4-Fluorobenzene	0.733	0.737	-0.5	92	0.00
69	p-Isopropyltoluene	2.912	2.409	17.3	83	0.00
70	n-Butylbenzene	2.825	2.373	16.0	83	0.00
71	2-Chlorotoluene	2.120	1.834	13.5	86	0.00
72 T	4-Ethyltoluene	1.877	1.648	12.2	87	0.00
73 T	1,3,5-Trimethylbenzene	1.575	1.364	13.4	86	0.00
74 T	1,2,4-Trimethylbenzene	1.719	1.455	15.4	87	0.00
75 T	1,3-Dichlorobenzene	1.003	0.835	16.7	82	0.00
76 T	1,4-Dichlorobenzene	1.003	0.830	17.2	81	0.00
77 T	1,2-Dichlorobenzene	0.965	0.768	20.4	81	0.00
78 T	Hexachloro-1,3-Butadiene	0.805	0.581	27.8	78	0.00
79 T	Naphthalene	1.527	1.263	17.3	78	0.00
80 T	Naphthalene,2-methyl-	0.401	0.346	13.7	80	0.00
81 T	1,2,4-Trichlorobenzene	0.683	0.580	15.1	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0