

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043209.D
 Acq On : 18 Nov 2025 12:49
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025

Quant Time: Nov 19 00:52:39 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	98952	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	266012	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	213547	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.371	95	152015	9.712	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.100%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.580	62	215m	0.038	ppbv	
32) 1,1,1-Trichloroethane	3.357	97	628m	0.032	ppbv	
35) Carbon Tetrachloride	3.758	117	664m	0.033	ppbv	
52) Tetrachloroethene	8.222	164	290m	0.031	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.963	83	640m	0.031	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
Data File : VL043209.D
Acq On : 18 Nov 2025 12:49
Operator : SY/MD
Sample : VSTDIC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.03

Quant Time: Nov 19 00:52:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Nov 19 00:48:15 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
Supervised By : Mahesh Dadoda 11/19/2025

