

## Report of Analysis

Client: GFE LLC  
Project: 470 TO 472 WESTBURY AVE  
Client Sample ID: OA1  
Lab Sample ID: Q3650-06  
Analytical Method: TO-15  
Sample Wt/Vol: 400 mL

Date Collected: 11/11/25  
Date Received: 11/14/25  
SDG No.: Q3650  
Matrix: Air  
Test: VOCMS Group2

| CAS Number     | Parameter              | Conc.<br>ppbv | Conc.<br>ug/M3 | Qua. | DF | MDL<br>ug/m3 | LOQ / CRQL<br>ug/m3 | Date Ana.      | BatchID  |
|----------------|------------------------|---------------|----------------|------|----|--------------|---------------------|----------------|----------|
| <b>TARGETS</b> |                        |               |                |      |    |              |                     |                |          |
| 75-01-4        | Vinyl Chloride         | 0.030         | 0.080          | U    | 1  | 0.080        | 0.080               | 11/18/25 19:58 | VL111825 |
| 142-82-5       | Heptane                | 0.17          | 0.70           | U    | 1  | 0.70         | 2.05                | 11/18/25 19:58 | VL111825 |
| 75-34-3        | 1,1-Dichloroethane     | 0.13          | 0.53           | U    | 1  | 0.53         | 2.02                | 11/18/25 19:58 | VL111825 |
| 110-82-7       | Cyclohexane            | 0.22          | 0.76           | U    | 1  | 0.76         | 1.72                | 11/18/25 19:58 | VL111825 |
| 156-59-2       | cis-1,2-Dichloroethene | 0.10          | 0.40           | U    | 1  | 0.40         | 1.98                | 11/18/25 19:58 | VL111825 |
| 71-55-6        | 1,1,1-Trichloroethane  | 0.020         | 0.11           | U    | 1  | 0.11         | 0.16                | 11/18/25 19:58 | VL111825 |
| 540-84-1       | 2,2,4-Trimethylpentane | 0.14          | 0.65           | U    | 1  | 0.65         | 2.34                | 11/18/25 19:58 | VL111825 |
| 71-43-2        | Benzene                | 0.11          | 0.35           | J    | 1  | 0.26         | 1.60                | 11/18/25 19:58 | VL111825 |
| 79-01-6        | Trichloroethene        | 0.020         | 0.11           | U    | 1  | 0.11         | 0.16                | 11/18/25 19:58 | VL111825 |
| 108-88-3       | Toluene                | 0.20          | 0.75           | J    | 1  | 0.60         | 1.88                | 11/18/25 19:58 | VL111825 |
| 127-18-4       | Tetrachloroethene      | 0.020         | 0.14           | U    | 1  | 0.14         | 0.20                | 11/18/25 19:58 | VL111825 |
| 100-41-4       | Ethyl Benzene          | 0.19          | 0.83           | U    | 1  | 0.83         | 2.17                | 11/18/25 19:58 | VL111825 |
| 179601-23-1    | m/p-Xylene             | 0.41          | 1.78           | U    | 1  | 1.78         | 4.34                | 11/18/25 19:58 | VL111825 |
| 95-47-6        | o-Xylene               | 0.21          | 0.91           | U    | 1  | 0.91         | 2.17                | 11/18/25 19:58 | VL111825 |
| 108-67-8       | 1,3,5-Trimethylbenzene | 0.18          | 0.88           | U    | 1  | 0.88         | 2.46                | 11/18/25 19:58 | VL111825 |
| 95-63-6        | 1,2,4-Trimethylbenzene | 0.18          | 0.88           | U    | 1  | 0.88         | 2.46                | 11/18/25 19:58 | VL111825 |
| 91-20-3        | Naphthalene            | 0.010         | 0.050          | U    | 1  | 0.050        | 0.52                | 11/18/25 19:58 | VL111825 |
| 110-54-3       | Hexane                 | 0.24          | 0.85           | J    | 1  | 0.56         | 1.76                | 11/18/25 19:58 | VL111825 |

### SURROGATES

|          |                         |      |  |          |      |         |
|----------|-------------------------|------|--|----------|------|---------|
| 460-00-4 | 1-Bromo-4-Fluorobenzene | 10.2 |  | 65 - 135 | 102% | SPK: 10 |
|----------|-------------------------|------|--|----------|------|---------|

### INTERNAL STANDARDS

|           | Area Count          |        |
|-----------|---------------------|--------|
| 74-97-5   | Bromochloromethane  | 92100  |
| 540-36-3  | 1,4-Difluorobenzene | 243000 |
| 3114-55-4 | Chlorobenzene-d5    | 191000 |

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements