

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q3654
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/19/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071344.D **Time Analyzed:** 17:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.496	4.397	4.597	511.570	500.000	2.3
Aroclor-1016-2	4.516	4.416	4.616	515.760	500.000	3.2
Aroclor-1016-3	4.573	4.474	4.674	504.590	500.000	0.9
Aroclor-1016-4	4.663	4.564	4.764	510.130	500.000	2.0
Aroclor-1016-5	4.948	4.849	5.049	516.900	500.000	3.4
Aroclor-1260-1	6.042	5.943	6.143	474.340	500.000	-5.1
Aroclor-1260-2	6.299	6.200	6.400	488.950	500.000	-2.2
Aroclor-1260-3	6.652	6.552	6.752	466.010	500.000	-6.8
Aroclor-1260-4	6.867	6.767	6.967	457.130	500.000	-8.6
Aroclor-1260-5	7.174	7.075	7.275	457.560	500.000	-8.5
Decachlorobiphenyl	8.536	8.437	8.637	44.700	50.000	-10.6
Tetrachloro-m-xylene	3.411	3.311	3.511	52.310	50.000	4.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q3654
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/19/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071344.D **Time Analyzed:** 17:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.870	3.771	3.971	590.240	500.000	18.0
Aroclor-1016-2	3.886	3.788	3.988	597.700	500.000	19.5
Aroclor-1016-3	4.047	3.950	4.150	577.740	500.000	15.5
Aroclor-1016-4	4.096	3.998	4.198	571.240	500.000	14.2
Aroclor-1016-5	4.293	4.195	4.395	587.860	500.000	17.6
Aroclor-1260-1	5.294	5.197	5.397	567.790	500.000	13.6
Aroclor-1260-2	5.483	5.386	5.586	568.210	500.000	13.6
Aroclor-1260-3	5.627	5.530	5.730	566.310	500.000	13.3
Aroclor-1260-4	6.091	5.993	6.193	573.540	500.000	14.7
Aroclor-1260-5	6.335	6.238	6.438	541.220	500.000	8.2
Decachlorobiphenyl	7.656	7.559	7.759	46.750	50.000	-6.5
Tetrachloro-m-xylene	2.856	2.756	2.956	58.960	50.000	17.9