



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### Initial and Continuing Calibration Verification

**Client:** Parkway Generation Operating LLC

**SDG No.:** Q3659

**Project:** Linden Gen - NJDEP WW Discharge Permit

**RunNo.:** LB137938

| Analyte                   |      | Units | Result | True Value | %<br>Recovery | Acceptance<br>Window (%R) | Analysis<br>Date |
|---------------------------|------|-------|--------|------------|---------------|---------------------------|------------------|
| Sample ID:<br>Corrosivity | ICV  | pH    | 7.00   | 7          | 100           | 90-110                    | 11/18/2025       |
| Sample ID:<br>Corrosivity | CCV1 | pH    | 2.01   | 2.00       | 101           | 90-110                    | 11/18/2025       |
| Sample ID:<br>Corrosivity | CCV2 | pH    | 12.02  | 12.00      | 100           | 90-110                    | 11/18/2025       |

### Initial and Continuing Calibration Verification

**Client:** Parkway Generation Operating LLC

**SDG No.:** Q3659

**Project:** Linden Gen - NJDEP WW Discharge Permit

**RunNo.:** LB137940

| Analyte                                    | Units | Result | True Value | %<br>Recovery | Acceptance<br>Window (%R) | Analysis<br>Date |
|--------------------------------------------|-------|--------|------------|---------------|---------------------------|------------------|
| Sample ID: <b>ICV1</b><br>Reactive Cyanide | mg/L  | 0.097  | 0.099      | 98            | 85-115                    | 11/18/2025       |
| Sample ID: <b>CCV1</b><br>Reactive Cyanide | mg/L  | 0.25   | 0.25       | 100           | 90-110                    | 11/18/2025       |
| Sample ID: <b>CCV2</b><br>Reactive Cyanide | mg/L  | 0.25   | 0.25       | 100           | 90-110                    | 11/18/2025       |

## Initial and Continuing Calibration Verification

**Client:** Parkway Generation Operating LLC

**SDG No.:** Q3659

**Project:** Linden Gen - NJDEP WW Discharge Permit

**RunNo.:** LB137963

| Analyte                                       | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|-----------------------------------------------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: <b>ICV</b><br>Hexavalent Chromium  | mg/L  | 0.492  | 0.5        | 98         | 90-110                 | 11/19/2025    |
| Sample ID: <b>CCV1</b><br>Hexavalent Chromium | mg/L  | 0.491  | 0.5        | 98         | 90-110                 | 11/19/2025    |
| Sample ID: <b>CCV2</b><br>Hexavalent Chromium | mg/L  | 0.491  | 0.5        | 98         | 90-110                 | 11/19/2025    |
| Sample ID: <b>CCV3</b><br>Hexavalent Chromium | mg/L  | 0.495  | 0.5        | 99         | 90-110                 | 11/19/2025    |

## Initial and Continuing Calibration Verification

**Client:** Parkway Generation Operating LLC

**SDG No.:** Q3659

**Project:** Linden Gen - NJDEP WW Discharge Permit

**RunNo.:** LB137963

| Analyte | Units | Result | True Value | %<br>Recovery | Acceptance<br>Window (%R) | Analysis<br>Date |
|---------|-------|--------|------------|---------------|---------------------------|------------------|
|---------|-------|--------|------------|---------------|---------------------------|------------------|