

Laboratory Control Sample Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q3659
Project:	Linden Gen - NJDEP WW Discharge Permit	Run No.:	LB137963

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170592BS							
Hexavalent Chromium	mg/Kg	20	19.5		98	1	84-110	11/19/2025