

Report of Analysis

Client:	PSEG	Date Collected:	
Project:	PSEG Ridgefield OH ROW Bridge	Date Received:	
Client Sample ID:	PB170607BS	SDG No.:	Q3660
Lab Sample ID:	PB170607BS	Matrix:	Water
Analytical Method:	NJEPH	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol:	2000 uL
Prep Method :	SW3510	Prep Date :	11/18/25
		Test:	EPH

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Datafile	Date Ana.	Prep BatchID
------------	-----------	-------	------	----	-----	------------	-------	----------	-----------	--------------

TARGETS

Total AliphaticEPH	Total AliphaticEPH	1460			66.4	180	ug/l			
Total AromaticEPH	Total AromaticEPH	0	U		0	0	ug/l			
Total EPH	Total EPH	1460			66.4	180	ug/l			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	PSEG	Date Collected:	
Project:	PSEG Ridgefield OH ROW Bridge	Date Received:	
Client Sample ID:	PB170607BS	SDG No.:	Q3660
Lab Sample ID:	PB170607BS	Matrix:	Water
Analytical Method:	NJEPH	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol:	2000 uL
Prep Method :	SW3510	Prep Date :	11/18/25
		Test:	EPH

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Datafile	Date Ana.	Prep BatchID
TARGETS										
Aliphatic C9-C12	Aliphatic C9-C12	175		1	8.28	30.0	ug/l	FG016943.D	11/18/25 17:19	PB170607
Aliphatic C12-C16	Aliphatic C12-C16	158		1	13.5	20.0	ug/l	FG016943.D	11/18/25 17:19	PB170607
Aliphatic C16-C21	Aliphatic C16-C21	272		1	8.34	30.0	ug/l	FG016943.D	11/18/25 17:19	PB170607
Aliphatic C21-C28	Aliphatic C21-C28	344		1	12.5	40.0	ug/l	FG016943.D	11/18/25 17:19	PB170607
Aliphatic C28-C40	Aliphatic C28-C40	512		1	23.8	60.0	ug/l	FG016943.D	11/18/25 17:19	PB170607
Total AliphaticEPH	Total AliphaticEPH	1460			66.4	180	ug/l			
Total AromaticEPH	Total AromaticEPH	0	U		0	0	ug/l			
Total EPH	Total EPH	1460			66.4	180	ug/l			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	PSEG	Date Collected:	
Project:	PSEG Ridgefield OH ROW Bridge	Date Received:	
Client Sample ID:	PB170607BS	SDG No.:	Q3660
Lab Sample ID:	PB170607BS	Matrix:	Water
Analytical Method:	NJEPH	% Solid:	0
Sample Wt/Vol:	1000 mL	Test:	EPH
Prep Method :	SW3510	Final Vol:	2000 uL
		Prep Date :	11/18/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Datafile	Date Ana.	Prep BatchID
TARGETS										
Aromatic C10-C12	Aromatic C10-C12	144		1	2.07	20.0	ug/l	FF016738.D	11/18/25 17:19	PB170607
Aromatic C12-C16	Aromatic C12-C16	237		1	4.26	30.0	ug/l	FF016738.D	11/18/25 17:19	PB170607
Aromatic C16-C21	Aromatic C16-C21	456		1	8.27	50.0	ug/l	FF016738.D	11/18/25 17:19	PB170607
Aromatic C21-C36	Aromatic C21-C36	797		1	23.8	80.0	ug/l	FF016738.D	11/18/25 17:19	PB170607
Total AliphaticEPH	Total AliphaticEPH	0	U		0	0	ug/l			
Total AromaticEPH	Total AromaticEPH	1630			38.4	180	ug/l			
Total EPH	Total EPH	1630			38.4	180	ug/l			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	PSEG	Date Collected:	
Project:	PSEG Ridgefield OH ROW Bridge	Date Received:	
Client Sample ID:	PB170607BS	SDG No.:	Q3660
Lab Sample ID:	PB170607BS	Matrix:	Water
Analytical Method:	NJEPH	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol:	2000 uL
Prep Method :	SW3510	Test:	EPH
		Prep Date	11/18/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS									
Aromatic C10-C12	Aromatic C10-C12	144	1		2.07	20.0	ug/l	11/18/25	PB170607
Aromatic C12-C16	Aromatic C12-C16	237	1		4.26	30.0	ug/l	11/18/25	PB170607
Aromatic C16-C21	Aromatic C16-C21	456	1		8.27	50.0	ug/l	11/18/25	PB170607
Aromatic C21-C36	Aromatic C21-C36	797	1		23.8	80.0	ug/l	11/18/25	PB170607
SURROGATES									
580-13-2	2-Bromonaphthalene (SURR)	50.1			40 - 140	100%	SPK: 50		
321-60-8	2-Fluorobiphenyl (SURR)	53.6			40 - 140	107%	SPK: 50		
84-15-1	ortho-Terphenyl (SURR)	39.5			40 - 140	79%	SPK: 50		

Quantitation Report For Aromatic EPH Range.

Lab Sample ID:	PB170607BS	Acq On:	18 Nov 2025 17:19
Client Sample ID:	PB170607BS	Operator:	YP\AJ
Data file:	FF016738.D	Misc:	
Instrument:	FID_F	ALS Vial:	74
Dilution Factor:	1	Sample Multiplier:	1.00

Compound	R.T.		Response	Conc	highest_standard	Units
Aromatic C10-C12	4.487	6.366	9534269	72.249	200	ug/ml
Aromatic C12-C16	6.367	9.073	16044621	118.703	300	ug/ml
Aromatic C16-C21	9.074	13.397	30550947	228.43	500	ug/ml
Aromatic C21-C36	13.398	18.864	46924442	398.793	800	ug/ml
Aromatic EPH	4.487	18.864	103054279	818.175		ug/ml
2-Bromonaphthalene (SURR)	8.009	8.009	5679273	50.12		ug/ml
2-Fluorobiphenyl (SURR)	8.885	8.885	3825396	53.64		ug/ml
ortho-Terphenyl (SURR)	11.952	11.952	5807625	39.55		ug/ml