

Continuing Calibration Report for SequenceID : FE112625AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE057064.D

Aliphatic C9-C12	10431462.000	60.000	3.202	6.826	173857.700	171162.641	-1.575
Aliphatic C12-C16	7203002.000	40.000	6.827	10.271	180075.050	178754.319	-0.739
Aliphatic C16-C21	10468149.000	60.000	10.272	13.644	174469.150	186505.496	6.454
Aliphatic C21-C28	13609143.000	80.000	13.645	17.314	170114.288	192766.144	11.751
Aliphatic C28-C40	20180217.000	120.000	17.315	22.241	168168.475	159314.647	-5.557
Aliphatic EPH	61891973.000	360.000	3.202	22.241	171922.147	175414.751	1.991