

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Alliance

Contract: SCIA01

ProjectID: 186 Arlington Blvd, North Arlington

Lab Code: ACE

SDG No.: Q3661

Calibration Sequence : FF102925			Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID	
85	9472915	111446	FF016603.D	
170	18165984	106859	FF016604.D	
340	37023074	108891	FF016605.D	
850	89698209	105527	FF016606.D	
1700	169764046	99861	FF016607.D	
AVG RF : 106517		% RSD : 4.074		AVG RT : 15.059