

Report of Analysis

Client: Sciacca General Contractors, LLC
Project: 186 Arlington Blvd, North Arlington
Client Sample ID: VW1118SBL01
Lab Sample ID: VW1118SBL01
Analytical Method: 8260D
Sample Wt/Vol: 5 g

Level: LOW
Final Vol: 5000 uL

Date Collected:
Date Received:
SDG No.: Q3661
Matrix: SOIL
% Solid: 100
Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	1.10	U	1	1.10	5.00	ug/Kg	11/18/25 10:13	VW111825
74-87-3	Chloromethane	1.10	U	1	1.10	5.00	ug/Kg	11/18/25 10:13	VW111825
75-01-4	Vinyl Chloride	0.79	U	1	0.79	5.00	ug/Kg	11/18/25 10:13	VW111825
74-83-9	Bromomethane	1.10	U	1	1.10	5.00	ug/Kg	11/18/25 10:13	VW111825
75-00-3	Chloroethane	1.30	U	1	1.30	5.00	ug/Kg	11/18/25 10:13	VW111825
75-69-4	Trichlorofluoromethane	1.20	U	1	1.20	5.00	ug/Kg	11/18/25 10:13	VW111825
76-13-1	1,1,2-Trichlorotrifluoroethane	1.10	U	1	1.10	5.00	ug/Kg	11/18/25 10:13	VW111825
75-35-4	1,1-Dichloroethene	1.00	U	1	1.00	5.00	ug/Kg	11/18/25 10:13	VW111825
67-64-1	Acetone	4.70	U	1	4.70	25.0	ug/Kg	11/18/25 10:13	VW111825
75-15-0	Carbon Disulfide	1.10	U	1	1.10	5.00	ug/Kg	11/18/25 10:13	VW111825
1634-04-4	Methyl tert-butyl Ether	0.73	U	1	0.73	5.00	ug/Kg	11/18/25 10:13	VW111825
79-20-9	Methyl Acetate	1.50	U	1	1.50	5.00	ug/Kg	11/18/25 10:13	VW111825
75-09-2	Methylene Chloride	3.50	U	1	3.50	10.0	ug/Kg	11/18/25 10:13	VW111825
156-60-5	trans-1,2-Dichloroethene	0.86	U	1	0.86	5.00	ug/Kg	11/18/25 10:13	VW111825
75-34-3	1,1-Dichloroethane	0.80	U	1	0.80	5.00	ug/Kg	11/18/25 10:13	VW111825
110-82-7	Cyclohexane	0.79	U	1	0.79	5.00	ug/Kg	11/18/25 10:13	VW111825
78-93-3	2-Butanone	6.50	U	1	6.50	25.0	ug/Kg	11/18/25 10:13	VW111825
56-23-5	Carbon Tetrachloride	0.97	U	1	0.97	5.00	ug/Kg	11/18/25 10:13	VW111825
156-59-2	cis-1,2-Dichloroethene	0.75	U	1	0.75	5.00	ug/Kg	11/18/25 10:13	VW111825
74-97-5	Bromochloromethane	1.20	U	1	1.20	5.00	ug/Kg	11/18/25 10:13	VW111825
67-66-3	Chloroform	0.84	U	1	0.84	5.00	ug/Kg	11/18/25 10:13	VW111825
71-55-6	1,1,1-Trichloroethane	0.93	U	1	0.93	5.00	ug/Kg	11/18/25 10:13	VW111825
108-87-2	Methylcyclohexane	0.91	U	1	0.91	5.00	ug/Kg	11/18/25 10:13	VW111825
71-43-2	Benzene	0.79	U	1	0.79	5.00	ug/Kg	11/18/25 10:13	VW111825
107-06-2	1,2-Dichloroethane	0.79	U	1	0.79	5.00	ug/Kg	11/18/25 10:13	VW111825
79-01-6	Trichloroethene	0.81	U	1	0.81	5.00	ug/Kg	11/18/25 10:13	VW111825
78-87-5	1,2-Dichloropropane	0.91	U	1	0.91	5.00	ug/Kg	11/18/25 10:13	VW111825
75-27-4	Bromodichloromethane	0.78	U	1	0.78	5.00	ug/Kg	11/18/25 10:13	VW111825
108-10-1	4-Methyl-2-Pentanone	3.60	U	1	3.60	25.0	ug/Kg	11/18/25 10:13	VW111825
108-88-3	Toluene	0.78	U	1	0.78	5.00	ug/Kg	11/18/25 10:13	VW111825
10061-02-6	t-1,3-Dichloropropene	0.65	U	1	0.65	5.00	ug/Kg	11/18/25 10:13	VW111825
10061-01-5	cis-1,3-Dichloropropene	0.62	U	1	0.62	5.00	ug/Kg	11/18/25 10:13	VW111825
79-00-5	1,1,2-Trichloroethane	0.92	U	1	0.92	5.00	ug/Kg	11/18/25 10:13	VW111825
591-78-6	2-Hexanone	3.70	U	1	3.70	25.0	ug/Kg	11/18/25 10:13	VW111825
124-48-1	Dibromochloromethane	0.87	U	1	0.87	5.00	ug/Kg	11/18/25 10:13	VW111825
106-93-4	1,2-Dibromoethane	0.88	U	1	0.88	5.00	ug/Kg	11/18/25 10:13	VW111825
127-18-4	Tetrachloroethene	1.10	U	1	1.10	5.00	ug/Kg	11/18/25 10:13	VW111825
108-90-7	Chlorobenzene	0.91	U	1	0.91	5.00	ug/Kg	11/18/25 10:13	VW111825
100-41-4	Ethyl Benzene	0.67	U	1	0.67	5.00	ug/Kg	11/18/25 10:13	VW111825
179601-23-1	m/p-Xylenes	1.20	U	1	1.20	10.0	ug/Kg	11/18/25 10:13	VW111825
95-47-6	o-Xylene	0.82	U	1	0.82	5.00	ug/Kg	11/18/25 10:13	VW111825
100-42-5	Styrene	0.71	U	1	0.71	5.00	ug/Kg	11/18/25 10:13	VW111825
75-25-2	Bromoform	0.86	U	1	0.86	5.00	ug/Kg	11/18/25 10:13	VW111825

Report of Analysis

Client: Sciacca General Contractors, LLC
Project: 186 Arlington Blvd, North Arlington
Client Sample ID: VW1118SBL01
Lab Sample ID: VW1118SBL01
Analytical Method: 8260D
Sample Wt/Vol: 5 g

Level: LOW
Final Vol: 5000 uL

Date Collected:
Date Received:
SDG No.: Q3661
Matrix: SOIL
% Solid: 100
Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	0.78	U	1	0.78	5.00	ug/Kg	11/18/25 10:13	VW111825
79-34-5	1,1,2,2-Tetrachloroethane	1.20	U	1	1.20	5.00	ug/Kg	11/18/25 10:13	VW111825
541-73-1	1,3-Dichlorobenzene	1.70	U	1	1.70	5.00	ug/Kg	11/18/25 10:13	VW111825
106-46-7	1,4-Dichlorobenzene	1.60	U	1	1.60	5.00	ug/Kg	11/18/25 10:13	VW111825
95-50-1	1,2-Dichlorobenzene	1.50	U	1	1.50	5.00	ug/Kg	11/18/25 10:13	VW111825
96-12-8	1,2-Dibromo-3-Chloropropane	1.80	U	1	1.80	5.00	ug/Kg	11/18/25 10:13	VW111825
120-82-1	1,2,4-Trichlorobenzene	3.00	U	1	3.00	5.00	ug/Kg	11/18/25 10:13	VW111825
87-61-6	1,2,3-Trichlorobenzene	3.20	U	1	3.20	5.00	ug/Kg	11/18/25 10:13	VW111825
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	64.7			63 - 155	129%	SPK: 50		
1868-53-7	Dibromofluoromethane	52.3			70 - 134	105%	SPK: 50		
2037-26-5	Toluene-d8	47.4			74 - 123	95%	SPK: 50		
460-00-4	4-Bromofluorobenzene	52.0			52 - 138	104%	SPK: 50		
INTERNAL STANDARDS									
		Area Count							
363-72-4	Pentafluorobenzene	213000							
540-36-3	1,4-Difluorobenzene	512000							
3114-55-4	Chlorobenzene-d5	534000							
3855-82-1	1,4-Dichlorobenzene-d4	261000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products