

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3661</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Sciaccia General Contractors, LLC</u>	Datafile :	<u>VW032471.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW1118SBSD01	Dichlorodifluoromethane	20	21.9	ug/Kg	110	7		64	136	20
	Chloromethane	20	22.0	ug/Kg	110	2		73	129	15
	Vinyl chloride	20	22.5	ug/Kg	113	2		73	133	15
	Bromomethane	20	22.2	ug/Kg	111	3		77	137	15
	Chloroethane	20	21.7	ug/Kg	109	3		69	130	15
	Trichlorofluoromethane	20	22.1	ug/Kg	111	19		69	134	27
	1,1,2-Trichlorotrifluoroethane	20	22.7	ug/Kg	114	4		81	123	20
	1,1-Dichloroethene	20	22.3	ug/Kg	112	2		79	121	16
	Acetone	100	100	ug/Kg	100	10		60	174	28
	Carbon disulfide	20	21.5	ug/Kg	108	3		73	125	15
	Methyl tert-butyl Ether	20	20.7	ug/Kg	104	0		77	129	20
	Methyl Acetate	20	20.0	ug/Kg	100	1		51	140	30
	Methylene Chloride	20	23.6	ug/Kg	118	3		54	158	20
	trans-1,2-Dichloroethene	20	22.1	ug/Kg	111	2		80	123	15
	1,1-Dichloroethane	20	22.6	ug/Kg	113	2		82	123	15
	Cyclohexane	20	22.0	ug/Kg	110	0		76	122	15
	2-Butanone	100	100	ug/Kg	100	0		69	131	29
	Carbon Tetrachloride	20	22.5	ug/Kg	113	5		76	129	15
	cis-1,2-Dichloroethene	20	21.6	ug/Kg	108	2		82	123	15
	Bromochloromethane	20	22.8	ug/Kg	114	0		80	127	15
	Chloroform	20	22.7	ug/Kg	114	1		82	125	15
	1,1,1-Trichloroethane	20	22.3	ug/Kg	112	3		80	126	15
	Methylcyclohexane	20	20.8	ug/Kg	104	1		77	123	15
	Benzene	20	21.8	ug/Kg	109	1		84	121	15
	1,2-Dichloroethane	20	21.8	ug/Kg	109	2		81	126	15
	Trichloroethene	20	21.3	ug/Kg	106	0		83	122	15
	1,2-Dichloropropane	20	21.8	ug/Kg	109	2		83	122	15
	Bromodichloromethane	20	21.4	ug/Kg	107	2		82	123	15
	4-Methyl-2-Pentanone	100	110	ug/Kg	110	0		70	135	26
	Toluene	20	21.9	ug/Kg	110	2		83	122	15
	t-1,3-Dichloropropene	20	20.4	ug/Kg	102	1		78	124	15
	cis-1,3-Dichloropropene	20	20.8	ug/Kg	104	0		81	122	15
	1,1,2-Trichloroethane	20	21.2	ug/Kg	106	0		82	125	15
	2-Hexanone	100	100	ug/Kg	100	0		66	138	27
	Dibromochloromethane	20	20.7	ug/Kg	104	0		79	125	15
	1,2-Dibromoethane	20	20.9	ug/Kg	104	3		80	125	16
	Tetrachloroethene	20	21.2	ug/Kg	106	3		83	125	15
	Chlorobenzene	20	21.5	ug/Kg	108	0		84	122	15
	Ethyl Benzene	20	21.9	ug/Kg	110	4		82	124	15
	m/p-Xylenes	40	44.4	ug/Kg	111	2		83	124	15
	o-Xylene	20	21.2	ug/Kg	106	1		83	123	15
	Styrene	20	21.9	ug/Kg	110	2		82	124	15
	Bromoform	20	19.3	ug/Kg	97	4		75	127	16
	Isopropylbenzene	20	21.7	ug/Kg	109	8		82	124	15
	1,1,2,2-Tetrachloroethane	20	22.3	ug/Kg	112	6		77	127	21
	1,3-Dichlorobenzene	20	22.5	ug/Kg	113	6		83	122	15
	1,4-Dichlorobenzene	20	21.6	ug/Kg	108	4		84	121	15
	1,2-Dichlorobenzene	20	21.7	ug/Kg	109	3		83	124	15



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3661 Analytical Method: SW8260D
Client: Sciacca General Contractors, LLC Datafile : VW032471.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VW1118SBSD01	1,2-Dibromo-3-Chloropropane	20	19.8	ug/Kg	99	3		66	134	20
	1,2,4-Trichlorobenzene	20	19.9	ug/Kg	100	1		78	127	20
	1,2,3-Trichlorobenzene	20	20.5	ug/Kg	103	3		70	137	16