

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: SCIA01  
 Lab Code: ACE SDG NO.: Q3661  
 Lab File ID: VW032468.D Date Analyzed: 11/18/2025  
 Instrument ID: MSVOA\_W Time Analyzed: 09:46  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 361815        | 7.95  | 661363        | 8.85  | 600329        | 11.63  |
| UPPER LIMIT    | 723630        | 8.453 | 1322730       | 9.349 | 1200660       | 12.129 |
| LOWER LIMIT    | 180908        | 7.453 | 330682        | 8.349 | 300165        | 11.129 |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| VOC            | 143859 *      | 7.96  | 306233 *      | 8.85  | 292701 *      | 11.64  |
| VW1118SBL01    | 212726        | 7.96  | 511741        | 8.85  | 533709        | 11.63  |
| VW1118SBS01    | 368383        | 7.96  | 685400        | 8.85  | 608119        | 11.63  |
| VW1118SBSD01   | 367695        | 7.96  | 692034        | 8.85  | 614192        | 11.63  |

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.