

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: SCIA01
 Lab Code: ACE SDG NO.: Q3661
 Lab File ID: VW032490.D Date Analyzed: 11/19/2025
 Instrument ID: MSVOA_W Time Analyzed: 10:01
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	346973	7.96	636403	8.85	579827	11.63
UPPER LIMIT	693946	8.459	1272810	9.349	1159650	12.129
LOWER LIMIT	173487	7.459	318202	8.349	289914	11.129
EPA SAMPLE NO.						
VOCRE	150105 *	7.97	333057	8.85	362725	11.63
VW1119SBL01	218109	7.96	508852	8.85	541801	11.63
VW1119SBS01	361405	7.96	660085	8.85	590674	11.63

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.