

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: SCIA01
 Lab Code: ACE SDG NO.: Q3661
 Lab File ID: VW032468.D Date Analyzed: 11/18/2025
 Instrument ID: MSVOA_W Time Analyzed: 09:46
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	279351	13.556				
UPPER LIMIT	558702	14.056				
LOWER LIMIT	139676	13.056				
EPA SAMPLE NO.						
VOC	141349	13.56				
VW1118SBL01	261053	13.56				
VW1118SBS01	294055	13.56				
VW1118SBSD01	277057	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.