



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: HDR, Inc.

SDG No.: Q3662

Contract: HDRI08

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	103	70 - 130	CV	11/19/2025	10:01	LB137962
CRI01	Aluminum	110	100	110	65 - 135	P	11/19/2025	12:57	LB137974
	Antimony	50.0	50.0	100	65 - 135	P	11/19/2025	12:57	LB137974
	Arsenic	23.9	20.0	120	65 - 135	P	11/19/2025	12:57	LB137974
	Barium	99.7	100	100	65 - 135	P	11/19/2025	12:57	LB137974
	Beryllium	6.43	6.0	107	65 - 135	P	11/19/2025	12:57	LB137974
	Cadmium	5.89	6.0	98	65 - 135	P	11/19/2025	12:57	LB137974
	Calcium	2090	2000	104	65 - 135	P	11/19/2025	12:57	LB137974
	Chromium	9.95	10.0	100	65 - 135	P	11/19/2025	12:57	LB137974
	Cobalt	29.4	30.0	98	65 - 135	P	11/19/2025	12:57	LB137974
	Copper	21.2	20.0	106	65 - 135	P	11/19/2025	12:57	LB137974
	Iron	103	100	103	65 - 135	P	11/19/2025	12:57	LB137974
	Lead	10.1	12.0	84	65 - 135	P	11/19/2025	12:57	LB137974
	Magnesium	2170	2000	108	65 - 135	P	11/19/2025	12:57	LB137974
	Manganese	21.8	20.0	109	65 - 135	P	11/19/2025	12:57	LB137974
	Nickel	40.0	40.0	100	65 - 135	P	11/19/2025	12:57	LB137974
	Potassium	1780	2000	89	65 - 135	P	11/19/2025	12:57	LB137974
	Selenium	18.5	20.0	92	65 - 135	P	11/19/2025	12:57	LB137974
	Silver	10.7	10.0	107	65 - 135	P	11/19/2025	12:57	LB137974
	Sodium	1470	2000	74	65 - 135	P	11/19/2025	12:57	LB137974
	Thallium	42.1	40.0	105	65 - 135	P	11/19/2025	12:57	LB137974
	Vanadium	42.5	40.0	106	65 - 135	P	11/19/2025	12:57	LB137974
	Zinc	42.5	40.0	106	65 - 135	P	11/19/2025	12:57	LB137974