

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: HDR, Inc.

SDG No.: Q3662

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170600TB		WATER		Batch Number:	PB170634		Prep Date:	11/19/2025	
	Mercury	0.76	<2	U	2.00	CV	11/20/2025	07:47	LB137970
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170634BL		WATER		Batch Number:	PB170634		Prep Date:	11/19/2025	
	Mercury	0.076	<0.2	U	0.20	CV	11/19/2025	14:39	LB137970

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: HDR, Inc.

SDG No.: Q3662

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170600TB		WATER		Batch Number:	PB170626		Prep Date:	11/19/2025	
	Arsenic	25.6	<50	U	100	P	11/19/2025	17:49	LB137974
	Barium	72.8	<250	U	500	P	11/19/2025	17:49	LB137974
	Cadmium	2.50	<15	U	30.0	P	11/19/2025	17:49	LB137974
	Chromium	24.3	<25	J	50.0	P	11/19/2025	17:49	LB137974
	Lead	11.5	<30	U	60.0	P	11/19/2025	17:49	LB137974
	Selenium	48.2	<50	U	100	P	11/19/2025	17:49	LB137974
	Silver	8.10	<25	U	50.0	P	11/19/2025	17:49	LB137974
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170626BL		WATER		Batch Number:	PB170626		Prep Date:	11/19/2025	
	Arsenic	25.6	<50	U	100	P	11/19/2025	13:36	LB137974
	Barium	72.8	<250	U	500	P	11/19/2025	13:36	LB137974
	Cadmium	2.50	<15	U	30.0	P	11/19/2025	13:36	LB137974
	Chromium	10.6	<25	U	50.0	P	11/19/2025	13:36	LB137974
	Lead	11.5	<30	U	60.0	P	11/19/2025	13:36	LB137974
	Selenium	48.2	<50	U	100	P	11/19/2025	13:36	LB137974
	Silver	8.10	<25	U	50.0	P	11/19/2025	13:36	LB137974