

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: HDR, Inc.

SDG No.: Q3662

Contract: HDRI08

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	8.44			-20	20	11/19/2025	13:02	LB137974
	Barium	3.47	6.0	58	-94	106	11/19/2025	13:02	LB137974
	Cadmium	6.34	1.0	634	-5	7	11/19/2025	13:02	LB137974
	Chromium	51.9	52.0	100	42	62	11/19/2025	13:02	LB137974
	Lead	7.25			-12	12	11/19/2025	13:02	LB137974
	Selenium	1.34			-20	20	11/19/2025	13:02	LB137974
	Silver	-6.41			-10	10	11/19/2025	13:02	LB137974
ICSAB01	Arsenic	112	104	108	88.4	120	11/19/2025	13:06	LB137974
	Barium	494	537	92	437	637	11/19/2025	13:06	LB137974
	Cadmium	992	972	102	826	1120	11/19/2025	13:06	LB137974
	Chromium	551	542	102	460	624	11/19/2025	13:06	LB137974
	Lead	57.4	49.0	117	37	61	11/19/2025	13:06	LB137974
	Selenium	54.5	46.0	118	26	66	11/19/2025	13:06	LB137974
	Silver	186	201	92	170	232	11/19/2025	13:06	LB137974