

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>HDR, Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3662</u>
Contract:	<u>HDRI08</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3652-04</u>	Client ID:	<u>TP-1DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3652-04DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	1480		1480		0		P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	53.7		53.1		1		P
Lead	ug/L	20	67.1		75.0		11		P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>HDR, Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3662</u>
Contract:	<u>HDRI08</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3652-04MS</u>	Client ID:	<u>TP-1MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3652-04MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	3470		3780		9		P
Barium	ug/L	20	2140		2170		1		P
Cadmium	ug/L	20	832		908		9		P
Chromium	ug/L	20	1780		1900		7		P
Lead	ug/L	20	4100		4480		9		P
Mercury	ug/L	20	36.5		36.6		0		CV
Selenium	ug/L	20	8580		9420		9		P
Silver	ug/L	20	320		349		9		P