

Instrument ID: FID_E

Daily Analysis Runlog For Sequence/QC Batch ID # FE103025AL

Review By	yogesh	Review On	10/30/2025 10:43:15 AM
Supervise By	Sohil	Supervise On	10/30/2025 4:01:31 PM
SubDirectory	FE103025AL	HP Acquire Method	HP Processing Method
			FE103025AL

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24768,PP24770,PP24771,PP24772,PP24773
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24771 PP24769,PP24774

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	MECL2	MECL2	FE056573.D	30 Oct 2025 09:19		YP\AJ	Ok
2	I.BLK	I.BLK	FE056574.D	30 Oct 2025 09:49		YP\AJ	Ok
3	100 PPM ALIPHATIC HC	100 PPM ALIPHATIC HC	FE056575.D	30 Oct 2025 10:20		YP\AJ	Ok
4	50 PPM ALIPHATIC HC	50 PPM ALIPHATIC HC	FE056576.D	30 Oct 2025 10:50		YP\AJ	Ok
5	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FE056577.D	30 Oct 2025 11:20		YP\AJ	Ok
6	10 PPM ALIPHATIC HC	10 PPM ALIPHATIC HC	FE056578.D	30 Oct 2025 11:51		YP\AJ	Ok
7	5 PPM ALIPHATIC HC	5 PPM ALIPHATIC HC	FE056579.D	30 Oct 2025 12:21		YP\AJ	Ok
8	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FE056580.D	30 Oct 2025 12:51		YP\AJ	Ok
9	I.BLK	I.BLK	FE056581.D	30 Oct 2025 13:21		YP\AJ	Ok
10	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FE056582.D	30 Oct 2025 13:52		YP\AJ	Ok
11	Q3485-01	TR-04-102925	FE056583.D	30 Oct 2025 14:52		YP\AJ	Ok,M
12	Q3485-02	TR-04-102925-E2	FE056584.D	30 Oct 2025 15:23		YP\AJ	Ok,M
13	Q3486-01	TP3	FE056585.D	30 Oct 2025 15:53		YP\AJ	Ok
14	Q3486-02	TP3-E2	FE056586.D	30 Oct 2025 16:25		YP\AJ	Ok
15	Q3486-03	TP4	FE056587.D	30 Oct 2025 16:56	Need 5X,Surrogate fail	YP\AJ	Dilution
16	Q3486-04	TP4-E2	FE056588.D	30 Oct 2025 17:26		YP\AJ	Ok
17	Q3486-06	TP5-E2	FE056589.D	30 Oct 2025 17:57		YP\AJ	Ok
18	Q3486-07	TP6	FE056590.D	30 Oct 2025 18:27	Surrogate fail	YP\AJ	ReRun

Instrument ID: FID_E

Daily Analysis Runlog For Sequence/QC Batch ID # FE103025AL

Review By	yogesh	Review On	10/30/2025 10:43:15 AM
Supervise By	Sohil	Supervise On	10/30/2025 4:01:31 PM
SubDirectory	FE103025AL	HP Acquire Method	HP Processing Method FE103025AL
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24768,PP24770,PP24771,PP24772,PP24773		
CCC	PP24771		
Internal Standard/PEM			
ICV/I.BLK	PP24769,PP24774		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q3486-08	TP6-E2	FE056591.D	30 Oct 2025 18:58	Surrogate fail	YP\AJ	ReRun
20	Q3488-01	MOO-25-0294-0298	FE056592.D	30 Oct 2025 19:28		YP\AJ	Ok
21	Q3488-04	MOO-25-0302-0310	FE056593.D	30 Oct 2025 19:59	Need 500X	YP\AJ	Dilution
22	Q3488-06	MOO-25-0311-0313	FE056594.D	30 Oct 2025 20:29	Need further 20X	YP\AJ	Dilution
23	Q3490-01	LAW-25-0151	FE056595.D	30 Oct 2025 21:00	Need further 10X	YP\AJ	Dilution
24	Q3488-02	MOO-25-0299	FE056596.D	30 Oct 2025 21:30	Need further 20X	YP\AJ	Dilution
25	I.BLK	I.BLK	FE056597.D	30 Oct 2025 23:01		YP\AJ	Ok
26	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FE056598.D	30 Oct 2025 23:32		YP\AJ	Ok

M : Manual Integration

Instrument ID: FID_E

Daily Analysis Runlog For Sequence/QC Batch ID # FE111925AL

Review By	rahul	Review On	11/19/2025 3:03:33 PM
Supervise By	Sohil	Supervise On	11/20/2025 4:00:25 PM
SubDirectory	FE111925AL	HP Acquire Method	HP Processing Method FE103025AL

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24768,PP24770,PP24771,PP24772,PP24773
CCC Internal Standard/PEM	PP24771
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24769,PP24774

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	MECL2	MECL2	FE056923.D	19 Nov 2025 07:04		YP\AJ	Ok
2	I.BLK	I.BLK	FE056924.D	19 Nov 2025 07:35		YP\AJ	Ok
3	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FE056925.D	19 Nov 2025 08:05		YP\AJ	Ok
4	Q3652-01DL	TP-1DL	FE056926.D	19 Nov 2025 08:52		YP\AJ	Dilution
5	Q3652-01DL2	TP-1DL2	FE056927.D	19 Nov 2025 09:22		YP\AJ	Ok,M
6	Q3656-01DL	T2-GRAVELDL	FE056928.D	19 Nov 2025 09:52		YP\AJ	Dilution
7	Q3656-01DL2	Q3656-01DL2	FE056929.D	19 Nov 2025 10:22	Need more dilution	YP\AJ	Not Ok
8	Q3656-02DL	T2-GRAVEL-E2DL	FE056930.D	19 Nov 2025 10:52		YP\AJ	Dilution
9	Q3656-02DL2	T2-GRAVEL-E2DL2	FE056931.D	19 Nov 2025 11:22		YP\AJ	Ok
10	I.BLK	I.BLK	FE056932.D	19 Nov 2025 12:22		YP\AJ	Ok
11	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FE056933.D	19 Nov 2025 12:53		YP\AJ	Ok
12	PB170632BL	PB170632BL	FE056934.D	19 Nov 2025 14:09		YP\AJ	Ok
13	PB170632BS	PB170632BS	FE056935.D	19 Nov 2025 14:39		YP\AJ	Ok
14	PB170632BSD	PB170632BSD	FE056936.D	19 Nov 2025 15:09		YP\AJ	Ok
15	Q3662-04	B1-0.0-1.0-20251117	FE056937.D	19 Nov 2025 15:39		YP\AJ	Ok
16	Q3664-01	ARS20-0001	FE056938.D	19 Nov 2025 16:09		YP\AJ	Ok
17	Q3664-02	ARS20-0001-E2	FE056939.D	19 Nov 2025 16:40		YP\AJ	Ok
18	Q3664-03	ARS20-0013	FE056940.D	19 Nov 2025 17:10		YP\AJ	Ok

Instrument ID: FID_E

Daily Analysis Runlog For Sequence/QC Batch ID # FE111925AL

Review By	rahul	Review On	11/19/2025 3:03:33 PM
Supervise By	Sohil	Supervise On	11/20/2025 4:00:25 PM
SubDirectory	FE111925AL	HP Acquire Method	HP Processing Method
			FE103025AL
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24768,PP24770,PP24771,PP24772,PP24773		
CCC	PP24771		
Internal Standard/PEM			
ICV/I.BLK	PP24769,PP24774		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q3664-04	ARS20-0013-E2	FE056941.D	19 Nov 2025 17:40		YP\AJ	Ok
20	Q3664-05	291399	FE056942.D	19 Nov 2025 18:10		YP\AJ	Ok
21	Q3664-06	291399-E2	FE056943.D	19 Nov 2025 18:40		YP\AJ	Ok
22	Q3669-01	EO-01-11182025	FE056944.D	19 Nov 2025 19:11	Need 2X	YP\AJ	Dilution
23	Q3669-01D	Q3669-01D	FE056945.D	19 Nov 2025 19:40		YP\AJ	Ok,M
24	Q3669-01MS	EO-01-11182025MS	FE056946.D	19 Nov 2025 20:11	FE056944.D	YP\AJ	Ok,M
25	Q3669-01MSD	EO-01-11182025MSD	FE056947.D	19 Nov 2025 20:41	FE056944.D!FE056946.D	YP\AJ	Ok,M
26	Q3669-02	EO-01-11182025-E2	FE056948.D	19 Nov 2025 21:11	Need 5X	YP\AJ	Dilution
27	Q3656-01DL2	Q3656-01DL2	FE056949.D	19 Nov 2025 21:41	Carry over	YP\AJ	Not Ok
28	I.BLK	I.BLK	FE056950.D	19 Nov 2025 22:42		YP\AJ	Ok
29	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FE056951.D	19 Nov 2025 23:12		YP\AJ	Ok

M : Manual Integration