

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** HDRI08
Lab Code: ACE **SDG NO.:** Q3662
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/19/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115205.D **Time Analyzed:** 09:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.481	5.387	5.587	549.520	500.000	9.9
Aroclor-1016-2	5.502	5.408	5.608	545.500	500.000	9.1
Aroclor-1016-3	5.564	5.470	5.670	574.770	500.000	15.0
Aroclor-1016-4	5.659	5.566	5.766	560.670	500.000	12.1
Aroclor-1016-5	5.951	5.858	6.058	536.370	500.000	7.3
Aroclor-1260-1	7.064	6.971	7.171	520.230	500.000	4.0
Aroclor-1260-2	7.318	7.225	7.425	559.820	500.000	12.0
Aroclor-1260-3	7.675	7.582	7.782	559.310	500.000	11.9
Aroclor-1260-4	7.897	7.803	8.003	542.930	500.000	8.6
Aroclor-1260-5	8.204	8.110	8.310	537.050	500.000	7.4
Decachlorobiphenyl	9.920	9.830	10.030	59.010	50.000	18.0
Tetrachloro-m-xylene	4.341	4.248	4.448	56.960	50.000	13.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** HDRI08
Lab Code: ACE **SDG NO.:** Q3662
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/19/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115205.D **Time Analyzed:** 09:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.721	4.628	4.828	532.660	500.000	6.5
Aroclor-1016-2	4.740	4.647	4.847	529.490	500.000	5.9
Aroclor-1016-3	4.915	4.821	5.021	534.350	500.000	6.9
Aroclor-1016-4	4.957	4.864	5.064	532.560	500.000	6.5
Aroclor-1016-5	5.169	5.076	5.276	533.160	500.000	6.6
Aroclor-1260-1	6.198	6.106	6.306	494.860	500.000	-1.0
Aroclor-1260-2	6.386	6.294	6.494	477.740	500.000	-4.5
Aroclor-1260-3	6.539	6.446	6.646	492.290	500.000	-1.5
Aroclor-1260-4	7.009	6.917	7.117	505.920	500.000	1.2
Aroclor-1260-5	7.250	7.158	7.358	509.780	500.000	2.0
Decachlorobiphenyl	8.632	8.541	8.741	56.520	50.000	13.0
Tetrachloro-m-xylene	3.645	3.548	3.748	53.850	50.000	7.7