

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Contract: HDRI08
 Lab Code: ACE SDG NO.: Q3662
 GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/04/2025 11/04/2025

Client Sample No.: CCAL02 Date Analyzed: 11/19/2025
 Lab Sample No.: AR1660CCC500 Data File : PO115220.D Time Analyzed: 15:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.481	5.387	5.587	523.910	500.000	4.8
Aroclor-1016-2	5.502	5.408	5.608	514.450	500.000	2.9
Aroclor-1016-3	5.564	5.470	5.670	537.140	500.000	7.4
Aroclor-1016-4	5.660	5.566	5.766	525.560	500.000	5.1
Aroclor-1016-5	5.951	5.858	6.058	509.520	500.000	1.9
Aroclor-1260-1	7.064	6.971	7.171	490.940	500.000	-1.8
Aroclor-1260-2	7.318	7.225	7.425	523.640	500.000	4.7
Aroclor-1260-3	7.676	7.582	7.782	530.550	500.000	6.1
Aroclor-1260-4	7.897	7.803	8.003	509.250	500.000	1.9
Aroclor-1260-5	8.204	8.110	8.310	514.380	500.000	2.9
Decachlorobiphenyl	9.919	9.830	10.030	55.840	50.000	11.7
Tetrachloro-m-xylene	4.342	4.248	4.448	54.770	50.000	9.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Contract: HDRI08
 Lab Code: ACE SDG NO.: Q3662
 GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/04/2025 11/04/2025

Client Sample No.: CCAL02 Date Analyzed: 11/19/2025
 Lab Sample No.: AR1660CCC500 Data File : PO115220.D Time Analyzed: 15:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.721	4.628	4.828	517.180	500.000	3.4
Aroclor-1016-2	4.741	4.647	4.847	505.300	500.000	1.1
Aroclor-1016-3	4.915	4.821	5.021	520.980	500.000	4.2
Aroclor-1016-4	4.957	4.864	5.064	519.780	500.000	4.0
Aroclor-1016-5	5.169	5.076	5.276	518.360	500.000	3.7
Aroclor-1260-1	6.199	6.106	6.306	469.130	500.000	-6.2
Aroclor-1260-2	6.386	6.294	6.494	453.920	500.000	-9.2
Aroclor-1260-3	6.539	6.446	6.646	464.350	500.000	-7.1
Aroclor-1260-4	7.009	6.917	7.117	468.040	500.000	-6.4
Aroclor-1260-5	7.251	7.158	7.358	474.470	500.000	-5.1
Decachlorobiphenyl	8.632	8.541	8.741	52.550	50.000	5.1
Tetrachloro-m-xylene	3.645	3.548	3.748	53.210	50.000	6.4