

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** HDRI08
Lab Code: ACE **SDG NO.:** Q3662
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/19/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071332.D **Time Analyzed:** 13:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.497	4.397	4.597	481.490	500.000	-3.7
Aroclor-1016-2	4.516	4.416	4.616	488.020	500.000	-2.4
Aroclor-1016-3	4.573	4.474	4.674	485.270	500.000	-2.9
Aroclor-1016-4	4.664	4.564	4.764	487.200	500.000	-2.6
Aroclor-1016-5	4.949	4.849	5.049	488.530	500.000	-2.3
Aroclor-1260-1	6.042	5.943	6.143	464.240	500.000	-7.2
Aroclor-1260-2	6.299	6.200	6.400	486.320	500.000	-2.7
Aroclor-1260-3	6.652	6.552	6.752	484.460	500.000	-3.1
Aroclor-1260-4	6.867	6.767	6.967	489.140	500.000	-2.2
Aroclor-1260-5	7.174	7.075	7.275	494.950	500.000	-1.0
Decachlorobiphenyl	8.536	8.437	8.637	49.110	50.000	-1.8
Tetrachloro-m-xylene	3.411	3.311	3.511	49.280	50.000	-1.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** HDRI08
Lab Code: ACE **SDG NO.:** Q3662
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/19/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071332.D **Time Analyzed:** 13:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	3.870	3.771	3.971	537.530	500.000	7.5
Aroclor-1016-2	3.886	3.788	3.988	553.750	500.000	10.8
Aroclor-1016-3	4.048	3.950	4.150	550.090	500.000	10.0
Aroclor-1016-4	4.097	3.998	4.198	536.250	500.000	7.3
Aroclor-1016-5	4.294	4.195	4.395	544.870	500.000	9.0
Aroclor-1260-1	5.295	5.197	5.397	539.360	500.000	7.9
Aroclor-1260-2	5.484	5.386	5.586	547.740	500.000	9.5
Aroclor-1260-3	5.628	5.530	5.730	539.920	500.000	8.0
Aroclor-1260-4	6.092	5.993	6.193	544.730	500.000	8.9
Aroclor-1260-5	6.336	6.238	6.438	543.770	500.000	8.8
Decachlorobiphenyl	7.657	7.559	7.759	54.640	50.000	9.3
Tetrachloro-m-xylene	2.855	2.756	2.956	54.730	50.000	9.5