

Surrogate Summary

SDG No.: **Q3662**

Client: **HDR, Inc.**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO114870.D	PIBLK-PO114870.D	Tetrachloro-m-xyl	1	20	15.5	77		70 (60)	130 (140)
		Decachlorobiphen	1	20	16.8	84		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	16.2	81		70 (60)	130 (140)
		Decachlorobiphen	2	20	17.3	86		70 (60)	130 (140)
I.BLK-PO115209.D	PIBLK-PO115209.D	Tetrachloro-m-xyl	1	20	25.2	126		70 (60)	130 (140)
		Decachlorobiphen	1	20	26.6	133	*	70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	24.7	124		70 (60)	130 (140)
		Decachlorobiphen	2	20	27.3	136	*	70 (60)	130 (140)
Q3662-02	B3-0.0-1.0-20251117	Tetrachloro-m-xyl	1	20	18.4	92		30 (21)	150 (165)
		Decachlorobiphen	1	20	15.1	75		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	19.6	98		30 (21)	150 (165)
		Decachlorobiphen	2	20	16.1	81		30 (10)	150 (170)
Q3662-04	B1-0.0-1.0-20251117	Tetrachloro-m-xyl	1	20	18.5	93		30 (21)	150 (165)
		Decachlorobiphen	1	20	17.4	87		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	19.6	98		30 (21)	150 (165)
		Decachlorobiphen	2	20	14.7	73		30 (10)	150 (170)
I.BLK-PO115224.D	PIBLK-PO115224.D	Tetrachloro-m-xyl	1	20	22.5	112		70 (60)	130 (140)
		Decachlorobiphen	1	20	22.6	113		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	22.2	111		70 (60)	130 (140)
		Decachlorobiphen	2	20	23.2	116		70 (60)	130 (140)
I.BLK-PQ071272.D	PIBLK-PQ071272.D	Tetrachloro-m-xyl	1	20	21.9	110		70 (60)	130 (140)
		Decachlorobiphen	1	20	22.2	111		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	22.2	111		70 (60)	130 (140)
		Decachlorobiphen	2	20	21.7	109		70 (60)	130 (140)
I.BLK-PQ071333.D	PIBLK-PQ071333.D	Tetrachloro-m-xyl	1	20	22.8	114		70 (60)	130 (140)
		Decachlorobiphen	1	20	22.9	115		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	26.5	132	*	70 (60)	130 (140)
		Decachlorobiphen	2	20	25.6	128		70 (60)	130 (140)
PB170627BL	PB170627BL	Tetrachloro-m-xyl	1	20	18.4	92		30 (21)	150 (165)
		Decachlorobiphen	1	20	18.6	93		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	21.5	107		30 (21)	150 (165)
		Decachlorobiphen	2	20	21.0	105		30 (10)	150 (170)
PB170627BS	PB170627BS	Tetrachloro-m-xyl	1	20	18.6	93		30 (21)	150 (165)
		Decachlorobiphen	1	20	18.5	93		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	19.8	99		30 (21)	150 (165)
		Decachlorobiphen	2	20	20.5	102		30 (10)	150 (170)
Q3669-01MS	EO-01-11182025MS	Tetrachloro-m-xyl	1	20	10.9	54		30 (21)	150 (165)
		Decachlorobiphen	1	20	12.2	61		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	12.7	64		30 (21)	150 (165)
		Decachlorobiphen	2	20	8.99	45		30 (10)	150 (170)
Q3669-01MSD	EO-01-11182025MSD	Tetrachloro-m-xyl	1	20	11.1	56		30 (21)	150 (165)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: Q3662

Client: HDR, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q3669-01MSD	EO-01-11182025MSD	Decachlorobiphen	1	20	12.2	61		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	13.1	65		30 (21)	150 (165)
		Decachlorobiphen	2	20	8.74	44		30 (10)	150 (170)
I.BLK-PQ071345.D	PIBLK-PQ071345.D	Tetrachloro-m-xyl	1	20	22.1	111		70 (60)	130 (140)
		Decachlorobiphen	1	20	20.1	101		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	25.9	129		70 (60)	130 (140)
		Decachlorobiphen	2	20	21.3	106		70 (60)	130 (140)