

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF110525

Review By	rahul	Review On	11/5/2025 2:56:50 PM
Supervise By	Jagrut	Supervise On	11/6/2025 10:21:33 AM
SubDirectory	BF110525	HP Acquire Method	BNA_F
		HP Processing Method	BF110525

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13181,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF144168.D	05 Nov 2025 12:21		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF144169.D	05 Nov 2025 12:50		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF144170.D	05 Nov 2025 13:20		RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BF144171.D	05 Nov 2025 13:50		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF144172.D	05 Nov 2025 14:20		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF144173.D	05 Nov 2025 14:50	Compound #41 kept on QR	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF144174.D	05 Nov 2025 15:49		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF144175.D	05 Nov 2025 16:19		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF144176.D	05 Nov 2025 16:49		RC/JU	Ok
10	SSTDICV040	ICVBF110525	BF144177.D	05 Nov 2025 17:21		RC/JU	Ok,M
11	PB170372BL	PB170372BL	BF144178.D	05 Nov 2025 18:10		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112025

Review By	Jagrut	Review On	11/21/2025 3:35:02 PM
Supervise By	mohammad	Supervise On	11/22/2025 12:06:20 AM
SubDirectory	BF112025	HP Acquire Method	BNA_F
		HP Processing Method	BF110525

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13183,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF144289.D	20 Nov 2025 12:03		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF144290.D	20 Nov 2025 12:32		RC/JU	Ok
3	PB170631BL	PB170631BL	BF144291.D	20 Nov 2025 13:25		RC/JU	Ok
4	PB170631BS	PB170631BS	BF144292.D	20 Nov 2025 13:54		RC/JU	Ok,M
5	PB170652BL	PB170652BL	BF144293.D	20 Nov 2025 14:24		RC/JU	Ok
6	PB170652BS	PB170652BS	BF144294.D	20 Nov 2025 14:54		RC/JU	Ok,M
7	Q3669-01	EO-01-11182025	BF144295.D	20 Nov 2025 15:30		RC/JU	Ok,M
8	Q3662-04	B1-0.0-1.0-20251117	BF144296.D	20 Nov 2025 15:59		RC/JU	Ok,M
9	Q3667-01	RIGHT-SIDE-WALL	BF144297.D	20 Nov 2025 16:28		RC/JU	Ok
10	Q3667-05	LEFT-SIDE-FLOOR	BF144298.D	20 Nov 2025 16:58		RC/JU	Ok
11	Q3681-01	ETGI-354	BF144299.D	20 Nov 2025 17:27		RC/JU	Ok,M
12	Q3681-01MS	ETGI-354MS	BF144300.D	20 Nov 2025 17:57	Comp. #54 recovery is zero, Need noncon	RC/JU	Ok
13	Q3681-01MSD	ETGI-354MSD	BF144301.D	20 Nov 2025 18:27	Comp. #54 recovery is zero, Need noncon	RC/JU	Ok
14	Q3662-02	B3-0.0-1.0-20251117	BF144302.D	20 Nov 2025 18:56		RC/JU	Ok,M
15	Q3662-02MS	B3-0.0-1.0-20251117MS	BF144303.D	20 Nov 2025 19:26		RC/JU	Ok
16	Q3662-02MSD	B3-0.0-1.0-20251117MSD	BF144304.D	20 Nov 2025 19:56	Comp. #54 recovery is zero, Need noncon	RC/JU	Ok

M : Manual Integration