

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3662

Client ID : SSTDCCC040

Date Analyzed: 11/20/2025

Lab File ID: BF144290.D

Time Analyzed: 12:32

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	59614	6.863	228905	8.15	123999	9.90
UPPER LIMIT	119228	7.363	457810	8.645	247998	10.404
LOWER LIMIT	29807	6.363	114453	7.645	61999.5	9.404
EPA SAMPLE NO.						
01 PB170631BL	60075	6.86	236262	8.15	137092	9.90
02 PB170631BS	64956	6.86	256491	8.15	147675	9.90
03 B1-0.0-1.0-20251117	39264	6.86	131569	8.15	67150	9.90
04 B3-0.0-1.0-20251117	46510	6.86	165994	8.15	89354	9.90
05 B3-0.0-1.0-20251117MS	47796	6.86	175247	8.15	88326	9.90
06 B3-0.0-1.0-20251117MSD	44073	6.86	159722	8.15	80476	9.90

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.