

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111025\
 Data File : VW032409.D
 Acq On : 10 Nov 2025 16:17
 Operator : SY/MD
 Sample : VSTDICCC050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 11/11/2025
 Supervised By :Semsettin Yesilyurt 11/11/2025

Quant Time: Nov 11 03:35:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 11 03:33:05 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	473870	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	839187	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.635	117	738113	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	330535	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	286867	46.956	ug/l	0.00
Spiked Amount 50.000	Range 63 - 155		Recovery =	93.920%		
35) Dibromofluoromethane	7.904	113	240784	47.923	ug/l	0.00
Spiked Amount 50.000	Range 70 - 134		Recovery =	95.840%		
50) Toluene-d8	10.325	98	948663	49.216	ug/l	0.00
Spiked Amount 50.000	Range 74 - 123		Recovery =	98.440%		
62) 4-Bromofluorobenzene	12.617	95	348133	47.811	ug/l	0.00
Spiked Amount 50.000	Range 17 - 146		Recovery =	95.620%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.046	85	117696	48.496	ug/l	100
3) Chloromethane	2.253	50	193400	46.930	ug/l	100
4) Vinyl Chloride	2.405	62	255152	49.059	ug/l	100
5) Bromomethane	2.826	94	169422	48.699	ug/l	100
6) Chloroethane	2.972	64	168250	49.446	ug/l	100
7) Trichlorofluoromethane	3.308	101	217703	53.389	ug/l	100
8) Diethyl Ether	3.722	74	161055	48.309	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.106	101	221350	49.832	ug/l	100
10) Methyl Iodide	4.307	142	352515	49.273	ug/l	100
11) Tert butyl alcohol	5.216	59	127100	240.595	ug/l	100
12) 1,1-Dichloroethene	4.082	96	255938	50.249	ug/l	100
13) Acrolein	3.929	56	213422	240.790	ug/l	100
14) Allyl chloride	4.704	41	493300	50.363	ug/l	100
15) Acrylonitrile	5.405	53	480718	249.389	ug/l	100
16) Acetone	4.161	43	440757	241.146	ug/l	100
17) Carbon Disulfide	4.423	76	761956	50.614	ug/l	100
18) Methyl Acetate	4.710	43	219594	48.517	ug/l	100
19) Methyl tert-butyl Ether	5.460	73	539886	50.284	ug/l	100
20) Methylene Chloride	4.954	84	303080	46.367	ug/l	100
21) trans-1,2-Dichloroethene	5.466	96	285957	49.100	ug/l	100
22) Diisopropyl ether	6.337	45	987895	49.536	ug/l	100
23) Vinyl Acetate	6.283	43	3459955	253.227	ug/l	100
24) 1,1-Dichloroethane	6.246	63	543885	49.778	ug/l	100
25) 2-Butanone	7.191	43	663205	246.293	ug/l	100
26) 2,2-Dichloropropane	7.185	77	293455	48.192	ug/l	100
27) cis-1,2-Dichloroethene	7.191	96	345581	49.660	ug/l	100
28) Bromochloromethane	7.532	49	237900	46.067	ug/l	100
29) Tetrahydrofuran	7.551	42	429579	251.347	ug/l	100
30) Chloroform	7.691	83	525290	49.110	ug/l	100
31) Cyclohexane	7.971	56	486431	48.128	ug/l	100
32) 1,1,1-Trichloroethane	7.892	97	387016	50.649	ug/l	100
36) 1,1-Dichloropropene	8.099	75	382172	51.624	ug/l	100
37) Ethyl Acetate	7.270	43	283001	51.410	ug/l	100
38) Carbon Tetrachloride	8.081	117	334794	50.897	ug/l	100
39) Methylcyclohexane	9.343	83	501951	52.340	ug/l	100
40) Benzene	8.337	78	1206225	50.596	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111025\
 Data File : VW032409.D
 Acq On : 10 Nov 2025 16:17
 Operator : SY/MD
 Sample : VSTDICCC050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 11/11/2025
 Supervised By :Semsettin Yesilyurt 11/11/2025

Quant Time: Nov 11 03:35:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 11 03:33:05 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.496	41	157157	51.970	ug/l	100
42) 1,2-Dichloroethane	8.410	62	353418	50.470	ug/l	100
43) Isopropyl Acetate	8.435	43	506886	52.003	ug/l	100
44) Trichloroethene	9.099	130	280938	50.643	ug/l	100
45) 1,2-Dichloropropane	9.374	63	298715	50.321	ug/l	100
46) Dibromomethane	9.465	93	177243	51.095	ug/l	100
47) Bromodichloromethane	9.654	83	404213	51.082	ug/l	100
48) Methyl methacrylate	9.441	41	254032	54.845	ug/l	100
49) 1,4-Dioxane	9.465	88	51534	1060.466	ug/l #	100
51) 4-Methyl-2-Pentanone	10.209	43	1411536	261.951	ug/l	100
52) Toluene	10.392	92	750381	50.957	ug/l	100
53) t-1,3-Dichloropropene	10.605	75	422565	51.852	ug/l	100
54) cis-1,3-Dichloropropene	10.075	75	485529	51.519	ug/l	100
55) 1,1,2-Trichloroethane	10.788	97	245652	51.620	ug/l	100
56) Ethyl methacrylate	10.648	69	384615	52.675	ug/l	100
57) 1,3-Dichloropropane	10.934	76	438483	51.590	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.928	63	947852	243.759	ug/l	100
59) 2-Hexanone	10.971	43	1024540	259.873	ug/l	100
60) Dibromochloromethane	11.129	129	268368	50.499	ug/l	100
61) 1,2-Dibromoethane	11.239	107	235648	50.842	ug/l	100
64) Tetrachloroethene	10.867	164	221732	50.729	ug/l	100
65) Chlorobenzene	11.654	112	836572	52.585	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.727	131	249919	50.760	ug/l	100
67) Ethyl Benzene	11.727	91	1434522	52.113	ug/l	100
68) m/p-Xylenes	11.836	106	1080126	104.508	ug/l	100
69) o-Xylene	12.166	106	503049	50.833	ug/l	100
70) Styrene	12.178	104	891578	52.373	ug/l	100
71) Bromoform	12.349	173	149754	51.352	ug/l	100
73) Isopropylbenzene	12.464	105	1311412	54.022	ug/l	100
74) N-amyl acetate	12.269	43	463709	54.971	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.714	83	306330	53.171	ug/l	100
76) 1,2,3-Trichloropropane	12.763	75	221825m	48.938	ug/l	
77) Bromobenzene	12.745	156	311375	54.174	ug/l	100
78) n-propylbenzene	12.800	91	1592697	53.874	ug/l	100
79) 2-Chlorotoluene	12.891	91	949870	52.633	ug/l	100
80) 1,3,5-Trimethylbenzene	12.940	105	1071434	53.498	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.513	75	108494	54.830	ug/l	100
82) 4-Chlorotoluene	12.983	91	994389	52.976	ug/l	100
83) tert-Butylbenzene	13.202	119	952056	54.647	ug/l	100
84) 1,2,4-Trimethylbenzene	13.245	105	1087654	53.395	ug/l	100
85) sec-Butylbenzene	13.379	105	1387291	53.468	ug/l	100
86) p-Isopropyltoluene	13.495	119	1135093	52.985	ug/l	100
87) 1,3-Dichlorobenzene	13.495	146	568988	52.213	ug/l	100
88) 1,4-Dichlorobenzene	13.574	146	586517	52.535	ug/l	100
89) n-Butylbenzene	13.818	91	1119974	53.512	ug/l	100
90) Hexachloroethane	14.086	117	206125	53.365	ug/l	100
91) 1,2-Dichlorobenzene	13.867	146	510688	51.050	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.482	75	52673	53.072	ug/l	100
93) 1,2,4-Trichlorobenzene	15.129	180	341171	53.593	ug/l	100
94) Hexachlorobutadiene	15.226	225	139533	52.766	ug/l	100
95) Naphthalene	15.360	128	858633	54.088	ug/l	100
96) 1,2,3-Trichlorobenzene	15.549	180	297797	51.526	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111025\
Data File : VW032409.D
Acq On : 10 Nov 2025 16:17
Operator : SY/MD
Sample : VSTDICCC050
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :Amit Patel 11/11/2025
Supervised By :Semsettin Yesilyurt 11/11/2025

Quant Time: Nov 11 03:35:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 11 03:33:05 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111025\
Data File : VW032409.D
Acq On : 10 Nov 2025 16:17
Operator : SY/MD
Sample : VSTDICCC050
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VSTDICCC050

Quant Time: Nov 11 03:35:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 11 03:33:05 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Amit Patel 11/11/2025
Supervised By :Semsettin Yesilyurt 11/11/2025

