



Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111825\
Data File : VW032467.D
Acq On : 18 Nov 2025 09:15
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
Title : SW846 8260
Last Update : Tue Nov 11 03:48:21 2025



AutoFind: Scans 1759, 1760, 1761; Background Corrected with Scan 1750

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.3	48973	PASS
75	95	30	60	49.3	132197	PASS
95	95	100	100	100.0	267925	PASS
96	95	5	9	6.5	17523	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	64.4	172437	PASS
175	174	5	9	7.7	13242	PASS
176	174	95	101	95.7	164992	PASS
177	176	5	9	6.3	10340	PASS