

Surrogate Summary

SDG No.: **Q3662**

Client: **HDR, Inc.**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3662-01	B3-0.5-1.0-20251117	1,2-Dichloroethane-d4	50	55.3	111		70 (63)	130 (155)
		Dibromofluoromethane	50	47.5	95		70 (70)	130 (134)
		Toluene-d8	50	43.5	87		70 (74)	130 (123)
		4-Bromofluorobenzene	50	52.1	104		70 (52)	130 (138)
Q3662-03	B1-0.5-1.0-20251117	1,2-Dichloroethane-d4	50	57.3	115		70 (63)	130 (155)
		Dibromofluoromethane	50	47.8	95		70 (70)	130 (134)
		Toluene-d8	50	44.4	89		70 (74)	130 (123)
		4-Bromofluorobenzene	50	56.1	112		70 (52)	130 (138)
VW1118SBL01	VW1118SBL01	1,2-Dichloroethane-d4	50	64.7	129		70 (63)	130 (155)
		Dibromofluoromethane	50	52.3	105		70 (70)	130 (134)
		Toluene-d8	50	47.4	95		70 (74)	130 (123)
		4-Bromofluorobenzene	50	52.0	104		70 (52)	130 (138)
VW1118SBS01	VW1118SBS01	1,2-Dichloroethane-d4	50	53.7	107		70 (63)	130 (155)
		Dibromofluoromethane	50	52.0	104		70 (70)	130 (134)
		Toluene-d8	50	52.0	104		70 (74)	130 (123)
		4-Bromofluorobenzene	50	51.7	103		70 (52)	130 (138)
VW1118SBSD0	VW1118SBSD01	1,2-Dichloroethane-d4	50	52.1	104		70 (63)	130 (155)
		Dibromofluoromethane	50	50.5	101		70 (70)	130 (134)
		Toluene-d8	50	51.2	102		70 (74)	130 (123)
		4-Bromofluorobenzene	50	49.7	99		70 (52)	130 (138)

