

Report of Analysis

Client:	PSEG	Date Collected:	
Project:	Orange Gas and Appliance Service (EO)	Date Received:	
Client Sample ID:	PB170632BS	SDG No.:	Q3669
Lab Sample ID:	PB170632BS	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	100
Sample Wt/Vol:	30.03 g	Test:	EPH_NF
Prep Method :		Final Vol:	2000 uL
		Prep Date :	11/19/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Datafile	Date Ana.	Prep BatchID
------------	-----------	-------	------	----	-----	------------	-------	----------	-----------	--------------

TARGETS

Total AliphaticEPH	Total AliphaticEPH	93.4			2.09	5.99	mg/kg			
Total EPH	Total EPH	93.4			2.09	5.99	mg/kg			

* As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C40 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C40 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	PSEG	Date Collected:	
Project:	Orange Gas and Appliance Service (EO)	Date Received:	
Client Sample ID:	PB170632BS	SDG No.:	Q3669
Lab Sample ID:	PB170632BS	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	100
Sample Wt/Vol:	30.03 g	Test:	EPH_NF
Prep Method :		Final Vol:	2000 uL
		Prep Date :	11/19/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Datafile	Date Ana.	Prep BatchID
------------	-----------	-------	------	----	-----	------------	-------	----------	-----------	--------------

TARGETS

Total AliphaticEPH	Total AliphaticEPH	93.4			2.09	5.99	mg/kg			
Total EPH	Total EPH	93.4			2.09	5.99	mg/kg			

* As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C40 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C40 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	PSEG	Date Collected:	
Project:	Orange Gas and Appliance Service (EO)	Date Received:	
Client Sample ID:	PB170632BS	SDG No.:	Q3669
Lab Sample ID:	PB170632BS	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	100
Sample Wt/Vol:	30.03 g	Test:	EPH_NF
Prep Method :		Final Vol:	2000 uL
		Prep Date :	11/19/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Datafile	Date Ana.	Prep BatchID
------------	-----------	-------	------	----	-----	------------	-------	----------	-----------	--------------

TARGETS

Total AliphaticEPH	Total AliphaticEPH	93.4			2.09	5.99	mg/kg			
Total EPH	Total EPH	93.4			2.09	5.99	mg/kg			

* As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C40 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C40 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	PSEG	Date Collected:	
Project:	Orange Gas and Appliance Service (EO)	Date Received:	
Client Sample ID:	PB170632BS	SDG No.:	Q3669
Lab Sample ID:	PB170632BS	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	100
Sample Wt/Vol:	30.03 g	Test:	EPH_NF
Prep Method :			
	Final Vol: 2000 uL		
	Prep Date : 11/19/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Datafile	Date Ana.	Prep BatchID
------------	-----------	-------	------	----	-----	------------	-------	----------	-----------	--------------

TARGETS

Total AliphaticEPH	Total AliphaticEPH	93.4			2.09	5.99	mg/kg			
Total EPH	Total EPH	93.4			2.09	5.99	mg/kg			

* As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C40 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C40 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	PSEG	Date Collected:	
Project:	Orange Gas and Appliance Service (EO)	Date Received:	
Client Sample ID:	PB170632BS	SDG No.:	Q3669
Lab Sample ID:	PB170632BS	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	100
Sample Wt/Vol:	30.03 g	Test:	EPH_NF
Prep Method :		Final Vol:	2000 uL
		Prep Date :	11/19/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Datafile	Date Ana.	Prep BatchID
------------	-----------	-------	------	----	-----	------------	-------	----------	-----------	--------------

TARGETS

Total AliphaticEPH	Total AliphaticEPH	93.4			2.09	5.99	mg/kg			
Total EPH	Total EPH	93.4			2.09	5.99	mg/kg			

* As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C40 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C40 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	PSEG	Date Collected:	
Project:	Orange Gas and Appliance Service (EO)	Date Received:	
Client Sample ID:	PB170632BS	SDG No.:	Q3669
Lab Sample ID:	PB170632BS	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	100
Sample Wt/Vol:	30.03 g	Test:	EPH_NF
Prep Method :			
	Final Vol: 2000 uL		
	Prep Date : 11/19/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Datafile	Date Ana.	Prep BatchID
------------	-----------	-------	------	----	-----	------------	-------	----------	-----------	--------------

TARGETS

Total AliphaticEPH	Total AliphaticEPH	93.4			2.09	5.99	mg/kg			
Total EPH	Total EPH	93.4			2.09	5.99	mg/kg			

* As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C40 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C40 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	PSEG	Date Collected:	
Project:	Orange Gas and Appliance Service (EO)	Date Received:	
Client Sample ID:	PB170632BS	SDG No.:	Q3669
Lab Sample ID:	PB170632BS	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	100
Sample Wt/Vol:	30.03 g	Test:	EPH_NF
Prep Method :		Final Vol:	2000 uL
		Prep Date :	11/19/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Datafile	Date Ana.	Prep BatchID
------------	-----------	-------	------	----	-----	------------	-------	----------	-----------	--------------

TARGETS

Total AliphaticEPH	Total AliphaticEPH	93.4			2.09	5.99	mg/kg			
Total EPH	Total EPH	93.4			2.09	5.99	mg/kg			

* As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C40 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C40 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	PSEG	Date Collected:	
Project:	Orange Gas and Appliance Service (EO)	Date Received:	
Client Sample ID:	PB170632BS	SDG No.:	Q3669
Lab Sample ID:	PB170632BS	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	100
Sample Wt/Vol:	30.03 g	Test:	EPH_NF
Prep Method :			
	Final Vol: 2000 uL		
	Prep Date : 11/19/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Datafile	Date Ana.	Prep BatchID
------------	-----------	-------	------	----	-----	------------	-------	----------	-----------	--------------

TARGETS

Total AliphaticEPH	Total AliphaticEPH	93.4			2.09	5.99	mg/kg			
Total EPH	Total EPH	93.4			2.09	5.99	mg/kg			

* As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C40 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C40 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	PSEG	Date Collected:	
Project:	Orange Gas and Appliance Service (EO)	Date Received:	
Client Sample ID:	PB170632BS	SDG No.:	Q3669
Lab Sample ID:	PB170632BS	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	100
Sample Wt/Vol:	30.03 g	Test:	EPH_NF
Prep Method :		Final Vol:	2000 uL
		Prep Date :	11/19/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Datafile	Date Ana.	Prep BatchID
------------	-----------	-------	------	----	-----	------------	-------	----------	-----------	--------------

TARGETS

Total AliphaticEPH	Total AliphaticEPH	93.4			2.09	5.99	mg/kg			
Total EPH	Total EPH	93.4			2.09	5.99	mg/kg			

* As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C40 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C40 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	PSEG	Date Collected:	
Project:	Orange Gas and Appliance Service (EO)	Date Received:	
Client Sample ID:	PB170632BS	SDG No.:	Q3669
Lab Sample ID:	PB170632BS	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	100
Sample Wt/Vol:	30.03 g	Test:	EPH_NF
Prep Method :			
	Final Vol: 2000 uL		
	Prep Date : 11/19/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Datafile	Date Ana.	Prep BatchID
TARGETS										
Aliphatic C28-C40	Aliphatic C28-C40	24.5		1	1.18	2.00	mg/kg	FE056935.D	11/19/25 14:39	PB170632
Aliphatic C9-C28	Aliphatic C9-C28	68.9		1	0.91	3.99	mg/kg	FE056935.D	11/19/25 14:39	PB170632
Total AliphaticEPH	Total AliphaticEPH	93.4			2.09	5.99	mg/kg			
Total EPH	Total EPH	93.4			2.09	5.99	mg/kg			

* As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C40 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C40 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	PSEG	Date Collected:	
Project:	Orange Gas and Appliance Service (EO)	Date Received:	
Client Sample ID:	PB170632BS	SDG No.:	Q3669
Lab Sample ID:	PB170632BS	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	100
Sample Wt/Vol:	30.03 g	Final Vol:	2000 uL
Prep Method :		Prep Date	11/19/25
		Test:	EPH_NF

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS									
Aliphatic C9-C28	Aliphatic C9-C28	68.9	1		0.91	3.99	mg/kg	11/19/25	PB170632
Aliphatic C28-C40	Aliphatic C28-C40	24.5	1		1.18	2.00	mg/kg	11/19/25	PB170632
SURROGATES									
3383-33-2	1-chlorooctadecane (SURR)	30.1			40 - 140	60%	SPK: 50		
84-15-1	ortho-Terphenyl (SURR)	34.3			40 - 140	69%	SPK: 50		

Quantitation Report For Aliphatic EPH Range.

Lab Sample ID:	PB170632BS	Acq On:	19 Nov 2025 14:39
Client Sample ID:	PB170632BS	Operator:	YP\AJ
Data file:	FE056935.D	Misc:	
Instrument:	FID_E	ALS Vial:	18
Dilution Factor:	1	Sample Multiplier:	1.00

Compound	R.T.		Response	Conc	highest_standard	Units
Aliphatic C9-C12	3.209	6.832	34744240	202.99	300	ug/ml
Aliphatic C12-C16	6.833	10.276	44613076	249.578	200	ug/ml
Aliphatic C16-C21	10.277	13.648	49316119	264.422	300	ug/ml
Aliphatic C21-C28	13.649	17.318	61401352	318.528	400	ug/ml
Aliphatic C28-C40	17.319	22.247	58526334	367.363	600	ug/ml
Aliphatic EPH	3.209	22.247	248601121	1400		ug/ml
ortho-Terphenyl (SURR)	11.946	11.946	7171296	34.33		ug/ml
1-chlorooctadecane (SURR)	13.383	13.383	4841612	30.05		ug/ml
Aliphatic C9-C28	3.209	17.318	190074787	1040	1200	ug/ml