



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q3671

Contract: CORE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB123	Mercury	0.076	+/-0.2	U	0.20	CV	11/20/2025	12:13	LB137986

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q3671

Contract: CORE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB19	Mercury	0.076	+/-0.2	U	0.20	CV	11/20/2025	12:18	LB137986
CCB20	Mercury	0.076	+/-0.2	U	0.20	CV	11/20/2025	12:51	LB137986
CCB21	Mercury	0.076	+/-0.2	U	0.20	CV	11/20/2025	13:08	LB137986

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q3671

Contract: CORE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	11/20/2025	13:21	LB137997
	Antimony	6.76	+/-25	U	50.0	P	11/20/2025	13:21	LB137997
	Arsenic	5.12	+/-10	U	20.0	P	11/20/2025	13:21	LB137997
	Barium	14.6	+/-50	U	100	P	11/20/2025	13:21	LB137997
	Beryllium	0.56	+/-3	U	6.00	P	11/20/2025	13:21	LB137997
	Cadmium	0.50	+/-3	U	6.00	P	11/20/2025	13:21	LB137997
	Calcium	234	+/-1000	U	2000	P	11/20/2025	13:21	LB137997
	Chromium	2.12	+/-5	U	10.0	P	11/20/2025	13:21	LB137997
	Cobalt	2.26	+/-15	U	30.0	P	11/20/2025	13:21	LB137997
	Copper	4.60	+/-10	U	20.0	P	11/20/2025	13:21	LB137997
	Iron	23.4	+/-50	U	100	P	11/20/2025	13:21	LB137997
	Lead	2.30	+/-6	U	12.0	P	11/20/2025	13:21	LB137997
	Magnesium	244	+/-1000	U	2000	P	11/20/2025	13:21	LB137997
	Manganese	5.94	+/-10	U	20.0	P	11/20/2025	13:21	LB137997
	Nickel	3.06	+/-20	U	40.0	P	11/20/2025	13:21	LB137997
	Potassium	918	+/-1000	U	2000	P	11/20/2025	13:21	LB137997
	Selenium	9.64	+/-10	U	20.0	P	11/20/2025	13:21	LB137997
	Silver	1.62	+/-5	U	10.0	P	11/20/2025	13:21	LB137997
	Sodium	868	+/-1000	U	2000	P	11/20/2025	13:21	LB137997
	Thallium	4.38	+/-20	U	40.0	P	11/20/2025	13:21	LB137997
	Vanadium	6.26	+/-20	U	40.0	P	11/20/2025	13:21	LB137997
	Zinc	16.7	+/-20	U	40.0	P	11/20/2025	13:21	LB137997

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q3671

Contract: CORE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	11/20/2025	14:01	LB137997
	Antimony	6.76	+/-25	U	50.0	P	11/20/2025	14:01	LB137997
	Arsenic	5.12	+/-10	U	20.0	P	11/20/2025	14:01	LB137997
	Barium	14.6	+/-50	U	100	P	11/20/2025	14:01	LB137997
	Beryllium	0.56	+/-3	U	6.00	P	11/20/2025	14:01	LB137997
	Cadmium	0.50	+/-3	U	6.00	P	11/20/2025	14:01	LB137997
	Calcium	234	+/-1000	U	2000	P	11/20/2025	14:01	LB137997
	Chromium	2.12	+/-5	U	10.0	P	11/20/2025	14:01	LB137997
	Cobalt	2.26	+/-15	U	30.0	P	11/20/2025	14:01	LB137997
	Copper	4.60	+/-10	U	20.0	P	11/20/2025	14:01	LB137997
	Iron	23.4	+/-50	U	100	P	11/20/2025	14:01	LB137997
	Lead	2.30	+/-6	U	12.0	P	11/20/2025	14:01	LB137997
	Magnesium	244	+/-1000	U	2000	P	11/20/2025	14:01	LB137997
	Manganese	5.94	+/-10	U	20.0	P	11/20/2025	14:01	LB137997
	Nickel	3.06	+/-20	U	40.0	P	11/20/2025	14:01	LB137997
	Potassium	918	+/-1000	U	2000	P	11/20/2025	14:01	LB137997
	Selenium	9.64	+/-10	U	20.0	P	11/20/2025	14:01	LB137997
	Silver	1.62	+/-5	U	10.0	P	11/20/2025	14:01	LB137997
	Sodium	868	+/-1000	U	2000	P	11/20/2025	14:01	LB137997
	Thallium	4.38	+/-20	U	40.0	P	11/20/2025	14:01	LB137997
	Vanadium	6.26	+/-20	U	40.0	P	11/20/2025	14:01	LB137997
	Zinc	16.7	+/-20	U	40.0	P	11/20/2025	14:01	LB137997
CCB02	Aluminum	11.3	+/-50	U	100	P	11/20/2025	15:46	LB137997
	Antimony	6.76	+/-25	U	50.0	P	11/20/2025	15:46	LB137997
	Arsenic	5.12	+/-10	U	20.0	P	11/20/2025	15:46	LB137997
	Barium	14.6	+/-50	U	100	P	11/20/2025	15:46	LB137997
	Beryllium	0.56	+/-3	U	6.00	P	11/20/2025	15:46	LB137997
	Cadmium	0.50	+/-3	U	6.00	P	11/20/2025	15:46	LB137997
	Calcium	234	+/-1000	U	2000	P	11/20/2025	15:46	LB137997
	Chromium	2.12	+/-5	U	10.0	P	11/20/2025	15:46	LB137997
	Cobalt	2.26	+/-15	U	30.0	P	11/20/2025	15:46	LB137997
	Copper	4.60	+/-10	U	20.0	P	11/20/2025	15:46	LB137997
	Iron	23.4	+/-50	U	100	P	11/20/2025	15:46	LB137997
	Lead	2.30	+/-6	U	12.0	P	11/20/2025	15:46	LB137997
	Magnesium	244	+/-1000	U	2000	P	11/20/2025	15:46	LB137997
	Manganese	5.94	+/-10	U	20.0	P	11/20/2025	15:46	LB137997
	Nickel	3.06	+/-20	U	40.0	P	11/20/2025	15:46	LB137997
	Potassium	918	+/-1000	U	2000	P	11/20/2025	15:46	LB137997
	Selenium	9.64	+/-10	U	20.0	P	11/20/2025	15:46	LB137997

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q3671

Contract: CORE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	11/20/2025	15:46	LB137997
	Sodium	868	+/-1000	U	2000	P	11/20/2025	15:46	LB137997
	Thallium	4.38	+/-20	U	40.0	P	11/20/2025	15:46	LB137997
	Vanadium	6.26	+/-20	U	40.0	P	11/20/2025	15:46	LB137997
	Zinc	16.7	+/-20	U	40.0	P	11/20/2025	15:46	LB137997

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q3671

Contract: CORE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	11/21/2025	13:13	LB138017
	Antimony	6.76	+/-25	U	50.0	P	11/21/2025	13:13	LB138017
	Arsenic	5.12	+/-10	U	20.0	P	11/21/2025	13:13	LB138017
	Barium	14.6	+/-50	U	100	P	11/21/2025	13:13	LB138017
	Beryllium	0.56	+/-3	U	6.00	P	11/21/2025	13:13	LB138017
	Cadmium	0.50	+/-3	U	6.00	P	11/21/2025	13:13	LB138017
	Calcium	234	+/-1000	U	2000	P	11/21/2025	13:13	LB138017
	Chromium	2.12	+/-5	U	10.0	P	11/21/2025	13:13	LB138017
	Cobalt	2.26	+/-15	U	30.0	P	11/21/2025	13:13	LB138017
	Copper	4.60	+/-10	U	20.0	P	11/21/2025	13:13	LB138017
	Iron	23.4	+/-50	U	100	P	11/21/2025	13:13	LB138017
	Lead	2.30	+/-6	U	12.0	P	11/21/2025	13:13	LB138017
	Magnesium	244	+/-1000	U	2000	P	11/21/2025	13:13	LB138017
	Manganese	5.94	+/-10	U	20.0	P	11/21/2025	13:13	LB138017
	Nickel	3.06	+/-20	U	40.0	P	11/21/2025	13:13	LB138017
	Potassium	918	+/-1000	U	2000	P	11/21/2025	13:13	LB138017
	Selenium	9.64	+/-10	U	20.0	P	11/21/2025	13:13	LB138017
	Silver	1.62	+/-5	U	10.0	P	11/21/2025	13:13	LB138017
	Sodium	868	+/-1000	U	2000	P	11/21/2025	13:13	LB138017
	Thallium	4.38	+/-20	U	40.0	P	11/21/2025	13:13	LB138017
	Vanadium	6.26	+/-20	U	40.0	P	11/21/2025	13:13	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	13:13	LB138017

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q3671

Contract: CORE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	11/21/2025	13:48	LB138017
	Antimony	6.76	+/-25	U	50.0	P	11/21/2025	13:48	LB138017
	Arsenic	5.12	+/-10	U	20.0	P	11/21/2025	13:48	LB138017
	Barium	14.6	+/-50	U	100	P	11/21/2025	13:48	LB138017
	Beryllium	0.56	+/-3	U	6.00	P	11/21/2025	13:48	LB138017
	Cadmium	0.50	+/-3	U	6.00	P	11/21/2025	13:48	LB138017
	Calcium	234	+/-1000	U	2000	P	11/21/2025	13:48	LB138017
	Chromium	2.12	+/-5	U	10.0	P	11/21/2025	13:48	LB138017
	Cobalt	2.26	+/-15	U	30.0	P	11/21/2025	13:48	LB138017
	Copper	4.60	+/-10	U	20.0	P	11/21/2025	13:48	LB138017
	Iron	23.4	+/-50	U	100	P	11/21/2025	13:48	LB138017
	Lead	2.30	+/-6	U	12.0	P	11/21/2025	13:48	LB138017
	Magnesium	244	+/-1000	U	2000	P	11/21/2025	13:48	LB138017
	Manganese	5.94	+/-10	U	20.0	P	11/21/2025	13:48	LB138017
	Nickel	3.06	+/-20	U	40.0	P	11/21/2025	13:48	LB138017
	Potassium	918	+/-1000	U	2000	P	11/21/2025	13:48	LB138017
	Selenium	9.64	+/-10	U	20.0	P	11/21/2025	13:48	LB138017
	Silver	1.62	+/-5	U	10.0	P	11/21/2025	13:48	LB138017
	Sodium	868	+/-1000	U	2000	P	11/21/2025	13:48	LB138017
	Thallium	4.38	+/-20	U	40.0	P	11/21/2025	13:48	LB138017
	Vanadium	6.26	+/-20	U	40.0	P	11/21/2025	13:48	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	13:48	LB138017
CCB02	Aluminum	11.5	+/-50	J	100	P	11/21/2025	14:52	LB138017
	Antimony	6.76	+/-25	U	50.0	P	11/21/2025	14:52	LB138017
	Arsenic	5.12	+/-10	U	20.0	P	11/21/2025	14:52	LB138017
	Barium	14.6	+/-50	U	100	P	11/21/2025	14:52	LB138017
	Beryllium	0.56	+/-3	U	6.00	P	11/21/2025	14:52	LB138017
	Cadmium	0.50	+/-3	U	6.00	P	11/21/2025	14:52	LB138017
	Calcium	234	+/-1000	U	2000	P	11/21/2025	14:52	LB138017
	Chromium	2.12	+/-5	U	10.0	P	11/21/2025	14:52	LB138017
	Cobalt	2.26	+/-15	U	30.0	P	11/21/2025	14:52	LB138017
	Copper	4.60	+/-10	U	20.0	P	11/21/2025	14:52	LB138017
	Iron	23.4	+/-50	U	100	P	11/21/2025	14:52	LB138017
	Lead	2.30	+/-6	U	12.0	P	11/21/2025	14:52	LB138017
	Magnesium	244	+/-1000	U	2000	P	11/21/2025	14:52	LB138017
	Manganese	5.94	+/-10	U	20.0	P	11/21/2025	14:52	LB138017
	Nickel	3.06	+/-20	U	40.0	P	11/21/2025	14:52	LB138017
	Potassium	918	+/-1000	U	2000	P	11/21/2025	14:52	LB138017
	Selenium	9.64	+/-10	U	20.0	P	11/21/2025	14:52	LB138017

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q3671

Contract: CORE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	11/21/2025	14:52	LB138017
	Sodium	868	+/-1000	U	2000	P	11/21/2025	14:52	LB138017
	Thallium	4.38	+/-20	U	40.0	P	11/21/2025	14:52	LB138017
	Vanadium	6.26	+/-20	U	40.0	P	11/21/2025	14:52	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	14:52	LB138017
CCB03	Aluminum	11.3	+/-50	U	100	P	11/21/2025	15:42	LB138017
	Antimony	6.76	+/-25	U	50.0	P	11/21/2025	15:42	LB138017
	Arsenic	5.12	+/-10	U	20.0	P	11/21/2025	15:42	LB138017
	Barium	14.6	+/-50	U	100	P	11/21/2025	15:42	LB138017
	Beryllium	0.56	+/-3	U	6.00	P	11/21/2025	15:42	LB138017
	Cadmium	0.50	+/-3	U	6.00	P	11/21/2025	15:42	LB138017
	Calcium	234	+/-1000	U	2000	P	11/21/2025	15:42	LB138017
	Chromium	2.12	+/-5	U	10.0	P	11/21/2025	15:42	LB138017
	Cobalt	2.26	+/-15	U	30.0	P	11/21/2025	15:42	LB138017
	Copper	4.60	+/-10	U	20.0	P	11/21/2025	15:42	LB138017
	Iron	23.4	+/-50	U	100	P	11/21/2025	15:42	LB138017
	Lead	2.30	+/-6	U	12.0	P	11/21/2025	15:42	LB138017
	Magnesium	244	+/-1000	U	2000	P	11/21/2025	15:42	LB138017
	Manganese	5.94	+/-10	U	20.0	P	11/21/2025	15:42	LB138017
	Nickel	3.06	+/-20	U	40.0	P	11/21/2025	15:42	LB138017
	Potassium	918	+/-1000	U	2000	P	11/21/2025	15:42	LB138017
	Selenium	9.64	+/-10	U	20.0	P	11/21/2025	15:42	LB138017
	Silver	1.62	+/-5	U	10.0	P	11/21/2025	15:42	LB138017
	Sodium	868	+/-1000	U	2000	P	11/21/2025	15:42	LB138017
	Thallium	4.38	+/-20	U	40.0	P	11/21/2025	15:42	LB138017
	Vanadium	6.26	+/-20	U	40.0	P	11/21/2025	15:42	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	15:42	LB138017
CCB04	Aluminum	11.3	+/-50	U	100	P	11/21/2025	16:32	LB138017
	Antimony	6.76	+/-25	U	50.0	P	11/21/2025	16:32	LB138017
	Arsenic	5.12	+/-10	U	20.0	P	11/21/2025	16:32	LB138017
	Barium	14.6	+/-50	U	100	P	11/21/2025	16:32	LB138017
	Beryllium	0.56	+/-3	U	6.00	P	11/21/2025	16:32	LB138017
	Cadmium	0.50	+/-3	U	6.00	P	11/21/2025	16:32	LB138017
	Calcium	234	+/-1000	U	2000	P	11/21/2025	16:32	LB138017
	Chromium	2.12	+/-5	U	10.0	P	11/21/2025	16:32	LB138017
	Cobalt	2.26	+/-15	U	30.0	P	11/21/2025	16:32	LB138017
	Copper	4.60	+/-10	U	20.0	P	11/21/2025	16:32	LB138017
	Iron	23.4	+/-50	U	100	P	11/21/2025	16:32	LB138017
	Lead	2.30	+/-6	U	12.0	P	11/21/2025	16:32	LB138017

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q3671

Contract: CORE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	11/21/2025	16:32	LB138017
	Manganese	5.94	+/-10	U	20.0	P	11/21/2025	16:32	LB138017
	Nickel	3.06	+/-20	U	40.0	P	11/21/2025	16:32	LB138017
	Potassium	918	+/-1000	U	2000	P	11/21/2025	16:32	LB138017
	Selenium	9.64	+/-10	U	20.0	P	11/21/2025	16:32	LB138017
	Silver	1.62	+/-5	U	10.0	P	11/21/2025	16:32	LB138017
	Sodium	868	+/-1000	U	2000	P	11/21/2025	16:32	LB138017
	Thallium	4.38	+/-20	U	40.0	P	11/21/2025	16:32	LB138017
	Vanadium	6.26	+/-20	U	40.0	P	11/21/2025	16:32	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	16:32	LB138017
CCB05	Aluminum	11.3	+/-50	U	100	P	11/21/2025	17:55	LB138017
	Antimony	6.76	+/-25	U	50.0	P	11/21/2025	17:55	LB138017
	Arsenic	5.12	+/-10	U	20.0	P	11/21/2025	17:55	LB138017
	Barium	14.6	+/-50	U	100	P	11/21/2025	17:55	LB138017
	Beryllium	0.56	+/-3	U	6.00	P	11/21/2025	17:55	LB138017
	Cadmium	0.50	+/-3	U	6.00	P	11/21/2025	17:55	LB138017
	Calcium	234	+/-1000	U	2000	P	11/21/2025	17:55	LB138017
	Chromium	2.12	+/-5	U	10.0	P	11/21/2025	17:55	LB138017
	Cobalt	2.26	+/-15	U	30.0	P	11/21/2025	17:55	LB138017
	Copper	4.60	+/-10	U	20.0	P	11/21/2025	17:55	LB138017
	Iron	23.4	+/-50	U	100	P	11/21/2025	17:55	LB138017
	Lead	2.30	+/-6	U	12.0	P	11/21/2025	17:55	LB138017
	Magnesium	244	+/-1000	U	2000	P	11/21/2025	17:55	LB138017
	Manganese	5.94	+/-10	U	20.0	P	11/21/2025	17:55	LB138017
	Nickel	3.06	+/-20	U	40.0	P	11/21/2025	17:55	LB138017
	Potassium	918	+/-1000	U	2000	P	11/21/2025	17:55	LB138017
	Selenium	9.64	+/-10	U	20.0	P	11/21/2025	17:55	LB138017
	Silver	1.62	+/-5	U	10.0	P	11/21/2025	17:55	LB138017
	Sodium	868	+/-1000	U	2000	P	11/21/2025	17:55	LB138017
	Thallium	4.38	+/-20	U	40.0	P	11/21/2025	17:55	LB138017
	Vanadium	6.26	+/-20	U	40.0	P	11/21/2025	17:55	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	17:55	LB138017
CCB06	Aluminum	11.3	+/-50	U	100	P	11/21/2025	18:45	LB138017
	Antimony	6.76	+/-25	U	50.0	P	11/21/2025	18:45	LB138017
	Arsenic	5.12	+/-10	U	20.0	P	11/21/2025	18:45	LB138017
	Barium	14.6	+/-50	U	100	P	11/21/2025	18:45	LB138017
	Beryllium	0.56	+/-3	U	6.00	P	11/21/2025	18:45	LB138017
	Cadmium	0.50	+/-3	U	6.00	P	11/21/2025	18:45	LB138017
	Calcium	234	+/-1000	U	2000	P	11/21/2025	18:45	LB138017

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q3671

Contract: CORE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	11/21/2025	18:45	LB138017
	Cobalt	2.26	+/-15	U	30.0	P	11/21/2025	18:45	LB138017
	Copper	4.60	+/-10	U	20.0	P	11/21/2025	18:45	LB138017
	Iron	23.4	+/-50	U	100	P	11/21/2025	18:45	LB138017
	Lead	2.30	+/-6	U	12.0	P	11/21/2025	18:45	LB138017
	Magnesium	244	+/-1000	U	2000	P	11/21/2025	18:45	LB138017
	Manganese	5.94	+/-10	U	20.0	P	11/21/2025	18:45	LB138017
	Nickel	3.06	+/-20	U	40.0	P	11/21/2025	18:45	LB138017
	Potassium	918	+/-1000	U	2000	P	11/21/2025	18:45	LB138017
	Selenium	9.64	+/-10	U	20.0	P	11/21/2025	18:45	LB138017
	Silver	1.62	+/-5	U	10.0	P	11/21/2025	18:45	LB138017
	Sodium	868	+/-1000	U	2000	P	11/21/2025	18:45	LB138017
	Thallium	4.38	+/-20	U	40.0	P	11/21/2025	18:45	LB138017
	Vanadium	6.26	+/-20	U	40.0	P	11/21/2025	18:45	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	18:45	LB138017
CCB07	Aluminum	16.8	+/-50	J	100	P	11/21/2025	19:35	LB138017
	Antimony	6.76	+/-25	U	50.0	P	11/21/2025	19:35	LB138017
	Arsenic	5.12	+/-10	U	20.0	P	11/21/2025	19:35	LB138017
	Barium	14.6	+/-50	U	100	P	11/21/2025	19:35	LB138017
	Beryllium	0.56	+/-3	U	6.00	P	11/21/2025	19:35	LB138017
	Cadmium	0.50	+/-3	U	6.00	P	11/21/2025	19:35	LB138017
	Calcium	234	+/-1000	U	2000	P	11/21/2025	19:35	LB138017
	Chromium	2.12	+/-5	U	10.0	P	11/21/2025	19:35	LB138017
	Cobalt	2.26	+/-15	U	30.0	P	11/21/2025	19:35	LB138017
	Copper	4.60	+/-10	U	20.0	P	11/21/2025	19:35	LB138017
	Iron	23.4	+/-50	U	100	P	11/21/2025	19:35	LB138017
	Lead	2.30	+/-6	U	12.0	P	11/21/2025	19:35	LB138017
	Magnesium	244	+/-1000	U	2000	P	11/21/2025	19:35	LB138017
	Manganese	5.94	+/-10	U	20.0	P	11/21/2025	19:35	LB138017
	Nickel	3.06	+/-20	U	40.0	P	11/21/2025	19:35	LB138017
	Potassium	918	+/-1000	U	2000	P	11/21/2025	19:35	LB138017
	Selenium	9.64	+/-10	U	20.0	P	11/21/2025	19:35	LB138017
	Silver	1.62	+/-5	U	10.0	P	11/21/2025	19:35	LB138017
	Sodium	868	+/-1000	U	2000	P	11/21/2025	19:35	LB138017
	Thallium	4.38	+/-20	U	40.0	P	11/21/2025	19:35	LB138017
	Vanadium	6.26	+/-20	U	40.0	P	11/21/2025	19:35	LB138017
	Zinc	16.7	+/-20	U	40.0	P	11/21/2025	19:35	LB138017