

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3672

Continuing Calib Date: 11/24/2025

Initial Calibration Date(s): 11/18/2025

11/18/2025

Continuing Calib Time: 18:25

Initial Calibration Time(s): 12:13

17:52

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.50       | 4.50      | 4.40      | 4.60 | 0.00       |
| Aroclor-1016-2       | (2) | 4.52       | 4.52      | 4.42      | 4.62 | 0.00       |
| Aroclor-1016-3       | (3) | 4.57       | 4.57      | 4.47      | 4.67 | 0.00       |
| Aroclor-1016-4       | (4) | 4.66       | 4.66      | 4.56      | 4.76 | 0.00       |
| Aroclor-1016-5       | (5) | 4.95       | 4.95      | 4.85      | 5.05 | 0.00       |
| Aroclor-1260-1       | (1) | 6.04       | 6.04      | 5.94      | 6.14 | 0.00       |
| Aroclor-1260-2       | (2) | 6.30       | 6.30      | 6.20      | 6.40 | 0.00       |
| Aroclor-1260-3       | (3) | 6.65       | 6.65      | 6.55      | 6.75 | 0.00       |
| Aroclor-1260-4       | (4) | 6.87       | 6.87      | 6.77      | 6.97 | 0.00       |
| Aroclor-1260-5       | (5) | 7.17       | 7.18      | 7.08      | 7.28 | 0.01       |
| Tetrachloro-m-xylene |     | 3.41       | 3.41      | 3.31      | 3.51 | 0.00       |
| Decachlorobiphenyl   |     | 8.54       | 8.54      | 8.44      | 8.64 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3672

Continuing Calib Date: 11/24/2025

Initial Calibration Date(s): 11/18/2025

11/18/2025

Continuing Calib Time: 18:25

Initial Calibration Time(s): 12:13

17:52

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 3.87       | 3.87      | 3.77      | 3.97 | 0.00       |
| Aroclor-1016-2       | (2) | 3.88       | 3.89      | 3.79      | 3.99 | 0.01       |
| Aroclor-1016-3       | (3) | 4.05       | 4.05      | 3.95      | 4.15 | 0.00       |
| Aroclor-1016-4       | (4) | 4.09       | 4.10      | 4.00      | 4.20 | 0.01       |
| Aroclor-1016-5       | (5) | 4.29       | 4.30      | 4.20      | 4.40 | 0.01       |
| Aroclor-1260-1       | (1) | 5.29       | 5.30      | 5.20      | 5.40 | 0.01       |
| Aroclor-1260-2       | (2) | 5.48       | 5.49      | 5.39      | 5.59 | 0.01       |
| Aroclor-1260-3       | (3) | 5.62       | 5.63      | 5.53      | 5.73 | 0.01       |
| Aroclor-1260-4       | (4) | 6.09       | 6.09      | 5.99      | 6.19 | 0.00       |
| Aroclor-1260-5       | (5) | 6.33       | 6.34      | 6.24      | 6.44 | 0.01       |
| Tetrachloro-m-xylene |     | 2.85       | 2.86      | 2.76      | 2.96 | 0.01       |
| Decachlorobiphenyl   |     | 7.65       | 7.66      | 7.56      | 7.76 | 0.01       |