

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3672
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071398.D **Time Analyzed:** 23:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.496	4.397	4.597	509.320	500.000	1.9
Aroclor-1016-2	4.515	4.416	4.616	516.580	500.000	3.3
Aroclor-1016-3	4.573	4.474	4.674	517.880	500.000	3.6
Aroclor-1016-4	4.663	4.564	4.764	502.710	500.000	0.5
Aroclor-1016-5	4.948	4.849	5.049	511.080	500.000	2.2
Aroclor-1260-1	6.042	5.943	6.143	498.240	500.000	-0.4
Aroclor-1260-2	6.299	6.200	6.400	518.460	500.000	3.7
Aroclor-1260-3	6.651	6.552	6.752	501.630	500.000	0.3
Aroclor-1260-4	6.866	6.767	6.967	502.840	500.000	0.6
Aroclor-1260-5	7.173	7.075	7.275	509.510	500.000	1.9
Decachlorobiphenyl	8.535	8.437	8.637	44.750	50.000	-10.5
Tetrachloro-m-xylene	3.410	3.311	3.511	49.380	50.000	-1.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3672
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071398.D **Time Analyzed:** 23:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.869	3.771	3.971	588.090	500.000	17.6
Aroclor-1016-2	3.885	3.788	3.988	577.680	500.000	15.5
Aroclor-1016-3	4.046	3.950	4.150	596.890	500.000	19.4
Aroclor-1016-4	4.095	3.998	4.198	582.220	500.000	16.4
Aroclor-1016-5	4.292	4.195	4.395	587.310	500.000	17.5
Aroclor-1260-1	5.293	5.197	5.397	578.610	500.000	15.7
Aroclor-1260-2	5.482	5.386	5.586	578.950	500.000	15.8
Aroclor-1260-3	5.626	5.530	5.730	582.200	500.000	16.4
Aroclor-1260-4	6.089	5.993	6.193	581.900	500.000	16.4
Aroclor-1260-5	6.334	6.238	6.438	585.640	500.000	17.1
Decachlorobiphenyl	7.654	7.559	7.759	54.810	50.000	9.6
Tetrachloro-m-xylene	2.855	2.756	2.956	56.450	50.000	12.9