

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3672
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071410.D **Time Analyzed:** 04:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.497	4.397	4.597	445.840	500.000	-10.8
Aroclor-1016-2	4.516	4.416	4.616	434.950	500.000	-13.0
Aroclor-1016-3	4.573	4.474	4.674	430.870	500.000	-13.8
Aroclor-1016-4	4.663	4.564	4.764	437.750	500.000	-12.5
Aroclor-1016-5	4.949	4.849	5.049	443.520	500.000	-11.3
Aroclor-1260-1	6.042	5.943	6.143	408.020	500.000	-18.4
Aroclor-1260-2	6.299	6.200	6.400	429.850	500.000	-14.0
Aroclor-1260-3	6.652	6.552	6.752	426.380	500.000	-14.7
Aroclor-1260-4	6.866	6.767	6.967	431.540	500.000	-13.7
Aroclor-1260-5	7.174	7.075	7.275	427.990	500.000	-14.4
Decachlorobiphenyl	8.535	8.437	8.637	41.280	50.000	-17.4
Tetrachloro-m-xylene	3.411	3.311	3.511	44.600	50.000	-10.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3672
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071410.D **Time Analyzed:** 04:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.868	3.771	3.971	513.860	500.000	2.8
Aroclor-1016-2	3.885	3.788	3.988	520.150	500.000	4.0
Aroclor-1016-3	4.047	3.950	4.150	511.890	500.000	2.4
Aroclor-1016-4	4.095	3.998	4.198	504.220	500.000	0.8
Aroclor-1016-5	4.293	4.195	4.395	517.370	500.000	3.5
Aroclor-1260-1	5.293	5.197	5.397	500.850	500.000	0.2
Aroclor-1260-2	5.482	5.386	5.586	505.310	500.000	1.1
Aroclor-1260-3	5.626	5.530	5.730	503.460	500.000	0.7
Aroclor-1260-4	6.089	5.993	6.193	504.810	500.000	1.0
Aroclor-1260-5	6.334	6.238	6.438	499.360	500.000	-0.1
Decachlorobiphenyl	7.654	7.559	7.759	50.260	50.000	0.5
Tetrachloro-m-xylene	2.855	2.756	2.956	51.890	50.000	3.8