

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3672
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071415.D **Time Analyzed:** 06:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.496	4.397	4.597	432.840	500.000	-13.4
Aroclor-1016-2	4.515	4.416	4.616	448.570	500.000	-10.3
Aroclor-1016-3	4.573	4.474	4.674	442.930	500.000	-11.4
Aroclor-1016-4	4.663	4.564	4.764	434.310	500.000	-13.1
Aroclor-1016-5	4.949	4.849	5.049	439.120	500.000	-12.2
Aroclor-1260-1	6.042	5.943	6.143	407.760	500.000	-18.4
Aroclor-1260-2	6.300	6.200	6.400	430.950	500.000	-13.8
Aroclor-1260-3	6.652	6.552	6.752	427.000	500.000	-14.6
Aroclor-1260-4	6.867	6.767	6.967	428.230	500.000	-14.4
Aroclor-1260-5	7.174	7.075	7.275	429.580	500.000	-14.1
Decachlorobiphenyl	8.536	8.437	8.637	42.050	50.000	-15.9
Tetrachloro-m-xylene	3.411	3.311	3.511	44.300	50.000	-11.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3672
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071415.D **Time Analyzed:** 06:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.868	3.771	3.971	506.580	500.000	1.3
Aroclor-1016-2	3.884	3.788	3.988	519.720	500.000	3.9
Aroclor-1016-3	4.046	3.950	4.150	513.070	500.000	2.6
Aroclor-1016-4	4.095	3.998	4.198	504.840	500.000	1.0
Aroclor-1016-5	4.292	4.195	4.395	513.960	500.000	2.8
Aroclor-1260-1	5.293	5.197	5.397	495.390	500.000	-0.9
Aroclor-1260-2	5.482	5.386	5.586	503.330	500.000	0.7
Aroclor-1260-3	5.625	5.530	5.730	499.750	500.000	-0.1
Aroclor-1260-4	6.089	5.993	6.193	502.050	500.000	0.4
Aroclor-1260-5	6.334	6.238	6.438	498.830	500.000	-0.2
Decachlorobiphenyl	7.654	7.559	7.759	50.170	50.000	0.3
Tetrachloro-m-xylene	2.855	2.756	2.956	51.370	50.000	2.7