

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3672
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 11/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071418.D **Time Analyzed:** 09:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.511	4.397	4.597	468.710	500.000	-6.3
Aroclor-1016-2	4.530	4.416	4.616	456.980	500.000	-8.6
Aroclor-1016-3	4.588	4.474	4.674	480.010	500.000	-4.0
Aroclor-1016-4	4.678	4.564	4.764	465.540	500.000	-6.9
Aroclor-1016-5	4.963	4.849	5.049	473.390	500.000	-5.3
Aroclor-1260-1	6.057	5.943	6.143	463.600	500.000	-7.3
Aroclor-1260-2	6.315	6.200	6.400	479.950	500.000	-4.0
Aroclor-1260-3	6.667	6.552	6.752	470.240	500.000	-6.0
Aroclor-1260-4	6.882	6.767	6.967	482.260	500.000	-3.5
Aroclor-1260-5	7.189	7.075	7.275	484.030	500.000	-3.2
Decachlorobiphenyl	8.554	8.437	8.637	44.460	50.000	-11.1
Tetrachloro-m-xylene	3.424	3.311	3.511	46.350	50.000	-7.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3672
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 11/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071418.D **Time Analyzed:** 09:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.868	3.771	3.971	569.930	500.000	14.0
Aroclor-1016-2	3.885	3.788	3.988	578.700	500.000	15.7
Aroclor-1016-3	4.046	3.950	4.150	567.410	500.000	13.5
Aroclor-1016-4	4.094	3.998	4.198	540.350	500.000	8.1
Aroclor-1016-5	4.292	4.195	4.395	557.710	500.000	11.5
Aroclor-1260-1	5.293	5.197	5.397	547.750	500.000	9.6
Aroclor-1260-2	5.483	5.386	5.586	553.130	500.000	10.6
Aroclor-1260-3	5.626	5.530	5.730	541.900	500.000	8.4
Aroclor-1260-4	6.091	5.993	6.193	583.940	500.000	16.8
Aroclor-1260-5	6.336	6.238	6.438	565.980	500.000	13.2
Decachlorobiphenyl	7.659	7.559	7.759	53.390	50.000	6.8
Tetrachloro-m-xylene	2.854	2.756	2.956	53.020	50.000	6.0