

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3672
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL06 **Date Analyzed:** 11/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071430.D **Time Analyzed:** 13:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.496	4.397	4.597	467.590	500.000	-6.5
Aroclor-1016-2	4.515	4.416	4.616	468.660	500.000	-6.3
Aroclor-1016-3	4.573	4.474	4.674	463.070	500.000	-7.4
Aroclor-1016-4	4.663	4.564	4.764	463.840	500.000	-7.2
Aroclor-1016-5	4.948	4.849	5.049	474.080	500.000	-5.2
Aroclor-1260-1	6.042	5.943	6.143	472.520	500.000	-5.5
Aroclor-1260-2	6.299	6.200	6.400	493.920	500.000	-1.2
Aroclor-1260-3	6.651	6.552	6.752	470.090	500.000	-6.0
Aroclor-1260-4	6.866	6.767	6.967	468.500	500.000	-6.3
Aroclor-1260-5	7.174	7.075	7.275	474.920	500.000	-5.0
Decachlorobiphenyl	8.536	8.437	8.637	42.850	50.000	-14.3
Tetrachloro-m-xylene	3.410	3.311	3.511	45.410	50.000	-9.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3672
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL06 **Date Analyzed:** 11/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071430.D **Time Analyzed:** 13:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.868	3.771	3.971	559.790	500.000	12.0
Aroclor-1016-2	3.884	3.788	3.988	576.450	500.000	15.3
Aroclor-1016-3	4.046	3.950	4.150	546.460	500.000	9.3
Aroclor-1016-4	4.094	3.998	4.198	528.280	500.000	5.7
Aroclor-1016-5	4.291	4.195	4.395	554.960	500.000	11.0
Aroclor-1260-1	5.292	5.197	5.397	553.910	500.000	10.8
Aroclor-1260-2	5.482	5.386	5.586	548.650	500.000	9.7
Aroclor-1260-3	5.625	5.530	5.730	544.610	500.000	8.9
Aroclor-1260-4	6.089	5.993	6.193	578.640	500.000	15.7
Aroclor-1260-5	6.334	6.238	6.438	554.740	500.000	10.9
Decachlorobiphenyl	7.654	7.559	7.759	52.010	50.000	4.0
Tetrachloro-m-xylene	2.855	2.756	2.956	52.480	50.000	5.0