

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3672
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL08 **Date Analyzed:** 11/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071471.D **Time Analyzed:** 04:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.497	4.397	4.597	490.540	500.000	-1.9
Aroclor-1016-2	4.516	4.416	4.616	494.220	500.000	-1.2
Aroclor-1016-3	4.573	4.474	4.674	484.310	500.000	-3.1
Aroclor-1016-4	4.664	4.564	4.764	480.250	500.000	-4.0
Aroclor-1016-5	4.949	4.849	5.049	487.140	500.000	-2.6
Aroclor-1260-1	6.043	5.943	6.143	487.540	500.000	-2.5
Aroclor-1260-2	6.299	6.200	6.400	496.510	500.000	-0.7
Aroclor-1260-3	6.652	6.552	6.752	490.270	500.000	-1.9
Aroclor-1260-4	6.867	6.767	6.967	484.300	500.000	-3.1
Aroclor-1260-5	7.174	7.075	7.275	489.760	500.000	-2.0
Decachlorobiphenyl	8.536	8.437	8.637	44.280	50.000	-11.4
Tetrachloro-m-xylene	3.411	3.311	3.511	47.490	50.000	-5.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3672
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL08 **Date Analyzed:** 11/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071471.D **Time Analyzed:** 04:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.868	3.771	3.971	490.430	500.000	-1.9
Aroclor-1016-2	3.884	3.788	3.988	501.320	500.000	0.3
Aroclor-1016-3	4.046	3.950	4.150	539.610	500.000	7.9
Aroclor-1016-4	4.095	3.998	4.198	524.920	500.000	5.0
Aroclor-1016-5	4.292	4.195	4.395	562.410	500.000	12.5
Aroclor-1260-1	5.292	5.197	5.397	467.500	500.000	-6.5
Aroclor-1260-2	5.481	5.386	5.586	466.830	500.000	-6.6
Aroclor-1260-3	5.624	5.530	5.730	452.590	500.000	-9.5
Aroclor-1260-4	6.089	5.993	6.193	499.630	500.000	-0.1
Aroclor-1260-5	6.333	6.238	6.438	500.020	500.000	0.0
Decachlorobiphenyl	7.653	7.559	7.759	42.520	50.000	-15.0
Tetrachloro-m-xylene	2.854	2.756	2.956	45.950	50.000	-8.1