

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3672
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL09 **Date Analyzed:** 11/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071474.D **Time Analyzed:** 10:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.511	4.397	4.597	512.130	500.000	2.4
Aroclor-1016-2	4.530	4.416	4.616	498.480	500.000	-0.3
Aroclor-1016-3	4.588	4.474	4.674	525.140	500.000	5.0
Aroclor-1016-4	4.678	4.564	4.764	521.200	500.000	4.2
Aroclor-1016-5	4.964	4.849	5.049	520.350	500.000	4.1
Aroclor-1260-1	6.058	5.943	6.143	477.880	500.000	-4.4
Aroclor-1260-2	6.314	6.200	6.400	497.890	500.000	-0.4
Aroclor-1260-3	6.667	6.552	6.752	506.130	500.000	1.2
Aroclor-1260-4	6.882	6.767	6.967	510.500	500.000	2.1
Aroclor-1260-5	7.190	7.075	7.275	513.380	500.000	2.7
Decachlorobiphenyl	8.555	8.437	8.637	44.490	50.000	-11.0
Tetrachloro-m-xylene	3.424	3.311	3.511	47.350	50.000	-5.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16
Lab Code: ACE **SDG NO.:** Q3672
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL09 **Date Analyzed:** 11/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071474.D **Time Analyzed:** 10:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.868	3.771	3.971	480.440	500.000	-3.9
Aroclor-1016-2	3.884	3.788	3.988	501.370	500.000	0.3
Aroclor-1016-3	4.046	3.950	4.150	482.430	500.000	-3.5
Aroclor-1016-4	4.095	3.998	4.198	477.500	500.000	-4.5
Aroclor-1016-5	4.292	4.195	4.395	482.090	500.000	-3.6
Aroclor-1260-1	5.293	5.197	5.397	474.740	500.000	-5.1
Aroclor-1260-2	5.483	5.386	5.586	483.780	500.000	-3.2
Aroclor-1260-3	5.626	5.530	5.730	474.980	500.000	-5.0
Aroclor-1260-4	6.090	5.993	6.193	477.920	500.000	-4.4
Aroclor-1260-5	6.335	6.238	6.438	483.460	500.000	-3.3
Decachlorobiphenyl	7.657	7.559	7.759	40.100	50.000	-19.8
Tetrachloro-m-xylene	2.853	2.756	2.956	45.100	50.000	-9.8