

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** TETR16  
**Lab Code:** ACE **SDG NO.:** Q3672  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

**Client Sample No.:** CCAL10 **Date Analyzed:** 11/24/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PQ071486.D **Time Analyzed:** 15:38

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.496 | 4.397     | 4.597 | 515.020            | 500.000           | 3.0   |
| Aroclor-1016-2       | 4.516 | 4.416     | 4.616 | 525.670            | 500.000           | 5.1   |
| Aroclor-1016-3       | 4.573 | 4.474     | 4.674 | 519.170            | 500.000           | 3.8   |
| Aroclor-1016-4       | 4.664 | 4.564     | 4.764 | 522.260            | 500.000           | 4.5   |
| Aroclor-1016-5       | 4.948 | 4.849     | 5.049 | 512.840            | 500.000           | 2.6   |
| Aroclor-1260-1       | 6.042 | 5.943     | 6.143 | 477.130            | 500.000           | -4.6  |
| Aroclor-1260-2       | 6.299 | 6.200     | 6.400 | 501.120            | 500.000           | 0.2   |
| Aroclor-1260-3       | 6.652 | 6.552     | 6.752 | 495.520            | 500.000           | -0.9  |
| Aroclor-1260-4       | 6.867 | 6.767     | 6.967 | 492.870            | 500.000           | -1.4  |
| Aroclor-1260-5       | 7.174 | 7.075     | 7.275 | 497.420            | 500.000           | -0.5  |
| Decachlorobiphenyl   | 8.537 | 8.437     | 8.637 | 42.900             | 50.000            | -14.2 |
| Tetrachloro-m-xylene | 3.410 | 3.311     | 3.511 | 46.930             | 50.000            | -6.1  |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** TETR16  
**Lab Code:** ACE **SDG NO.:** Q3672  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

**Client Sample No.:** CCAL10 **Date Analyzed:** 11/24/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PQ071486.D **Time Analyzed:** 15:38

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 3.867 | 3.771     | 3.971 | 496.080            | 500.000           | -0.8  |
| Aroclor-1016-2       | 3.884 | 3.788     | 3.988 | 512.700            | 500.000           | 2.5   |
| Aroclor-1016-3       | 4.045 | 3.950     | 4.150 | 494.830            | 500.000           | -1.0  |
| Aroclor-1016-4       | 4.094 | 3.998     | 4.198 | 485.710            | 500.000           | -2.9  |
| Aroclor-1016-5       | 4.291 | 4.195     | 4.395 | 494.130            | 500.000           | -1.2  |
| Aroclor-1260-1       | 5.291 | 5.197     | 5.397 | 488.180            | 500.000           | -2.4  |
| Aroclor-1260-2       | 5.481 | 5.386     | 5.586 | 497.580            | 500.000           | -0.5  |
| Aroclor-1260-3       | 5.624 | 5.530     | 5.730 | 484.620            | 500.000           | -3.1  |
| Aroclor-1260-4       | 6.088 | 5.993     | 6.193 | 484.460            | 500.000           | -3.1  |
| Aroclor-1260-5       | 6.332 | 6.238     | 6.438 | 484.670            | 500.000           | -3.1  |
| Decachlorobiphenyl   | 7.653 | 7.559     | 7.759 | 38.510             | 50.000            | -23.0 |
| Tetrachloro-m-xylene | 2.854 | 2.756     | 2.956 | 46.100             | 50.000            | -7.8  |