

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3673

Continuing Calib Date: 11/20/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 12:05

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1268-1 (1)	8.50	8.51	8.41	8.61	0.01
Aroclor-1268-2 (2)	8.59	8.60	8.50	8.70	0.01
Aroclor-1268-3 (3)	8.81	8.82	8.72	8.92	0.01
Aroclor-1268-4 (4)	9.20	9.22	9.12	9.32	0.02
Aroclor-1268-5 (5)	9.60	9.61	9.51	9.71	0.01
Tetrachloro-m-xylene	4.34	4.35	4.25	4.45	0.01
Decachlorobiphenyl	9.92	9.93	9.83	10.03	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR16

Lab Code: ACE

SDG NO.: Q3673

Continuing Calib Date: 11/20/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 12:05

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1268-1 (1)	7.53	7.54	7.44	7.64	0.01
Aroclor-1268-2 (2)	7.60	7.61	7.51	7.71	0.01
Aroclor-1268-3 (3)	7.80	7.81	7.71	7.91	0.01
Aroclor-1268-4 (4)	8.09	8.10	8.00	8.20	0.01
Aroclor-1268-5 (5)	8.38	8.39	8.29	8.49	0.01
Tetrachloro-m-xylene	3.65	3.65	3.55	3.75	0.01
Decachlorobiphenyl	8.63	8.64	8.54	8.74	0.01