284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16

Lab Code: ACE **SDG NO.:** Q3673

Continuing Calib Date: 11/21/2025 **Initial Calibration Date(s):** 11/04/2025 11/04/2025

Continuing Calib Time: 15:46 **Initial Calibration Time(s):** 09:39 18:31

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1268-1	(1)	8.50	8.51	8.41	8.61	0.01
Aroclor-1268-2	(2)	8.59	8.60	8.50	8.70	0.01
Aroclor-1268-3	(3)	8.81	8.82	8.72	8.92	0.01
Aroclor-1268-4	(4)	9.21	9.22	9.12	9.32	0.01
Aroclor-1268-5	(5)	9.60	9.61	9.51	9.71	0.01
Tetrachloro-m-xylene		4.34	4.35	4.25	4.45	0.01
Decachlorobiphenyl		9.92	9.93	9.83	10.03	0.01

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR16

Lab Code: ACE **SDG NO.:** Q3673

Continuing Calib Date: 11/21/2025 **Initial Calibration Date(s):** 11/04/2025 11/04/2025

Continuing Calib Time: 15:46 **Initial Calibration Time(s):** 09:39 18:31

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1268-1	(1)	7.53	7.54	7.44	7.64	0.01
Aroclor-1268-2	(2)	7.60	7.61	7.51	7.71	0.01
Aroclor-1268-3	(3)	7.81	7.81	7.71	7.91	0.01
Aroclor-1268-4	(4)	8.09	8.10	8.00	8.20	0.01
Aroclor-1268-5	(5)	8.38	8.39	8.29	8.49	0.01
Tetrachloro-m-xylene		3.65	3.65	3.55	3.75	0.01
Decachlorobiphenyl		8.63	8.64	8.54	8.74	0.01